



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Madison County S/N 80-001307
 Date In 5/23/17 Date Out 5/24/2017 Ship P/U H/D CMI EE

Intake Performed By JOE

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By JOE

Lab Temp °C 20.2
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 153
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm .144 (.139 - .169)
 36mm .157 (.156 - .190)
 53mm .273 (.228 - .278)
 103mm .496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # ATP103 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>503764</u>	<u>2016118</u> <u>11/15/18</u>
0.20	<u>611621</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>AG626604</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By JOE
Simulator Temperatures °C
 External Digital Therm. ID#: 300505
 34°C +- .2 Serial #: DR2035
 34°C +- .2 Serial #: 503764
 34°C +- .2 Serial #: 611621

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By JOE

Barometric Pressure 1001 Gauge
 ID# 26932 994 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-A

Simulator	Serial Number
0.00	<u>62879</u>
Interferent	<u>68144</u>
0.05	<u>DR2035</u>
0.08	<u>503764</u>
0.20	<u>611621</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:
OK 5/30/17

J. Zaham 5/30/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
05/24/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:23
Control Test	0.049	10:24
Air Blank	0.000	10:24
Control Test	0.049	10:25
Air Blank	0.000	10:25
Control Test	0.049	10:26
Air Blank	0.000	10:27
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
05/24/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:18
Control Test	0.079	10:19
Air Blank	0.000	10:19
Control Test	0.079	10:20
Air Blank	0.000	10:20
Control Test	0.079	10:21
Air Blank	0.000	10:22
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
05/24/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:11
Control Test	0.195	10:12
Air Blank	0.000	10:13
Control Test	0.196	10:13
Air Blank	0.000	10:14
Control Test	0.196	10:14
Air Blank	0.000	10:15
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

P Murphy

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
05/24/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:38
Control Test	0.079	10:38
Air Blank	0.000	10:39
Control Test	0.079	10:39
Air Blank	0.000	10:40
Control Test	0.079	10:40
Air Blank	0.000	10:40
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

P Murphy

Operator's Signature

JP

JP