

INSTRUMENT PROCESSING SHEET

Agency Okeechobee County SO S/N 80-001320
 Date In 3/24/17 Date Out 3/30/17 Ship P/U H/D CMI EE

RECEIVED
 APR 03 2017
 FDLE
 Alcohol Testing Program

Intake Performed By TR

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By SP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 141
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .156 (.139 - .169)
 36mm .175 (.156 - .190)
 53mm .242 (.228 - .278)
 103mm .496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G3709</u>	<u>201603D</u> <u>3-8-18</u>
0.08	<u>DR1279</u>	<u>201611B</u> <u>11-15-18</u>
0.20	<u>DR3856</u>	<u>201604C</u> <u>4-5-18</u>
0.08 DGS	N/A	<u>AG16261005</u> <u>9-22-18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure
 ID# 28427 1013 Gauge
1012 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>G3709</u>
0.08	<u>DR1279</u>
0.20	<u>DR3856</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC RDS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
 Quality Control Review

4/3/17
 Date

STABILITY CHECKS - INSTRUMENT # 80-001320 - OKEECHOBEE COUNTY SO

OKEECHOBEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001320
 03/30/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:37
Control Test	0.051	09:38
Air Blank	0.000	09:38
Control Test	0.050	09:39
Air Blank	0.000	09:39
Control Test	0.050	09:40
Air Blank	0.000	09:40
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

SP

Operator's Signature

OKEECHOBEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001320
 03/30/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:41
Control Test	0.080	09:42
Air Blank	0.000	09:43
Control Test	0.080	09:43
Air Blank	0.000	09:44
Control Test	0.080	09:44
Air Blank	0.000	09:45
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

OKEECHOBEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001320
 03/30/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:46
Control Test	0.202	09:47
Air Blank	0.000	09:47
Control Test	0.201	09:48
Air Blank	0.000	09:48
Control Test	0.202	09:49
Air Blank	0.000	09:50
Control Test Stats		
Average	0.2017	
Std Dev	0.0006	
Rel Std Dev(%)	0.2863	

SP

Operator's Signature

OKEECHOBEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001320
 03/30/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:33
Control Test	0.078	09:34
Air Blank	0.000	09:34
Control Test	0.079	09:35
Air Blank	0.000	09:35
Control Test	0.079	09:35
Air Blank	0.000	09:36
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

DGS

RWS BK

SP

Operator's Signature