

INSTRUMENT PROCESSING SHEET

Agency Indian River Sheriff's Office s/n 80-001328 Alcohol Testing Program Date Out 8/23/2017 Date In 8/22/2017 Ship P/U DH/D DCMI DEE Performed By Private Reyres Daniel Region Da Quality Checks Performed By Flow Calibration Performed By Intake Flow Calibration N/A FCF Flow Calibration Complete Registration ■ Lab Temp °C M Annual Breath Tube Screen ☐ Replace External O-Rings □ Return from CMI □ 5L/min - 1/11..... □ 15L/min - 103 mm DLE □ 30L/min - 103 mm DLE □ Return from Enforcement ☐ Instrument Set Up Verified ☐ R-Value Electronics ☐ Flow Verification (L/s) ☐ Other _ Flow Column # □ R-Value Visual Inspection: Post Calibration Verification () 32mm ____ (.139 - .169) √ Case ✓ Handle 36mm _____ (.156 - .190) Flow Column # ✓ Dry Gas Holder ✓ Feet 53mm _____ (.228 - .278) 32mm _____ (.139 - .169) ✓ Keyboard/Plug
✓ Back/Plugs 103mm _____ (.447 - .547) 36mm _____ (.156 - .190) ✓ Screws tight ✓ Breath Hose ■ Barometric Pressure Check 53mm _____ (.228 - .278) Gauge ID# 103mm ____ (.447 - .547) Other Equipment: Stability Checks Maintenance Performed By Power cord Simulator Serial # Lot #/Exp Printer Cable ■ Battery Replacement 0.05 Dry Gas Regulator Replacement Other ____ ■ Breath Tube Replacement Other 0.08 Notes: Quality Checks Cont. Performed By Simulator Temperatures °C 0.20 External Digital Therm. ID#: □ 34°C +-.2 Serial #: _____ N/A 0.08 DGS ☐ 34°C +-.2 Serial #: _____ ☐ 34°C +-.2 Serial #: **Calibration Adjustment** Performed By **Department Inspection** Performed By □ Calibration Adjustment N/A Barometric Pressure Gauge Calibration Adjustment Complete ID# Instrument Barometric Pressure Gauge ID# Simulator Serial Number Lot Number Expiration Mouth Alcohol Solution Lot # Acetone Stock Solution Lot # 0.000 N/A N/A Simulator Serial Number 0.040 0.00 0.100 Interferent 0.200 0.05 0.300 0.08 0.080 DGS N/A 0.20 Post Calibration Adjustment Stability Checks Simulator Serial Number Lot Number Expiration **Attachments** 0.05 Form 41 Calibration Adjustment 0.08 Pre-Stability Tests Post-Stability Tests ☐ Flow Calibration ☐ Other _____ 0.20 0.08 DGS N/A Notes/Suggested Service: ☐ Instrument Complies with Chapter 11D-8, FAC

M

facility for evaluation.

Instrument does not turn on. Instrument will be sent to repair

Instrument Does Not Comply with Chapter 11D-8, FAC

Conduct an Agency Inspection Before Evidentiary Use

Return to/Place into Evidentiary Use

Remain Out of Evidentiary Use