



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Orange County

S/N 80-001416

Date In 2/14/17

Date Out 3/2/17

Ship P/U H/D CMI EE

Intake Performed By DS

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case ok Handle ok
 Dry Gas Holder ok Feet ok
 Keyboard/Plug ok Back/Plugs ok
 Screws tight ok Breath Hose ok

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By DS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 126
 Flow Verification (L/s)
 Flow Column # ATP105
 32mm 148 (.139 - .169)
 36mm 167 (.156 - .190)
 53mm 234 (.228 - .278)
 103mm 500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD1011</u>	<u>201601F</u> <u>1/26/18</u>
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	<u>N/A</u>	<u>AG-626604</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
MAR 02 2017
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DS

Barometric Pressure 1016 Gauge
 ID# 26932 1016 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2017A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD1011</u>
0.20	<u>SD1025</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes:
QA/PC OK @ JPM 3/2/2017

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kuitland

Quality Control Review

3/2/17
Date

80-001416
 Stability checks
 3/2/17 P

INTOXILYZER 8000
 Instrument Initialization
 06:11 03/02/2017

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001416
 03/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:00
Control Test	0.079	07:01
Air Blank	0.000	07:01
Control Test	0.078	07:02
Air Blank	0.000	07:03
Control Test	0.078	07:03
Air Blank	0.000	07:04
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Operator's Signature

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001416
 03/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:06
Control Test	0.048	07:06
Air Blank	0.000	07:07
Control Test	0.047	07:08
Air Blank	0.000	07:08
Control Test	0.047	07:09
Air Blank	0.000	07:09
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	

Operator's Signature

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001416
 03/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:56
Control Test	0.080	06:57
Air Blank	0.000	06:57
Control Test	0.080	06:57
Air Blank	0.000	06:58
Control Test	0.079	06:58
Air Blank	0.000	06:59
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

Operator's Signature

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001416
 03/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:12
Control Test	0.197	07:13
Air Blank	0.000	07:13
Control Test	0.195	07:14
Air Blank	0.000	07:15
Control Test	0.194	07:15
Air Blank	0.000	07:16
Control Test Stats		
Average	0.1953	
Std Dev	0.0015	
Rel Std Dev(%)	0.7820	

RYM
 BSK

Operator's Signature