



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency UF PD S/N 80-001714
Date In 10/10/17 Date Out 10/19/17

Ship P/U H/D CMI EE

Intake section with checkboxes for Registration, Annual, Return from CMI, etc. Includes visual inspection and other equipment fields.

Quality Checks section with checkboxes for Lab Temp, Breath Tube Screen, Replace External O-Rings, etc. Includes a simulator table.

Flow Calibration and Maintenance sections with checkboxes for Flow Calibration N/A, Battery Replacement, etc. Includes a quality checks continuation section.

RECEIVED stamp: OCT 19 2017 FDLE Alcohol Testing Program

Calibration Adjustment section with checkboxes for Calibration Adjustment N/A, etc. Includes a simulator table.

Department Inspection section with checkboxes for Barometric Pressure, etc. Includes a simulator table.

Attachments section with checkboxes for Form 41, Pre-Stability Tests, etc.

Notes/Suggested Service section with handwritten notes and a signature.

Compliance section with checkboxes for Instrument Complies with Chapter 11D-8, FAC, etc.

80-0017124
 Stability Checks
 10/19/17

INTOXILYZER 8000
 Instrument Initialization
 07:07 10/19/2017

UNIVERSITY OF FL PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001714
 10/19/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:45
Control Test	0.078	07:45
Air Blank	0.000	07:46
Control Test	0.079	07:46
Air Blank	0.000	07:47
Control Test	0.078	07:47
Air Blank	0.000	07:47
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DGS

Operator's Signature

UNIVERSITY OF FL PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001714
 10/19/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:50
Control Test	0.198	07:51
Air Blank	0.000	07:51
Control Test	0.198	07:52
Air Blank	0.000	07:52
Control Test	0.198	07:53
Air Blank	0.000	07:53
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

UNIVERSITY OF FL PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001714
 10/19/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:55
Control Test	0.080	07:56
Air Blank	0.000	07:57
Control Test	0.080	07:57
Air Blank	0.000	07:58
Control Test	0.080	07:58
Air Blank	0.000	07:59
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Operator's Signature

SP

10/19/17
 SP