



INSTRUMENT PROCESSING SHEET

Agency Alachua PDS/N 80-001720Florida Department of
Law EnforcementDate In 10/26/17DI Completion Date 11/20/17☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake Performed By <u>DS</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DS</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>180</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP05</u> 32 mm <u>167</u> (.139 - .169) 36 mm <u>180</u> (.156 - .190) 53 mm <u>257</u> (.228 - .278) 103 mm <u>515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>G2835</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG708807</u> <u>3/29/19</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>G2835</u>	<u>201707D</u> <u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>	0.080 DGS	N/A	<u>AG708807</u> <u>3/29/19</u>	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>DS</u> ☒ Lab Temp °C <u>22.0</u> External Digital Therm. ID#: <u>30024</u> ☒ 34°C +/- .2 Serial #: <u>G2835</u> ☒ 34°C +/- .2 Serial #: <u>DR1279</u> ☒ 34°C +/- .2 Serial #: <u>SD1025</u>																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	<u>G2835</u>	<u>201707D</u> <u>7/25/19</u>																																																
0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>																																																
0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>																																																
0.080 DGS	N/A	<u>AG708807</u> <u>3/29/19</u>																																																
Final Release Date FDLE DEC 14 2017 Alcohol Testing Program	Calibration Adjustment <u>10206</u> Performed By <u>DS</u> Barometric Pressure Gauge <u>28427</u> ID # <u>28427</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>SD1016</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>SD1024</u></td> <td><u>16320</u></td> <td><u>10/21/18</u></td> </tr> <tr> <td>0.100</td> <td><u>G2834</u></td> <td><u>17060</u></td> <td><u>2/14/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1011</u></td> <td><u>17090</u></td> <td><u>2/24/19</u></td> </tr> <tr> <td>0.300</td> <td><u>DR1275</u></td> <td><u>17140</u></td> <td><u>5/15/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>17817080A2</u></td> <td><u>8/5/19</u></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>G2835</u></td> <td><u>201707D</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>201707C</u></td> <td><u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: _____ _____ _____ _____ _____		Simulator	Serial Number	Lot Number	Expiration	0.000	<u>SD1016</u>	N/A	N/A	0.040	<u>SD1024</u>	<u>16320</u>	<u>10/21/18</u>	0.100	<u>G2834</u>	<u>17060</u>	<u>2/14/19</u>	0.200	<u>SD1011</u>	<u>17090</u>	<u>2/24/19</u>	0.300	<u>DR1275</u>	<u>17140</u>	<u>5/15/19</u>	0.080 DGS	N/A	<u>17817080A2</u>	<u>8/5/19</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>G2835</u>	<u>201707D</u>	<u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u>	<u>7/25/19</u>	0.200	<u>SD1025</u>	<u>201707C</u>	<u>7/24/19</u>	0.080 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																															
0.000	<u>SD1016</u>	N/A	N/A																																															
0.040	<u>SD1024</u>	<u>16320</u>	<u>10/21/18</u>																																															
0.100	<u>G2834</u>	<u>17060</u>	<u>2/14/19</u>																																															
0.200	<u>SD1011</u>	<u>17090</u>	<u>2/24/19</u>																																															
0.300	<u>DR1275</u>	<u>17140</u>	<u>5/15/19</u>																																															
0.080 DGS	N/A	<u>17817080A2</u>	<u>8/5/19</u>																																															
Simulator	Serial Number	Lot Number	Expiration																																															
0.050	<u>G2835</u>	<u>201707D</u>	<u>7/25/19</u>																																															
0.080	<u>DR1279</u>	<u>201707E</u>	<u>7/25/19</u>																																															
0.200	<u>SD1025</u>	<u>201707C</u>	<u>7/24/19</u>																																															
0.080 DGS	N/A																																																	
Department Inspection Performed By <u>DS</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1019</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2880</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2840</u></td> </tr> <tr> <td>0.050</td> <td><u>G2835</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> </tr> </tbody> </table> Attachments ☒ Form 41 x2 ☐ Flow Calibration ☒ Stability Checks ☐ Form 40 ☒ Calibration Adjustment ☐ Other _____ ☒ Post-Stability Checks ☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use <u>DS 12/14/17</u> Tech Review / Date <u>JJ Dake 12/14/17</u> Admin Review / Date			Simulator	Serial Number	0.000	<u>G2880</u>	Interferent	<u>G2840</u>	0.050	<u>G2835</u>	0.080	<u>DR1279</u>	0.200	<u>SD1025</u>																																				
Simulator	Serial Number																																																	
0.000	<u>G2880</u>																																																	
Interferent	<u>G2840</u>																																																	
0.050	<u>G2835</u>																																																	
0.080	<u>DR1279</u>																																																	
0.200	<u>SD1025</u>																																																	

80-001720
Stability checks
11/2/17

INTOXILYZER 8000
Instrument Initialization
05:32 11/02/2017

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000
11/02/2017
Software: 8100.27
SN 80-001720

Test	g/210L	Time
Air Blank	0.000	06:13
Control Test	0.081	06:14
Air Blank	0.000	06:14
Control Test	0.082	06:15
Air Blank	0.000	06:15
Control Test	0.082	06:15
Air Blank	0.000	06:16
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

DG

Operator's Signature

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000
11/02/2017
Software: 8100.27
SN 80-001720

Test	g/210L	Time
Air Blank	0.000	06:19
Control Test	0.203	06:19
Air Blank	0.000	06:20
Control Test	0.203	06:21
Air Blank	0.000	06:21
Control Test	0.203	06:22
Air Blank	0.000	06:22
Control Test Stats		
Average	0.2030	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000
11/02/2017
Software: 8100.27
SN 80-001720

Test	g/210L	Time
Air Blank	0.000	06:25
Control Test	0.081	06:25
Air Blank	0.000	06:26
Control Test	0.082	06:27
Air Blank	0.000	06:27
Control Test	0.082	06:28
Air Blank	0.000	06:28
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

Operator's Signature

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000
11/02/2017
Software: 8100.27
SN 80-001720

Test	g/210L	Time
Air Blank	0.000	06:30
Control Test	0.050	06:31
Air Blank	0.000	06:31
Control Test	0.052	06:32
Air Blank	0.000	06:33
Control Test	0.051	06:33
Air Blank	0.000	06:34
Control Test Stats		
Average	0.0510	
Std Dev	0.0010	
Rel Std Dev(%)	1.9608	

DG

Operator's Signature

12/14/17

INTOXILYZER 8000
Instrument Initialization
08:08 11/20/2017

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001720
11/20/2017 08:39:03

Auto Calibration
Max Power Res Value = 31
Auto Range Res Value = 13

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12581, Sum Io = 13724
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0920 (-0.0030)
Sample #2 = 0.1010 (0.0340)
Sample #3 = 0.1270 (0.0470)
Sample #4 = 0.1160 (0.0820)
Avg % Abs = 0.1147 (0.0543)
STD DEV = 0.0131 (0.0248)
REL STD DEV = 11.382 (45.692)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.3350 (0.1720)
Sample #2 = 0.1530 (0.3710)
Sample #3 = 0.0440 (0.4340)
Sample #4 = 0.3380 (0.4480)
Avg % Abs = 0.1783 (0.4177)
STD DEV = 0.1486 (0.0410)
REL STD DEV = 83.343 (9.820)

12/4/17
22

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12551, Sum Io = 13681
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8420 (-0.0280)
Sample #2 = 0.8220 (0.0070)
Sample #3 = 0.8420 (0.0410)
Sample #4 = 0.8140 (0.0770)
Avg % Abs = 0.8260 (0.0417)
STD DEV = 0.0144 (0.0350)
REL STD DEV = 1.746 (84.011)

**** AUTO CAL FAIL

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.6770 (-0.0500)
Sample #2 = 1.4480 (0.3150)
Sample #3 = 1.3610 (-0.0100)
Sample #4 = 1.2530 (0.0570)
Avg % Abs = 1.3540 (0.1207)
STD DEV = 0.0977 (0.1716)
REL STD DEV = 7.215 (142.209)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12544, Sum Io = 13763
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8320 (-0.0270)
Sample #2 = 0.8480 (-0.0110)
Sample #3 = 0.8400 (0.0050)
Sample #4 = 0.8320 (0.0360)
Avg % Abs = 0.8400 (0.0100)
STD DEV = 0.0080 (0.0239)
REL STD DEV = 0.952 (238.956)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.7670 (-0.2080)
Sample #2 = 1.8740 (-0.0950)
Sample #3 = 1.4670 (0.3170)
Sample #4 = 1.7080 (0.3190)
Avg % Abs = 1.6830 (0.1803)
STD DEV = 0.2046 (0.2384)
REL STD DEV = 12.160 (132.226)

80-001720
Cal adjustment.
11/20/17

Fault det. on .040
Re try and Fault det
again. Stopped Cal Adjust
and will redo again to
Curt.

22

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001720
11/20/2017 09:13:17

Auto Calibration
Max Power Res Value = 32
Auto Range Res Value = 13

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12561, Sum Io = 13621
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1330 (-0.0220)
Sample #2 = 0.1120 (0.0390)
Sample #3 = 0.1110 (0.0850)
Sample #4 = 0.0980 (0.1240)
Avg % Abs = 0.1070 (0.0827)
STD DEV = 0.0078 (0.0425)
REL STD DEV = 7.299 (51.469)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1190 (-0.0320)
Sample #2 = -0.4210 (-0.2880)
Sample #3 = 0.4100 (-0.4150)
Sample #4 = -0.1200 (-0.0580)
Avg % Abs = -0.0437 (-0.2537)
STD DEV = 0.4207 (0.1810)
REL STD DEV = 963.494 (71.338)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12535, Sum Io = 13689
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8290 (0.0000)
Sample #2 = 0.8240 (0.0310)
Sample #3 = 0.8170 (0.0600)
Sample #4 = 0.8220 (0.0840)
Avg % Abs = 0.8210 (0.0583)
STD DEV = 0.0036 (0.0265)
REL STD DEV = 0.439 (45.496)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5690 (-0.1620)
Sample #2 = 1.3510 (0.2630)
Sample #3 = 1.5840 (0.2830)
Sample #4 = 1.6640 (0.3590)
Avg % Abs = 1.5330 (0.3017)
STD DEV = 0.1626 (0.0506)
REL STD DEV = 10.608 (16.790)

**** AUTO CAL FAIL

80-001720

Cal. Adjust

11/20/17

Auto Cal
Std. Dev. High

Fault Det on 04 Solution

Reason unknown.

- send to
Authorized Repair
Facility.

RMB

11/14/17
JO



INSTRUMENT PROCESSING SHEET

Florida Department of
Law Enforcement

Agency Alachua PD

S/N 80-001720

Date In 10/26/17

DI Completion Date 11/20/17

☐ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>AS</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ Final Release Date _____ _____ _____	Quality Checks Performed By <u>AS</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>180</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP05</u> 32 mm <u>167</u> (.139 - .169) 36 mm <u>180</u> (.156 - .190) 53 mm <u>257</u> (.228 - .278) 103 mm <u>515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>G2835</td> <td>201707D 7/25/19</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E 7/25/19</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>201707C 7/24/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG708807 3/29/19</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	G2835	201707D 7/25/19	0.080	DR1279	201707E 7/25/19	0.200	SD1025	201707C 7/24/19	0.080 DGS	N/A	AG708807 3/29/19	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>AS</u> ☒ Lab Temp °C <u>22.0</u> External Digital Therm. ID#: <u>300504</u> ☒ 34°C +/- .2 Serial #: <u>G2835</u> ☒ 34°C +/- .2 Serial #: <u>DR1279</u> ☒ 34°C +/- .2 Serial #: <u>SD1025</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	G2835	201707D 7/25/19																																																												
0.080	DR1279	201707E 7/25/19																																																												
0.200	SD1025	201707C 7/24/19																																																												
0.080 DGS	N/A	AG708807 3/29/19																																																												
Calibration Adjustment <u>1020</u> Performed By <u>AS</u> Barometric Pressure Gauge <u>28427</u> ID # <u>28427</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1016</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1024</td> <td>16320</td> <td>10/21/18</td> </tr> <tr> <td>0.100</td> <td>G2834</td> <td>17060</td> <td>2/14/19</td> </tr> <tr> <td>0.200</td> <td>SD1011</td> <td>17090</td> <td>2/24/19</td> </tr> <tr> <td>0.300</td> <td>DR1275</td> <td>17140</td> <td>5/15/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>17817080A2</td> <td>8/5/19</td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>G2835</td> <td>201707D</td> <td>7/25/19</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201707E</td> <td>7/25/19</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>201707C</td> <td>7/24/19</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	SD1016	N/A	N/A	0.040	SD1024	16320	10/21/18	0.100	G2834	17060	2/14/19	0.200	SD1011	17090	2/24/19	0.300	DR1275	17140	5/15/19	0.080 DGS	N/A	17817080A2	8/5/19	Simulator	Serial Number	Lot Number	Expiration	0.050	G2835	201707D	7/25/19	0.080	DR1279	201707E	7/25/19	0.200	SD1025	201707C	7/24/19	0.080 DGS	N/A			Department Inspection Performed By <u>AS</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1019</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2880</td> </tr> <tr> <td>Interferent</td> <td>G2840</td> </tr> <tr> <td>0.050</td> <td>G2835</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> </tr> </tbody> </table> Attachments ☐ Form 41 ☒ Stability Checks ☒ Calibration Adjustment ☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____		Simulator	Serial Number	0.000	G2880	Interferent	G2840	0.050	G2835	0.080	DR1279	0.200	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000	SD1016	N/A	N/A																																																											
0.040	SD1024	16320	10/21/18																																																											
0.100	G2834	17060	2/14/19																																																											
0.200	SD1011	17090	2/24/19																																																											
0.300	DR1275	17140	5/15/19																																																											
0.080 DGS	N/A	17817080A2	8/5/19																																																											
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050	G2835	201707D	7/25/19																																																											
0.080	DR1279	201707E	7/25/19																																																											
0.200	SD1025	201707C	7/24/19																																																											
0.080 DGS	N/A																																																													
Simulator	Serial Number																																																													
0.000	G2880																																																													
Interferent	G2840																																																													
0.050	G2835																																																													
0.080	DR1279																																																													
0.200	SD1025																																																													
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____		☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date _____ Admin Review / Date _____																																																												

INTOXILYZER 8000
Instrument Initialization
08:08 11/20/2017

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001720
11/20/2017 08:39:03

Auto Calibration
Max Power Res Value = 31
Auto Range Res Value = 13

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12581, Sum Io = 13724

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	= 0.0920	(-0.0030)
Sample #2	= 0.1010	(0.0340)
Sample #3	= 0.1270	(0.0470)
Sample #4	= 0.1160	(0.0820)
Avg % Abs	= 0.1147	(0.0543)
STD DEV	= 0.0131	(0.0248)
REL STD DEV	= 11.382	(45.692)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	= 0.3350	(0.1720)
Sample #2	= 0.1530	(0.3710)
Sample #3	= 0.0440	(0.4340)
Sample #4	= 0.3380	(0.4480)
Avg % Abs	= 0.1783	(0.4177)
STD DEV	= 0.1486	(0.0410)
REL STD DEV	= 83.343	(9.820)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12561, Sum Io = 13581

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	= 0.8420	(-0.0280)
Sample #2	= 0.8220	(0.0070)
Sample #3	= 0.8420	(0.0410)
Sample #4	= 0.8140	(0.0770)
Avg % Abs	= 0.8260	(0.0417)
STD DEV	= 0.0144	(0.0350)
REL STD DEV	= 1.746	(84.011)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	= 1.6770	(-0.0500)
Sample #2	= 1.4480	(0.3150)
Sample #3	= 1.3610	(-0.0100)
Sample #4	= 1.2530	(0.0570)
Avg % Abs	= 1.3540	(0.1207)
STD DEV	= 0.0977	(0.1716)
REL STD DEV	= 7.215	(142.209)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12544, Sum Io = 13763

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	= 0.8320	(-0.0270)
Sample #2	= 0.8480	(-0.0110)
Sample #3	= 0.8400	(0.0050)
Sample #4	= 0.8320	(0.0360)
Avg % Abs	= 0.8400	(0.0100)
STD DEV	= 0.0080	(0.0239)
REL STD DEV	= 0.952	(238.956)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	= 1.7670	(-0.2080)
Sample #2	= 1.8740	(-0.0950)
Sample #3	= 1.4670	(0.3170)
Sample #4	= 1.7080	(0.3190)
Avg % Abs	= 1.6830	(0.1803)
STD DEV	= 0.2046	(0.2384)
REL STD DEV	= 12.160	(132.226)

80-001720
Cal adjustment +
11/20/17

Fault det. on .040
Re try and Fault det
again. Stopped Cal Adjust
and will redo again to
check.

**** AUTO CAL FAIL

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001720
11/20/2017 09:13:17

Auto Calibration
Max Power Res Value = 32
Auto Range Res Value = 13

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12561, Sum Io = 13621

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1330	(-0.0220)
Sample #2 =	0.1120	(0.0390)
Sample #3 =	0.1110	(0.0850)
Sample #4 =	0.0980	(0.1240)
Avg % Abs =	0.1070	(0.0827)
STD DEV =	0.0078	(0.0425)
REL STD DEV =	7.299	(51.469)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1190	(-0.0320)
Sample #2 =	-0.4210	(-0.2880)
Sample #3 =	0.4100	(-0.4150)
Sample #4 =	-0.1200	(-0.0580)
Avg % Abs =	-0.0437	(-0.2537)
STD DEV =	0.4207	(0.1810)
REL STD DEV =	963.494	(71.338)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12535, Sum Io = 13689

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.8290	(0.0000)
Sample #2 =	0.8240	(0.0310)
Sample #3 =	0.8170	(0.0600)
Sample #4 =	0.8220	(0.0840)
Avg % Abs =	0.8210	(0.0583)
STD DEV =	0.0036	(0.0265)
REL STD DEV =	0.439	(45.496)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.5690	(-0.1620)
Sample #2 =	1.3510	(0.2630)
Sample #3 =	1.5840	(0.2830)
Sample #4 =	1.6640	(0.3590)
Avg % Abs =	1.5330	(0.3017)
STD DEV =	0.1626	(0.0506)
REL STD DEV =	10.608	(16.790)

**** AUTO CAL FAIL

80-001720

Cal. Adjust

11/20/17

Auto Cal
Std. Dev. High

Fault Det on 04 Solution

Reason unknown.

- send to
Authorized Repair
Facility.

80-001720
Stability checks
11/2/17

INTOXILYZER 8000
Instrument Initialization
05:32 11/02/2017

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000
11/02/2017
Software: 8100.27
SN 80-001720

Test	g/210L	Time
Air Blank	0.000	06:13
Control Test	0.081	06:14
Air Blank	0.000	06:14
Control Test	0.082	06:15
Air Blank	0.000	06:15
Control Test	0.082	06:15
Air Blank	0.000	06:16
Control Test	0.082	06:16
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

DE

Operator's Signature

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000
11/02/2017
Software: 8100.27
SN 80-001720

Test	g/210L	Time
Air Blank	0.000	06:19
Control Test	0.203	06:19
Air Blank	0.000	06:20
Control Test	0.203	06:21
Air Blank	0.000	06:21
Control Test	0.203	06:22
Air Blank	0.000	06:22
Control Test	0.203	06:22
Average	0.2030	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000
11/02/2017
Software: 8100.27
SN 80-001720

Test	g/210L	Time
Air Blank	0.000	06:25
Control Test	0.081	06:25
Air Blank	0.000	06:26
Control Test	0.082	06:27
Air Blank	0.000	06:27
Control Test	0.082	06:28
Air Blank	0.000	06:28
Control Test	0.082	06:28
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

Operator's Signature

ALACHUA PD
Intoxilyzer - Alcohol Analyzer
Model 8000
11/02/2017
Software: 8100.27
SN 80-001720

Test	g/210L	Time
Air Blank	0.000	06:30
Control Test	0.050	06:31
Air Blank	0.000	06:31
Control Test	0.052	06:32
Air Blank	0.000	06:33
Control Test	0.051	06:33
Air Blank	0.000	06:34
Control Test	0.051	06:34
Average	0.0510	
Std Dev	0.0010	
Rel Std Dev(%)	1.9608	

Operator's Signature