



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Lakeland PD

S/N 80-001865

Date In 5/3/17

Date Out 5/4/17

Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input checked="" type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>[Signature]</u> Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>213</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>0.152</u> (.139 - .169) 36mm <u>0.160</u> (.156 - .190) 53mm <u>0.234</u> (.228 - .278) 103mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>DR2035</td> <td>201603D 3/8/18</td> </tr> <tr> <td>0.08</td> <td>SD3964</td> <td>201611B 11/15/18</td> </tr> <tr> <td>0.20</td> <td>G11621</td> <td>201604C 4/15/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG626605 9/22/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	DR2035	201603D 3/8/18	0.08	SD3964	201611B 11/15/18	0.20	G11621	201604C 4/15/18	0.08 DGS	N/A	AG626605 9/22/18	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # <u>MAY</u> <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	DR2035	201603D 3/8/18															
0.08	SD3964	201611B 11/15/18															
0.20	G11621	201604C 4/15/18															
0.08 DGS	N/A	AG626605 9/22/18															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____															

RECEIVED
MAY 04 2017
ALCOHOL TESTING PROGRAM

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000		N/A	N/A																				
0.040																							
0.100																							
0.200																							
0.400																							
0.080 DGS	N/A																						
<input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																				
0.05																							
0.08																							
0.20																							
0.08 DGS	N/A																						

Department Inspection Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Barometric Pressure ID# <u>28421</u> _____ Gauge <u>1007</u> _____ Instrument <u>1007</u>	
Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u>	
Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	DR2035
0.08	SD3964
0.20	G11621
Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	

Notes: Please update menu level 2 to a unique password. [Signature]

PA/RC OK SP

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC
<input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC
<input checked="" type="checkbox"/> Return to/Place into Evidentiary Use
<input type="checkbox"/> Remain Out of Evidentiary Use
<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

[Signature]
Quality Control Review

5/4/17
Date

Stability Checks # 80-001865 Lakeland P.D. 5/4/17 *RMS*

DGS

LAKELAND PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001865
05/04/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:25
Control Test	0.051	09:26
Air Blank	0.000	09:27
Control Test	0.050	09:27
Air Blank	0.000	09:28
Control Test	0.051	09:28
Air Blank	0.000	09:29
Control Test Stats	0.0507	
Average	0.0006	
Std Dev	1.1395	
Rel Std Dev(%)		

RMS

Operator's Signature

LAKELAND PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001865
05/04/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:30
Control Test	0.081	09:31
Air Blank	0.000	09:31
Control Test	0.080	09:32
Air Blank	0.000	09:32
Control Test	0.081	09:33
Air Blank	0.000	09:34
Control Test Stats	0.0807	
Average	0.0006	
Std Dev	0.7157	
Rel Std Dev(%)		

RMS

Operator's Signature

LAKELAND PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001865
05/04/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:35
Control Test	0.199	09:36
Air Blank	0.000	09:36
Control Test	0.198	09:37
Air Blank	0.000	09:37
Control Test	0.198	09:38
Air Blank	0.000	09:39
Control Test Stats	0.1983	
Average	0.0006	
Std Dev	0.2911	
Rel Std Dev(%)		

RMS

Operator's Signature

LAKELAND PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001865
05/04/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:20
Control Test	0.079	09:21
Air Blank	0.000	09:21
Control Test	0.079	09:22
Air Blank	0.000	09:22
Control Test	0.079	09:22
Air Blank	0.000	09:23
Control Test Stats	0.0790	
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SR

JD

RMS

Operator's Signature