



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Pasco County S.O. S/N 80-003347
Date In 6/13/17 Date Out 6/13/17 Ship P/U H/D CMI EE

Intake Performed By <u>KMB</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____	Quality Checks Performed By <u>KMB</u> <input checked="" type="checkbox"/> Lab Temp °C <u>24.2</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>210</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>APP103</u> 32mm <u>0.144</u> (.139 - .169) 36mm <u>0.164</u> (.156 - .190) 53mm <u>0.230</u> (.228 - .278) 103mm <u>0.500</u> (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)															
	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD3962</u></td> <td><u>201603D</u> <u>3/8/18</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> <td><u>201611B</u> <u>11/15/18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3933</u></td> <td><u>201604C</u> <u>4/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG612405</u> <u>5/3/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>SD3962</u>	<u>201603D</u> <u>3/8/18</u>	0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>	0.20	<u>SD3933</u>	<u>201604C</u> <u>4/5/18</u>	0.08 DGS	N/A	<u>AG612405</u> <u>5/3/18</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Quality Checks Cont. Performed By <u>KMB</u> Simulator Temperatures °C External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3962</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3964</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3933</u>
Simulator	Serial #	Lot #/Exp															
0.05	<u>SD3962</u>	<u>201603D</u> <u>3/8/18</u>															
0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>															
0.20	<u>SD3933</u>	<u>201604C</u> <u>4/5/18</u>															
0.08 DGS	N/A	<u>AG612405</u> <u>5/3/18</u>															

Calibration Adjustment Performed By _____ <input checked="" type="checkbox"/> Calibration Adjustment N/A <input type="checkbox"/> Calibration Adjustment Complete Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>KMB</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1017</u> Gauge ID# <u>28421</u> <u>1019</u> Instrument Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td><u>G2880</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2834</u></td> </tr> <tr> <td>0.05</td> <td><u>SD3962</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3933</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	<u>G2880</u>	Interferent	<u>G2834</u>	0.05	<u>SD3962</u>	0.08	<u>SD3964</u>	0.20	<u>SD3933</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	<u>G2880</u>																																																												
Interferent	<u>G2834</u>																																																												
0.05	<u>SD3962</u>																																																												
0.08	<u>SD3964</u>																																																												
0.20	<u>SD3933</u>																																																												
	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____																																																												

Notes/Suggested Service:
QA/QC OK SP
JJ Dehan
Quality Control Review
Date 6/13/17

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC
- Return to/Place into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

Stability Checks # 80-003347 Pasco County S.O. 6/13/17 *AMS*

AMS

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003347
06/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:20
Control Test	0.050	09:21
Air Blank	0.000	09:22
Control Test	0.050	09:22
Air Blank	0.000	09:23
Control Test	0.050	09:23
Air Blank	0.000	09:24
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AMS
Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003347
06/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:26
Control Test	0.080	09:27
Air Blank	0.000	09:28
Control Test	0.080	09:28
Air Blank	0.000	09:29
Control Test	0.080	09:30
Air Blank	0.000	09:30
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AMS
Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003347
06/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.200	09:32
Air Blank	0.000	09:33
Control Test	0.200	09:34
Air Blank	0.000	09:34
Control Test	0.199	09:35
Air Blank	0.000	09:35
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

AMS
Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003347
06/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:37
Control Test	0.077	09:37
Air Blank	0.000	09:38
Control Test	0.078	09:38
Air Blank	0.000	09:38
Control Test	0.078	09:39
Air Blank	0.000	09:39
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

SP

AMS
Operator's Signature

AMS