

INSTRUMENT PROCESSING SHEET

Agency FHP - Troop H

S/N 80-003406

Date In 4/12/17

Date Out 4/13/17

Ship P/U H/D CMI EE

Intake Performed By <u>DD</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	Quality Checks Performed By <u>QSYM</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>195</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>.156</u> (.139 - .169) 36mm <u>.164</u> (.156 - .190) 53mm <u>.230</u> (.228 - .278) 103mm <u>.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>DR2035</u></td> <td><u>201603D</u> <u>3/8/18</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> <td><u>201611B</u> <u>11/15/18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD1025</u></td> <td><u>201604C</u> <u>4/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>RG626604</u> <u>9/22/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>	0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>	0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>	0.08 DGS	N/A	<u>RG626604</u> <u>9/22/18</u>	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>															
0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>															
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>															
0.08 DGS	N/A	<u>RG626604</u> <u>9/22/18</u>															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____															

RECEIVED
 APR 18 2017
 ALCOHOL TESTING PROGRAM

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000		N/A	N/A																				
0.040																							
0.100																							
0.200																							
0.400																							
0.080 DGS	N/A																						
<input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																				
0.05																							
0.08																							
0.20																							
0.08 DGS	N/A																						

Department Inspection Performed By <u>QSYM</u> <input checked="" type="checkbox"/> Barometric Pressure ID# <u>28421</u> Gauge <u>1020</u> Instrument <u>1020</u>	
Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u>	
Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD3964</u>
0.20	<u>SD1025</u>
Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Other _____	

Notes: Investigated possible power problem. No problem found.
QA/QC OK SP

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

JJ Dahan
Quality Control Review

4/18/17
Date

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
04/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:08
Control Test	0.049	08:08
Air Blank	0.000	08:09
Control Test	0.050	08:10
Air Blank	0.000	08:10
Control Test	0.050	08:11
Air Blank	0.000	08:11
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

P Murphy

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
04/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:03
Control Test	0.080	08:03
Air Blank	0.000	08:04
Control Test	0.080	08:05
Air Blank	0.000	08:05
Control Test	0.081	08:06
Air Blank	0.000	08:06
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

P Murphy

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
04/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:58
Control Test	0.198	07:59
Air Blank	0.000	07:59
Control Test	0.198	08:00
Air Blank	0.000	08:00
Control Test	0.198	08:01
Air Blank	0.000	08:02
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
04/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:23
Control Test	0.082	08:23
Air Blank	0.000	08:24
Control Test	0.081	08:24
Air Blank	0.000	08:25
Control Test	0.081	08:25
Air Blank	0.000	08:26
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

DGS

P Murphy

Operator's Signature

SP JD