

## INSTRUMENT PROCESSING SHEET

Agency H. I. Best AFB S/N 80-004771  
 Date In 6/2/17 Date Out 6/2/17  Ship  P/U  H/D  CMI  EE

| <b>Intake</b> Performed By <u>[Signature]</u><br><input type="checkbox"/> Registration<br><input checked="" type="checkbox"/> Annual<br><input type="checkbox"/> Return from CMI<br><input type="checkbox"/> Return from Enforcement Electronics<br><input type="checkbox"/> Other _____<br>Visual Inspection:<br><u>OK</u> Case <u>OK</u> Handle<br><u>OK</u> Dry Gas Holder <u>OK</u> Feet<br><u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs<br><u>OK</u> Screws tight <u>OK</u> Breath Hose<br>Other Equipment:<br><input type="checkbox"/> Power cord<br><input type="checkbox"/> Printer Cable<br><input type="checkbox"/> Other _____<br>Notes: _____<br>_____<br>_____<br>_____ | <b>Quality Checks</b> Performed By _____<br><input type="checkbox"/> Lab Temp °C _____<br><input type="checkbox"/> Breath Tube Screen<br><input type="checkbox"/> Replace External O-Rings<br><input type="checkbox"/> Instrument Set Up Verified<br><input type="checkbox"/> R-Value _____<br><input type="checkbox"/> Flow Verification (L/s)<br>Flow Column # _____<br>32mm _____ (.139 - .169)<br>36mm _____ (.156 - .190)<br>53mm _____ (.228 - .278)<br>103mm _____ (.447 - .547)<br><input type="checkbox"/> Barometric Pressure Check<br>Gauge ID # _____<br><input type="checkbox"/> Stability Checks<br><table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> </tr> </tbody> </table> | Simulator | Serial # | Lot #/Exp | 0.05 |  |  | 0.08 |  |  | 0.20 |  |  | 0.08 DGS | N/A |  | <b>Flow Calibration</b> Performed By _____<br><input type="checkbox"/> Flow Calibration N/A<br><input type="checkbox"/> Flow Calibration Complete<br>Flow Column # _____<br><input type="checkbox"/> 5L/min - 17mm<br><input type="checkbox"/> 15L/min - 53mm<br><input type="checkbox"/> 30L/min - 103mm<br><input type="checkbox"/> R-Value _____<br><input type="checkbox"/> Post Calibration Verification (L/s)<br>Flow Column # _____<br>32mm _____ (.139 - .169)<br>36mm _____ (.156 - .190)<br>53mm _____ (.228 - .278)<br>103mm _____ (.447 - .547)<br><b>Maintenance</b> Performed By _____<br><input type="checkbox"/> Battery Replacement<br><input type="checkbox"/> Dry Gas Regulator Replacement<br><input type="checkbox"/> Breath Tube Replacement<br><input type="checkbox"/> Other _____<br><b>Quality Checks Cont.</b> Performed By _____<br><b>Simulator Temperatures °C</b><br>External Digital Therm. ID#: _____<br><input type="checkbox"/> 34°C +/- .2 Serial #: _____<br><input type="checkbox"/> 34°C +/- .2 Serial #: _____<br><input type="checkbox"/> 34°C +/- .2 Serial #: _____ |
|--|--|-----------|----------|-----------|------|--|--|------|--|--|------|--|--|----------|-----|--|--|
| Simulator  | Serial #   | Lot #/Exp |          |           |      |  |  |      |  |  |      |  |  |          |     |  |  |
| 0.05   |  |           |          |           |      |  |  |      |  |  |      |  |  |          |     |  |  |
| 0.08   |  |           |          |           |      |  |  |      |  |  |      |  |  |          |     |  |  |
| 0.20   |  |           |          |           |      |  |  |      |  |  |      |  |  |          |     |  |  |
| 0.08 DGS   | N/A  |           |          |           |      |  |  |      |  |  |      |  |  |          |     |  |  |

| <b>Calibration Adjustment</b> Performed By _____<br><input type="checkbox"/> Calibration Adjustment N/A<br><input type="checkbox"/> Calibration Adjustment Complete<br>Barometric Pressure Gauge _____ ID # _____<br><table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks<br><table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> | Simulator     | Serial Number | Lot Number | Expiration | 0.000 |  | N/A | N/A | 0.040 |  |  |  | 0.100 |  |  |  | 0.200 |  |  |  | 0.300 |  |  |  | 0.080 DGS | N/A |  |  | Simulator | Serial Number | Lot Number | Expiration | 0.05 |  |  |  | 0.08 |  |  |  | 0.20 |  |  |  | 0.08 DGS | N/A |  |  | <b>Department Inspection</b> Performed By _____<br><input type="checkbox"/> Barometric Pressure _____ Gauge<br>ID# _____ Instrument<br>Mouth Alcohol Solution Lot # _____<br>Acetone Stock Solution Lot # _____<br><table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.05</td> <td></td> </tr> <tr> <td>0.08</td> <td></td> </tr> <tr> <td>0.20</td> <td></td> </tr> </tbody> </table> <b>Attachments</b><br><input type="checkbox"/> Form 41<br><input type="checkbox"/> Pre-Stability Tests<br><input type="checkbox"/> Flow Calibration<br><input type="checkbox"/> Calibration Adjustment<br><input type="checkbox"/> Post-Stability Tests<br><input type="checkbox"/> Other _____ | Simulator | Serial Number | 0.00 |  | Interferent |  | 0.05 |  | 0.08 |  | 0.20 |  |
|---|---------------|---------------|------------|------------|-------|--|-----|-----|-------|--|--|--|-------|--|--|--|-------|--|--|--|-------|--|--|--|-----------|-----|--|--|-----------|---------------|------------|------------|------|--|--|--|------|--|--|--|------|--|--|--|----------|-----|--|--|--|-----------|---------------|------|--|-------------|--|------|--|------|--|------|--|
| Simulator   | Serial Number | Lot Number    | Expiration |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.000   |               | N/A           | N/A        |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.040   |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.100   |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.200   |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.300   |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.080 DGS   | N/A           |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| Simulator   | Serial Number | Lot Number    | Expiration |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.05  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.08  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.20  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.08 DGS  | N/A           |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| Simulator   | Serial Number |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.00  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| Interferent   |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.05  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.08  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.20  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |  |           |               |      |  |             |  |      |  |      |  |      |  |

Notes/Suggested Service:  
Unable to operate, power supply problem, suggest sending to repair  
ATOC/DWS 12/14/17  
[Signature] Date 12/14/17  
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use pgm 1/29/18  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use



Alcohol Testing Program

**INSTRUMENT PROCESSING SHEET**

Agency Hulbert AFB S/N 80-004771  
Date In 6/2/17 Date Out 6/2/17  Ship  P/U  H/D  CMI  EE

| <b>Intake</b> Performed By <u>[Signature]</u><br><input type="checkbox"/> Registration<br><input checked="" type="checkbox"/> Annual<br><input type="checkbox"/> Return from CMI<br><input type="checkbox"/> Return from Enforcement Electronics<br><input type="checkbox"/> Other _____<br>Visual Inspection:<br><u>OK</u> Case <u>OK</u> Handle<br><u>OK</u> Dry Gas Holder <u>OK</u> Feet<br><u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs<br><u>OK</u> Screws tight <u>OK</u> Breath Hose<br>Other Equipment:<br><input type="checkbox"/> Power cord<br><input type="checkbox"/> Printer Cable<br><input type="checkbox"/> Other _____<br>Notes: _____<br>_____<br>_____<br>_____ | <b>Quality Checks</b> Performed By _____<br><input type="checkbox"/> Lab Temp °C _____<br><input type="checkbox"/> Breath Tube Screen<br><input type="checkbox"/> Replace External O-Rings<br><input type="checkbox"/> Instrument Set Up Verified<br><input type="checkbox"/> R-Value _____<br><input type="checkbox"/> Flow Verification (L/s)<br>Flow Column # _____<br>32mm _____ (.139 - .169)<br>36mm _____ (.156 - .190)<br>53mm _____ (.228 - .278)<br>103mm _____ (.447 - .547)<br><input type="checkbox"/> Barometric Pressure Check<br>Gauge ID # _____<br><input type="checkbox"/> Stability Checks<br><table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr><td>0.05</td><td></td><td></td></tr> <tr><td>0.08</td><td></td><td></td></tr> <tr><td>0.20</td><td></td><td></td></tr> <tr><td>0.08 DGS</td><td>N/A</td><td></td></tr> </tbody> </table> | Simulator | Serial # | Lot #/Exp | 0.05 |  |  | 0.08 |  |  | 0.20 |  |  | 0.08 DGS | N/A |  | <b>Flow Calibration</b> Performed By _____<br><input type="checkbox"/> Flow Calibration N/A<br><input type="checkbox"/> Flow Calibration Complete<br>Flow Column # _____<br><input type="checkbox"/> 5L/min - 17mm<br><input type="checkbox"/> 15L/min - 53mm<br><input type="checkbox"/> 30L/min - 103mm<br><input type="checkbox"/> R-Value _____<br><input type="checkbox"/> Post Calibration Verification (L/s)<br>Flow Column # _____<br>32mm _____ (.139 - .169)<br>36mm _____ (.156 - .190)<br>53mm _____ (.228 - .278)<br>103mm _____ (.447 - .547)<br><b>Maintenance</b> Performed By _____<br><input type="checkbox"/> Battery Replacement<br><input type="checkbox"/> Dry Gas Regulator Replacement<br><input type="checkbox"/> Breath Tube Replacement<br><input type="checkbox"/> Other _____<br><b>Quality Checks Cont.</b> Performed By _____<br><b>Simulator Temperatures °C</b><br>External Digital Therm. ID#: _____<br><input type="checkbox"/> 34°C +/- .2 Serial #: _____<br><input type="checkbox"/> 34°C +/- .2 Serial #: _____<br><input type="checkbox"/> 34°C +/- .2 Serial #: _____ |
|--|--|-----------|----------|-----------|------|--|--|------|--|--|------|--|--|----------|-----|--|--|
| Simulator  | Serial #   | Lot #/Exp |          |           |      |  |  |      |  |  |      |  |  |          |     |  |  |
| 0.05   |  |           |          |           |      |  |  |      |  |  |      |  |  |          |     |  |  |
| 0.08   |  |           |          |           |      |  |  |      |  |  |      |  |  |          |     |  |  |
| 0.20   |  |           |          |           |      |  |  |      |  |  |      |  |  |          |     |  |  |
| 0.08 DGS   | N/A  |           |          |           |      |  |  |      |  |  |      |  |  |          |     |  |  |

| <b>Calibration Adjustment</b> Performed By _____<br><input type="checkbox"/> Calibration Adjustment N/A<br><input type="checkbox"/> Calibration Adjustment Complete<br>Barometric Pressure Gauge _____ ID # _____<br><table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks<br><table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.05</td><td></td><td></td><td></td></tr> <tr><td>0.08</td><td></td><td></td><td></td></tr> <tr><td>0.20</td><td></td><td></td><td></td></tr> <tr><td>0.08 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> | Simulator     | Serial Number | Lot Number | Expiration | 0.000 |  | N/A | N/A | 0.040 |  |  |  | 0.100 |  |  |  | 0.200 |  |  |  | 0.300 |  |  |  | 0.080 DGS | N/A |  |  | Simulator | Serial Number | Lot Number | Expiration | 0.05 |  |  |  | 0.08 |  |  |  | 0.20 |  |  |  | 0.08 DGS | N/A |  |  | <b>Department Inspection</b> Performed By _____<br><input type="checkbox"/> Barometric Pressure _____ Gauge<br>ID# _____ Instrument<br>Mouth Alcohol Solution Lot # _____<br>Acetone Stock Solution Lot # _____<br><table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.00</td><td></td></tr> <tr><td>Interferent</td><td></td></tr> <tr><td>0.05</td><td></td></tr> <tr><td>0.08</td><td></td></tr> <tr><td>0.20</td><td></td></tr> </tbody> </table> <b>Attachments</b><br><input type="checkbox"/> Form 41<br><input type="checkbox"/> Pre-Stability Tests<br><input type="checkbox"/> Flow Calibration<br><input type="checkbox"/> Calibration Adjustment<br><input type="checkbox"/> Post-Stability Tests<br><input type="checkbox"/> Other _____ | Simulator | Serial Number | 0.00 |  | Interferent |  | 0.05 |  | 0.08 |  | 0.20 |  |
|---|---------------|---------------|------------|------------|-------|--|-----|-----|-------|--|--|--|-------|--|--|--|-------|--|--|--|-------|--|--|--|-----------|-----|--|--|-----------|---------------|------------|------------|------|--|--|--|------|--|--|--|------|--|--|--|----------|-----|--|--|---|-----------|---------------|------|--|-------------|--|------|--|------|--|------|--|
| Simulator   | Serial Number | Lot Number    | Expiration |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.000   |               | N/A           | N/A        |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.040   |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.100   |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.200   |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.300   |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.080 DGS   | N/A           |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| Simulator   | Serial Number | Lot Number    | Expiration |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.05  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.08  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.20  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.08 DGS  | N/A           |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| Simulator   | Serial Number |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.00  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| Interferent   |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.05  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.08  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |
| 0.20  |               |               |            |            |       |  |     |     |       |  |  |  |       |  |  |  |       |  |  |  |       |  |  |  |           |     |  |  |           |               |            |            |      |  |  |  |      |  |  |  |      |  |  |  |          |     |  |  |   |           |               |      |  |             |  |      |  |      |  |      |  |

|  |  |
|--|--|
| Notes/Suggested Service:<br><u>Unable to operate, power supply problem, suggest sending to repair</u><br><u>DA/OC/ASB 12/14/17</u><br><u>[Signature]</u> | <input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC<br><input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC<br><input checked="" type="checkbox"/> Return to/Place into Evidentiary Use<br><input type="checkbox"/> Remain Out of Evidentiary Use<br><input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use |
|--|--|

Quality Control Review \_\_\_\_\_ Date 12/14/17