



INSTRUMENT PROCESSING SHEET

Agency Okaloosa CountyS/N 80-005057Florida Department of
Law EnforcementDate In 12/5/17DI Completion Date 12/8/17☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	Performed By <u>DD</u>	Quality Checks	Performed By <u>DD</u>	Flow Calibration	Performed By <u>DD</u>																																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>154</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.165</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks		Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																									
Final Release Date FDLE DEC 11 2017 Alcohol Testing Program		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>62835</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG715202</u> <u>6/1/19</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>	0.080 DGS	N/A	<u>AG715202</u> <u>6/1/19</u>	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____	Temperature Checks Performed By <u>DD</u> ☒ Lab Temp °C <u>22.2</u> External Digital Therm. ID#: <u>500505</u> ☒ 34°C +/- .2 Serial #: <u>62835</u> ☒ 34°C +/- .2 Serial #: <u>DR1279</u> ☒ 34°C +/- .2 Serial #: <u>SD1025</u>																																										
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>																																																											
0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>																																																											
0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>																																																											
0.080 DGS	N/A	<u>AG715202</u> <u>6/1/19</u>																																																											
Calibration Adjustment Performed By <u>DD</u> Barometric Pressure Gauge <u>1011</u> ID # <u>30793</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>SD1016</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>SD1024</u></td> <td><u>16320</u></td> <td><u>10/21/18</u></td> </tr> <tr> <td>0.100</td> <td><u>62834</u></td> <td><u>17280</u></td> <td><u>9/11/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1011</u></td> <td><u>17090</u></td> <td><u>2/24/19</u></td> </tr> <tr> <td>0.300</td> <td><u>DR1275</u></td> <td><u>16410</u></td> <td><u>12/19/18</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>17817030A3</u></td> <td><u>8/5/19</u></td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>62835</u></td> <td><u>201707D</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>201707C</u></td> <td><u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG715202</u></td> <td><u>6/1/19</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>SD1016</u>	N/A	N/A	0.040	<u>SD1024</u>	<u>16320</u>	<u>10/21/18</u>	0.100	<u>62834</u>	<u>17280</u>	<u>9/11/19</u>	0.200	<u>SD1011</u>	<u>17090</u>	<u>2/24/19</u>	0.300	<u>DR1275</u>	<u>16410</u>	<u>12/19/18</u>	0.080 DGS	N/A	<u>17817030A3</u>	<u>8/5/19</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>62835</u>	<u>201707D</u>	<u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u>	<u>7/25/19</u>	0.200	<u>SD1025</u>	<u>201707C</u>	<u>7/24/19</u>	0.080 DGS	N/A	<u>AG715202</u>	<u>6/1/19</u>	Department Inspection Performed By <u>DD</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1010</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>62880</u></td> </tr> <tr> <td>Interferent</td> <td><u>62840</u></td> </tr> <tr> <td>0.050</td> <td><u>62835</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>62880</u>	Interferent	<u>62840</u>	0.050	<u>62835</u>	0.080	<u>DR1279</u>	0.200	<u>SD1025</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	<u>SD1016</u>	N/A	N/A																																																										
0.040	<u>SD1024</u>	<u>16320</u>	<u>10/21/18</u>																																																										
0.100	<u>62834</u>	<u>17280</u>	<u>9/11/19</u>																																																										
0.200	<u>SD1011</u>	<u>17090</u>	<u>2/24/19</u>																																																										
0.300	<u>DR1275</u>	<u>16410</u>	<u>12/19/18</u>																																																										
0.080 DGS	N/A	<u>17817030A3</u>	<u>8/5/19</u>																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	<u>62835</u>	<u>201707D</u>	<u>7/25/19</u>																																																										
0.080	<u>DR1279</u>	<u>201707E</u>	<u>7/25/19</u>																																																										
0.200	<u>SD1025</u>	<u>201707C</u>	<u>7/24/19</u>																																																										
0.080 DGS	N/A	<u>AG715202</u>	<u>6/1/19</u>																																																										
Simulator	Serial Number																																																												
0.000	<u>62880</u>																																																												
Interferent	<u>62840</u>																																																												
0.050	<u>62835</u>																																																												
0.080	<u>DR1279</u>																																																												
0.200	<u>SD1025</u>																																																												
Notes/Suggested Service: <u>performed calibration adjust to bring</u> <u>barometric readings closer to nominal</u> _____ _____ _____	☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use _____ _____																																																												
	Tech Review / Date <u>SP 12/11/17</u> Admin Review / Date <u>12/11/17</u>																																																												

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
12/08/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:00
Control Test	0.049	08:01
Air Blank	0.000	08:01
Control Test	0.049	08:02
Air Blank	0.000	08:03
Control Test	0.050	08:03
Air Blank	0.000	08:04
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	


Operator's Signature

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
12/08/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:11
Control Test	0.080	08:12
Air Blank	0.000	08:13
Control Test	0.080	08:13
Air Blank	0.000	08:14
Control Test	0.080	08:14
Air Blank	0.000	08:15
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
12/08/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:06
Control Test	0.198	08:07
Air Blank	0.000	08:07
Control Test	0.199	08:08
Air Blank	0.000	08:08
Control Test	0.198	08:09
Air Blank	0.000	08:09
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	


Operator's Signature

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
12/08/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:31
Control Test	0.081	08:32
Air Blank	0.000	08:32
Control Test	0.081	08:32
Air Blank	0.000	08:33
Control Test	0.081	08:33
Air Blank	0.000	08:34
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS


Operator's Signature

8P

12/11/17
SO

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
12/08/2017 08:46:35

Auto Calibration
Max Power Res Value = 30
Auto Range Res Value = 16

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12532, 9um Io = 12698

Channel 1

Sample	% Abs	(% Abs Ref)
Sample #1	0.0460	(-0.0230)
Sample #2	0.0280	(0.0270)
Sample #3	0.0420	(0.0450)
Sample #4	0.0670	(0.0520)
Avg % Abs	0.0457	(0.0413)
STD DEV	0.0198	(0.0129)
REL STD DEV	43.263	(31.202)

Channel 2

Sample	% Abs	(% Abs Ref)
Sample #1	0.1220	(-0.0200)
Sample #2	0.1010	(0.0060)
Sample #3	0.1230	(0.0000)
Sample #4	0.1390	(0.0100)
Avg % Abs	0.1210	(0.0053)
STD DEV	0.0191	(0.0050)
REL STD DEV	15.768	(94.373)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12521, 9um Io = 12695

Channel 1

Sample	% Abs	(% Abs Ref)
Sample #1	0.7740	(0.0000)
Sample #2	0.7190	(0.0330)
Sample #3	0.7710	(0.0270)
Sample #4	0.7420	(0.0400)
Avg % Abs	0.7440	(0.0333)
STD DEV	0.0261	(0.0065)
REL STD DEV	3.502	(19.519)

Channel 2

Sample	% Abs	(% Abs Ref)
Sample #1	1.4970	(-0.0150)
Sample #2	1.4940	(0.0100)
Sample #3	1.4940	(0.0120)
Sample #4	1.4930	(0.0030)
Avg % Abs	1.4937	(0.0083)
STD DEV	0.0006	(0.0047)
REL STD DEV	0.039	(56.710)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12513, 9um Io = 12692

Channel 1

Sample	% Abs	(% Abs Ref)
Sample #1	1.8160	(-0.0020)
Sample #2	1.8290	(-0.0020)
Sample #3	1.8020	(0.0120)
Sample #4	1.8410	(0.0050)
Avg % Abs	1.8240	(0.0050)
STD DEV	0.0200	(0.0070)
REL STD DEV	1.095	(140.000)

Channel 2

Sample	% Abs	(% Abs Ref)
Sample #1	3.5320	(-0.0010)
Sample #2	3.5570	(-0.0120)
Sample #3	3.5210	(0.0000)
Sample #4	3.5650	(-0.0040)
Avg % Abs	3.5477	(-0.0053)
STD DEV	0.0234	(0.0061)
REL STD DEV	0.661	(114.564)

Sol Value = 0.200 g/210L ***

Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12511, 9um Io = 12691

Channel 1

Sample	% Abs	(% Abs Ref)
Sample #1	3.5380	(-0.0060)
Sample #2	3.5300	(0.0200)
Sample #3	3.5540	(0.0340)
Sample #4	3.5380	(0.0340)
Avg % Abs	3.5407	(0.0293)
STD DEV	0.0122	(0.0081)
REL STD DEV	0.345	(27.555)

Channel 2

Sample	% Abs	(% Abs Ref)
Sample #1	6.7230	(-0.0050)
Sample #2	6.7970	(0.0150)
Sample #3	6.8340	(-0.0060)
Sample #4	6.8170	(0.0080)
Avg % Abs	6.8160	(0.0057)
STD DEV	0.0185	(0.0107)
REL STD DEV	0.272	(188.694)

Sol Value = 0.300 g/210L ***

Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12506, 9um Io = 12689

Channel 1

Sample	% Abs	(% Abs Ref)
Sample #1	5.1930	(-0.0040)
Sample #2	5.2490	(-0.0120)
Sample #3	5.2070	(0.0160)
Sample #4	5.2560	(-0.0030)
Avg % Abs	5.2373	(0.0003)
STD DEV	0.0265	(0.0143)
REL STD DEV	0.506	(4288.356)

Channel 2

Sample	% Abs	(% Abs Ref)
Sample #1	9.8270	(-0.0060)
Sample #2	9.9140	(-0.0110)
Sample #3	9.8950	(0.0030)
Sample #4	9.9160	(-0.0130)
Avg % Abs	9.9083	(-0.0070)
STD DEV	0.0116	(0.0087)
REL STD DEV	0.117	(124.540)

AUTO CAL DATA

Channel 1

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.046
Std Dev = 0.02 Rel Std Dev = 43.26
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.744
Std Dev = 0.03 Rel Std Dev = 3.50
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.824
Std Dev = 0.02 Rel Std Dev = 1.10
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.541
Std Dev = 0.01 Rel Std Dev = 0.35
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.237
Std Dev = 0.03 Rel Std Dev = 0.51
Zero Order Coef = -108.66
First Order Coef = 2651.12
Second Order Coef = 18.67
Standard Deviation = 22.231617

Channel 2

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.121
Std Dev = 0.02 Rel Std Dev = 15.77
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.494
Std Dev = 0.00 Rel Std Dev = 0.04
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.548
Std Dev = 0.02 Rel Std Dev = 0.66
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.816
Std Dev = 0.02 Rel Std Dev = 0.27
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.908
Std Dev = 0.01 Rel Std Dev = 0.12
Zero Order Coef = -147.92
First Order Coef = 1343.14
Second Order Coef = 11.40
Standard Deviation = 14.682058

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.039	0.0006
0.100	0.101	-0.0006
0.200	0.200	0.0002
0.300	0.300	-0.0001

88

12/11/17
52

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	0.0004
0.100	0.100	0.0000
0.200	0.200	-0.0003
0.300	0.300	0.0001

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1
 Sample #1 = 3593.00
 Sample #2 = 3590.00
 Sample #3 = 3537.00
 Sample #4 = 3642.00
 Average Result = 3589.6667
 STD DEV = 52.5008
 REL STD DEV = 1.463

 ***** CHANNEL 2
 Sample #1 = 3518.00
 Sample #2 = 3497.00
 Sample #3 = 3485.00
 Sample #4 = 3496.00
 Average Result = 3492.6667
 STD DEV = 6.6583
 REL STD DEV = 0.191

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1011
 3 um H2O Adjust (mg/l*10,000) = 220
 9 um H2O Adjust (mg/l*10,000) = 317
 **** AUTO CAL PASS

Post Calibration
 Adjust Stabilities

OKALOOSA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005057
 12/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:37
Control Test	0.049	09:38
Air Blank	0.000	09:38
Control Test	0.049	09:39
Air Blank	0.000	09:40
Control Test	0.050	09:40
Air Blank	0.000	09:41
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

P Murphy
 Operator's Signature

OKALOOSA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005057
 12/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:43
Control Test	0.081	09:43
Air Blank	0.000	09:44
Control Test	0.080	09:44
Air Blank	0.000	09:45
Control Test	0.080	09:46
Air Blank	0.000	09:46
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

P Murphy
 Operator's Signature

OKALOOSA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005057
 12/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:47
Control Test	0.200	09:48
Air Blank	0.000	09:49
Control Test	0.200	09:49
Air Blank	0.000	09:50
Control Test	0.200	09:51
Air Blank	0.000	09:51
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
 Operator's Signature

OKALOOSA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005057
 12/08/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:53
Control Test	0.080	09:54
Air Blank	0.000	09:54
Control Test	0.079	09:54
Air Blank	0.000	09:55
Control Test	0.079	09:55
Air Blank	0.000	09:56
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

P Murphy
 Operator's Signature

8P

12/11/17
 JO