



# INSTRUMENT PROCESSING SHEET

Agency Crestview PDS/N 80-005058Florida Department of  
Law EnforcementDate In 12/7/17DI Completion Date 12/8/17 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>JDB</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>PSJM</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>242</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.161</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.257</u> (.228 - .278) 103 mm <u>.512</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																												
<b>Final Release Date</b> <div style="text-align: center; padding: 10px;"> <b>FDLE</b>           DEC 11 2017   <b>Alcohol Testing Program</b> </div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>62835</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG715202</u> <u>6/1/19</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>	0.080 DGS	N/A	<u>AG715202</u> <u>6/1/19</u>	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>PSJM</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.2</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>62835</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1025</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>																																																												
0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>																																																												
0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>																																																												
0.080 DGS	N/A	<u>AG715202</u> <u>6/1/19</u>																																																												
<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> Performed By <u>PSJM</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1010</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>62880</u></td> </tr> <tr> <td>Interferent</td> <td><u>62840</u></td> </tr> <tr> <td>0.050</td> <td><u>62835</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>62880</u>	Interferent	<u>62840</u>	0.050	<u>62835</u>	0.080	<u>DR1279</u>	0.200	<u>SD1025</u>	<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input type="checkbox"/> Post-Stability Checks
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000		N/A	N/A																																																											
0.040																																																														
0.100																																																														
0.200																																																														
0.300																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050																																																														
0.080																																																														
0.200																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number																																																													
0.000	<u>62880</u>																																																													
Interferent	<u>62840</u>																																																													
0.050	<u>62835</u>																																																													
0.080	<u>DR1279</u>																																																													
0.200	<u>SD1025</u>																																																													
Notes/Suggested Service: _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between; margin-top: 10px;"> <div style="text-align: center;"> <u>SP 12/11/17</u>              Tech Review / Date           </div> <div style="text-align: center;"> <u>JJ [Signature] 12/11/17</u>              Admin Review / Date           </div> </div>																																																													

CRESTVIEW PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005058  
12/08/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:30
Control Test	0.049	09:31
Air Blank	0.000	09:32
Control Test	0.049	09:32
Air Blank	0.000	09:33
Control Test	0.049	09:33
Air Blank	0.000	09:34
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*P Murphy*  
-----  
Operator's Signature

CRESTVIEW PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005058  
12/08/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:35
Control Test	0.080	09:36
Air Blank	0.000	09:37
Control Test	0.080	09:37
Air Blank	0.000	09:38
Control Test	0.080	09:38
Air Blank	0.000	09:39
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*P Murphy*  
-----  
Operator's Signature

CRESTVIEW PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005058  
12/08/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:41
Control Test	0.198	09:41
Air Blank	0.000	09:42
Control Test	0.200	09:43
Air Blank	0.000	09:43
Control Test	0.199	09:44
Air Blank	0.000	09:44
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

*P Murphy*  
-----  
Operator's Signature

SP

CRESTVIEW PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005058  
12/08/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:51
Control Test	0.078	09:51
Air Blank	0.000	09:51
Control Test	0.080	09:52
Air Blank	0.000	09:52
Control Test	0.080	09:52
Air Blank	0.000	09:53
Control Test Stats		
Average	0.0793	
Std Dev	0.0012	
Rel Std Dev(%)	1.4555	

DGS

*P Murphy*  
-----  
Operator's Signature

12/11/17  
JA