



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Hernando County Sheriff's OfficeS/N 80-005249Date In 09/29/2017Date Out 10/04/2017☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>PSB</u> ☐ Registration ☒ Annual ☒ Return from CMI ☐ Return from Enforcement Electronics ☐ Other _____ Visual Inspection: ☒ Case ☒ Handle ☒ Dry Gas Holder ☒ Feet ☒ Keyboard/Plug ☒ Back/Plugs ☒ Screws tight ☒ Breath Hose Other Equipment: ☒ Power cord ☐ Printer Cable ☒ Other <u>Static Bag</u> Notes: _____ _____ _____ _____ _____		Quality Checks Performed By <u>PSB</u> ☒ Lab Temp °C <u>22.0</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>230</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 102</u> 32mm <u>0.160</u> (.139 - .169) 36mm <u>0.183</u> (.156 - .190) 53mm <u>0.250</u> (.228 - .278) 103mm <u>0.523</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3962</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.08</td> <td>SD1013</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.20</td> <td>DR3856</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG708807 03/29/2019</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.05	SD3962	201707D 07/25/2019	0.08	SD1013	201707E 07/25/2019	0.20	DR3856	201707C 07/24/2019	0.08 DGS	N/A	AG708807 03/29/2019	Flow Calibration Performed By _____ ☒ Flow Calibration N/A ☐ Flow Calibration Complete Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Quality Checks Cont. Performed By <u>PSB</u> Simulator Temperatures °C External Digital Therm. ID#: <u>300505</u> ☒ 34°C +- .2 Serial #: <u>SD3962</u> ☒ 34°C +- .2 Serial #: <u>SD1013</u> ☒ 34°C +- .2 Serial #: <u>DR3856</u>																																												
Simulator	Serial #	Lot #/Exp																																																													
0.05	SD3962	201707D 07/25/2019																																																													
0.08	SD1013	201707E 07/25/2019																																																													
0.20	DR3856	201707C 07/24/2019																																																													
0.08 DGS	N/A	AG708807 03/29/2019																																																													
Calibration Adjustment Performed By _____ ☒ Calibration Adjustment N/A ☐ Calibration Adjustment Complete Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>PSB</u> ☒ Barometric Pressure <u>1021</u> Gauge ID# <u>26932</u> <u>1020</u> Instrument Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD3962</td> </tr> <tr> <td>0.08</td> <td>SD1013</td> </tr> <tr> <td>0.20</td> <td>DR3856</td> </tr> </tbody> </table>		Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD3962	0.08	SD1013	0.20	DR3856
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number	Lot Number	Expiration																																																												
0.05																																																															
0.08																																																															
0.20																																																															
0.08 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.00	SD1019																																																														
Interferent	SD1021																																																														
0.05	SD3962																																																														
0.08	SD1013																																																														
0.20	DR3856																																																														
Notes/Suggested Service: <u>Please remember to change level 2 menu password</u> <u>to something unique. PSB</u> <u>QA/QC OK SP</u> <u>10/5/17</u>		Attachments ☒ Form 41 ☒ Pre-Stability Tests ☐ Flow Calibration ☐ Calibration Adjustment ☐ Post-Stability Tests ☐ Other _____																																																													
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																															

Quality Control Review

Date

Stability Checks # 80-005249 Hernando County S.O. 10/4/17 *Ans*

Ans

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
10/04/2017
Software: 8100.27

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
10/04/2017
Software: 8100.27

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
10/04/2017
Software: 8100.27

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
10/04/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:10
Control Test	0.081	13:10
Air Blank	0.000	13:11
Control Test	0.081	13:11
Air Blank	0.000	13:12
Control Test	0.080	13:12
Air Blank	0.000	13:12
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Test	g/210L	Time
Air Blank	0.000	13:05
Control Test	0.199	13:06
Air Blank	0.000	13:07
Control Test	0.199	13:07
Air Blank	0.000	13:08
Control Test	0.198	13:09
Air Blank	0.000	13:09
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

Test	g/210L	Time
Air Blank	0.000	13:00
Control Test	0.080	13:00
Air Blank	0.000	13:01
Control Test	0.082	13:01
Air Blank	0.000	13:02
Control Test	0.080	13:03
Air Blank	0.000	13:03
Control Test Stats		
Average	0.0807	
Std Dev	0.0012	
Rel Std Dev(%)	1.4314	

Test	g/210L	Time
Air Blank	0.000	12:55
Control Test	0.050	12:55
Air Blank	0.000	12:56
Control Test	0.049	12:57
Air Blank	0.000	12:57
Control Test	0.050	12:58
Air Blank	0.000	12:58
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

SP

Ans

Operator's Signature

Ans

Operator's Signature

Ans

Operator's Signature

Ans

Operator's Signature

10/5/17
JD



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Hernando County S/N 80-005249
 Date In 1/25/17 Date Out 3/7/17 ☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake Performed By <u>DE</u> ☐ Registration ☒ Annual ☐ Return from CMI ☐ Return from Enforcement Electronics ☐ Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: ☒ Power cord ☐ Printer Cable ☒ Other <u>Static Bag</u> Notes: <u>All indicated amb fail issues. DE</u>	Quality Checks Performed By _____ ☐ Breath Tube Screen ☐ Replace O-Rings ☐ Instrument Set Up Verified ☐ R-Value _____ ☐ Flow Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) ☐ Barometric Pressure Check Gauge ID # _____ ☐ Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05			0.08			0.20			0.08 DGS	N/A		Flow Calibration Performed By _____ ☐ Flow Calibration N/A ☐ Flow Calibration Complete Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Suggested Service _____ _____ _____
Simulator	Serial #	Lot #/Exp															
0.05																	
0.08																	
0.20																	
0.08 DGS	N/A																

Optical Bench Calibration Performed By _____ ☐ Optical Bench Calibration N/A ☐ Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By _____ ☐ Barometric Pressure _____ Gauge ID# _____ Instrument Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.05</td> <td></td> </tr> <tr> <td>0.08</td> <td></td> </tr> <tr> <td>0.20</td> <td></td> </tr> </tbody> </table> Attachments ☐ Form 41 ☐ Pre-Stability Tests ☐ Flow Calibration ☐ Optical Bench Cal ☐ Post-Stability Tests ☐ Other _____	Simulator	Serial Number	0.00		Interferent		0.05		0.08		0.20	
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00																																																													
Interferent																																																													
0.05																																																													
0.08																																																													
0.20																																																													

Notes: <u>Repeatedly tested for amb/purge issues. Nothing found. Sending to CMI as a precaution. DE</u>	☐ Instrument Complies with Chapter 11D-8, FAC <u>Not Determined</u> ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use <u>10/5/17 JD</u>
--	--

Quality Control Review

Date

Revised November 2012

Stability Checks #80-005249 Hernando County S.O. 2/2/17

No Indication of Ambient Fails. DMS

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
02/02/2017
Software: 8100.27

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
02/02/2017
Software: 8100.27

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
02/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:37
Control Test	0.080	16:37
Air Blank	0.000	16:37
Control Test	0.081	16:38
Air Blank	0.000	16:38
Control Test	0.080	16:39
Air Blank	0.000	16:39
Control Test	0.082	16:39
Air Blank	0.000	16:40
Control Test	0.081	16:40
Air Blank	0.000	16:41
Control Test	0.081	16:41
Air Blank	0.000	16:42
Control Test	0.081	16:42
Air Blank	0.000	16:42
Control Test	0.081	16:43
Air Blank	0.000	16:43
Control Test	0.081	16:43
Air Blank	0.000	16:44
Control Test	0.081	16:44
Air Blank	0.000	16:45
Control Test	0.082	16:45
Air Blank	0.000	16:46
Control Test	0.081	16:46
Air Blank	0.000	16:46
Control Test	0.080	16:47
Air Blank	0.000	16:47
Control Test	0.081	16:48
Air Blank	0.000	16:48
Control Test	0.081	16:48
Air Blank	0.000	16:49
Control Test	0.082	16:49
Air Blank	0.000	16:50
Control Test	0.081	16:50
Air Blank	0.000	16:51
Control Test	0.081	16:51
Air Blank	0.000	16:51
Control Test	0.081	16:52
Air Blank	0.000	16:52
Control Test	0.081	16:53
Air Blank	0.000	16:53
Control Test Stats		
Average	0.0810	
Std Dev	0.0006	
Rel Std Dev(%)	0.6938	

Test	g/210L	Time
Air Blank	0.000	16:19
Control Test	0.081	16:19
Air Blank	0.000	16:19
Control Test	0.081	16:20
Air Blank	0.000	16:20
Control Test	0.081	16:21
Air Blank	0.000	16:21
Control Test	0.081	16:21
Air Blank	0.000	16:22
Control Test	0.080	16:22
Air Blank	0.000	16:23
Control Test	0.081	16:23
Air Blank	0.000	16:24
Control Test	0.081	16:24
Air Blank	0.000	16:24
Control Test	0.080	16:25
Air Blank	0.000	16:25
Control Test	0.081	16:26
Air Blank	0.000	16:26
Control Test	0.081	16:26
Air Blank	0.000	16:27
Control Test	0.081	16:27
Air Blank	0.000	16:28
Control Test	0.081	16:28
Air Blank	0.000	16:29
Control Test	0.082	16:29
Air Blank	0.000	16:29
Control Test	0.081	16:30
Air Blank	0.000	16:30
Control Test	0.081	16:31
Air Blank	0.000	16:31
Control Test	0.081	16:31
Air Blank	0.000	16:32
Control Test	0.081	16:32
Air Blank	0.000	16:33
Control Test	0.081	16:33
Air Blank	0.000	16:34
Control Test	0.081	16:34
Air Blank	0.000	16:34
Control Test	0.081	16:35
Air Blank	0.000	16:35
Control Test Stats		
Average	0.0809	
Std Dev	0.0004	
Rel Std Dev(%)	0.4868	

Test	g/210L	Time
Air Blank	0.000	16:01
Control Test	0.081	16:01
Air Blank	0.000	16:02
Control Test	0.081	16:02
Air Blank	0.000	16:03
Control Test	0.081	16:03
Air Blank	0.000	16:03
Control Test	0.082	16:04
Air Blank	0.000	16:04
Control Test	0.082	16:05
Air Blank	0.000	16:05
Control Test	0.082	16:05
Air Blank	0.000	16:06
Control Test	0.081	16:06
Air Blank	0.000	16:07
Control Test	0.081	16:07
Air Blank	0.000	16:08
Control Test	0.083	16:08
Air Blank	0.000	16:08
Control Test	0.080	16:09
Air Blank	0.000	16:09
Control Test	0.081	16:09
Air Blank	0.000	16:10
Control Test	0.081	16:10
Air Blank	0.000	16:11
Control Test	0.082	16:11
Air Blank	0.000	16:12
Control Test	0.080	16:12
Air Blank	0.000	16:12
Control Test	0.081	16:13
Air Blank	0.000	16:13
Control Test	0.081	16:14
Air Blank	0.000	16:14
Control Test	0.081	16:14
Air Blank	0.000	16:15
Control Test	0.081	16:15
Air Blank	0.000	16:16
Control Test	0.082	16:16
Air Blank	0.000	16:17
Control Test	0.080	16:17
Air Blank	0.000	16:17
Control Test Stats		
Average	0.0812	
Std Dev	0.0008	
Rel Std Dev(%)	0.9455	

Operator's Signature

Operator's Signature

Operator's Signature

10/5/17
JB

Stability Checks # 80-005249 Hernando County S.O. 2/15/17 *RMS*

No indication of Ambient Fails

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
02/15/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:49
Control Test	0.194	09:50
Air Blank	0.000	09:50
Control Test	0.195	09:51
Air Blank	0.000	09:51
Control Test	0.196	09:52
Air Blank	0.000	09:53
Control Test	0.197	09:53
Air Blank	0.000	09:54
Control Test	0.196	09:54
Air Blank	0.000	09:55
Control Test	0.197	09:56
Air Blank	0.000	09:56
Control Test	0.197	09:57
Air Blank	0.000	09:57
Control Test	0.197	09:58
Air Blank	0.000	09:59
Control Test	0.196	09:59
Air Blank	0.000	10:00
Control Test	0.198	10:01
Air Blank	0.000	10:01
Control Test	0.197	10:02
Air Blank	0.000	10:02
Control Test	0.197	10:03
Air Blank	0.000	10:04
Control Test	0.197	10:04
Air Blank	0.000	10:05
Control Test	0.196	10:06
Air Blank	0.000	10:06
Control Test	0.197	10:07
Air Blank	0.000	10:07
Control Test	0.197	10:08
Air Blank	0.000	10:09
Control Test	0.197	10:09
Air Blank	0.000	10:10
Control Test	0.197	10:11
Air Blank	0.000	10:11
Control Test	0.197	10:12
Air Blank	0.000	10:12
Control Test	0.196	10:13
Air Blank	0.000	10:14
Control Test Stats		
Average	0.1966	
Std Dev	0.0009	
Rel Std Dev(%)	0.4513	

RMS
Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
02/15/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:15
Control Test	0.198	10:15
Air Blank	0.000	10:16
Control Test	0.198	10:17
Air Blank	0.000	10:17
Control Test	0.197	10:18
Air Blank	0.000	10:19
Control Test	0.197	10:19
Air Blank	0.000	10:20
Control Test	0.197	10:21
Air Blank	0.000	10:21
Control Test	0.197	10:22
Air Blank	0.000	10:22
Control Test	0.197	10:22
Air Blank	0.000	10:23
Control Test	0.197	10:24
Air Blank	0.000	10:24
Control Test	0.196	10:25
Air Blank	0.000	10:26
Control Test	0.196	10:26
Air Blank	0.000	10:27
Control Test	0.196	10:27
Air Blank	0.000	10:27
Control Test	0.198	10:28
Air Blank	0.000	10:29
Control Test	0.197	10:29
Air Blank	0.000	10:30
Control Test	0.197	10:31
Air Blank	0.000	10:31
Control Test	0.197	10:32
Air Blank	0.000	10:33
Control Test	0.197	10:33
Air Blank	0.000	10:34
Control Test	0.197	10:34
Air Blank	0.000	10:35
Control Test	0.197	10:36
Air Blank	0.000	10:36
Control Test	0.197	10:37
Air Blank	0.000	10:38
Control Test	0.196	10:38
Air Blank	0.000	10:39
Control Test	0.197	10:40
Air Blank	0.000	10:40
Control Test Stats		
Average	0.1970	
Std Dev	0.0006	
Rel Std Dev(%)	0.2853	

RMS
Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
02/15/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:41
Control Test	0.198	10:42
Air Blank	0.000	10:42
Control Test	0.197	10:43
Air Blank	0.000	10:43
Control Test	0.197	10:44
Air Blank	0.000	10:45
Control Test	0.197	10:45
Air Blank	0.000	10:46
Control Test	0.197	10:47
Air Blank	0.000	10:47
Control Test	0.197	10:48
Air Blank	0.000	10:48
Control Test	0.197	10:49
Air Blank	0.000	10:50
Control Test	0.197	10:50
Air Blank	0.000	10:51
Control Test	0.197	10:52
Air Blank	0.000	10:52
Control Test	0.197	10:53
Air Blank	0.000	10:54
Control Test	0.197	10:54
Air Blank	0.000	10:55
Control Test	0.196	10:55
Air Blank	0.000	10:56
Control Test	0.197	10:57
Air Blank	0.000	10:57
Control Test	0.197	10:58
Air Blank	0.000	10:59
Control Test	0.197	10:59
Air Blank	0.000	11:00
Control Test	0.197	11:01
Air Blank	0.000	11:01
Control Test	0.197	11:02
Air Blank	0.000	11:02
Control Test	0.197	11:03
Air Blank	0.000	11:04
Control Test	0.197	11:04
Air Blank	0.000	11:05
Control Test	0.197	11:06
Air Blank	0.000	11:06
Control Test Stats		
Average	0.1970	
Std Dev	0.0003	
Rel Std Dev(%)	0.1647	

RMS
Operator's Signature

10/5/17
JD

Stability Checks #80-005249 Hernando County S.O. 2/15/17 *RMB*

No Indication of Ambient Issues

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
02/15/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:26
Control Test	0.197	11:26
Air Blank	0.000	11:27
Control Test	0.199	11:28
Air Blank	0.000	11:28
Control Test	0.198	11:29
Air Blank	0.000	11:29
Control Test	0.197	11:30
Air Blank	0.000	11:31
Control Test	0.197	11:31
Air Blank	0.000	11:32
Control Test	0.197	11:33
Air Blank	0.000	11:33
Control Test	0.197	11:34
Air Blank	0.000	11:34
Control Test	0.198	11:35
Air Blank	0.000	11:36
Control Test	0.197	11:36
Air Blank	0.000	11:37
Control Test	0.197	11:38
Air Blank	0.000	11:38
Control Test	0.197	11:39
Air Blank	0.000	11:40
Control Test	0.196	11:40
Air Blank	0.000	11:41
Control Test	0.196	11:41
Air Blank	0.000	11:42
Control Test	0.196	11:43
Air Blank	0.000	11:43
Control Test	0.195	11:44
Air Blank	0.000	11:45
Control Test	0.196	11:45
Air Blank	0.000	11:46
Control Test	0.195	11:47
Air Blank	0.000	11:47
Control Test	0.195	11:48
Air Blank	0.000	11:48
Control Test	0.196	11:49
Air Blank	0.000	11:50
Control Test	0.196	11:50
Air Blank	0.000	11:51
Control Test Stats		
Average	0.1966	
Std Dev	0.0010	
Rel Std Dev(%)	0.5322	

RMB
Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
02/15/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:52
Control Test	0.082	11:53
Air Blank	0.000	11:53
Control Test	0.082	11:53
Air Blank	0.000	11:54
Control Test	0.083	11:54
Air Blank	0.000	11:55
Control Test	0.083	11:55
Air Blank	0.000	11:56
Control Test	0.083	11:56
Air Blank	0.000	11:56
Control Test	0.084	11:57
Air Blank	0.000	11:57
Control Test	0.083	11:58
Air Blank	0.000	11:58
Control Test	0.083	11:58
Air Blank	0.000	11:59
Control Test	0.083	11:59
Air Blank	0.000	12:00
Control Test	0.083	12:00
Air Blank	0.000	12:01
Control Test	0.084	12:01
Air Blank	0.000	12:01
Control Test	0.083	12:02
Air Blank	0.000	12:02
Control Test	0.083	12:02
Air Blank	0.000	12:03
Control Test	0.083	12:03
Air Blank	0.000	12:04
Control Test	0.083	12:04
Air Blank	0.000	12:05
Control Test	0.083	12:05
Air Blank	0.000	12:05
Control Test	0.084	12:06
Air Blank	0.000	12:06
Control Test	0.083	12:07
Air Blank	0.000	12:07
Control Test	0.083	12:07
Air Blank	0.000	12:08
Control Test	0.083	12:08
Air Blank	0.000	12:09
Control Test Stats		
Average	0.0830	
Std Dev	0.0005	
Rel Std Dev(%)	0.6146	

RMB
Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
02/15/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:17
Control Test	0.083	12:17
Air Blank	0.000	12:18
Control Test	0.084	12:18
Air Blank	0.000	12:19
Control Test	0.083	12:19
Air Blank	0.000	12:19
Control Test	0.083	12:20
Air Blank	0.000	12:20
Control Test	0.083	12:21
Air Blank	0.000	12:21
Control Test	0.083	12:21
Air Blank	0.000	12:22
Control Test	0.083	12:22
Air Blank	0.000	12:23
Control Test	0.084	12:23
Air Blank	0.000	12:24
Control Test	0.083	12:24
Air Blank	0.000	12:24
Control Test	0.083	12:25
Air Blank	0.000	12:25
Control Test	0.083	12:26
Air Blank	0.000	12:26
Control Test	0.082	12:26
Air Blank	0.000	12:27
Control Test	0.082	12:27
Air Blank	0.000	12:28
Control Test	0.081	12:28
Air Blank	0.000	12:29
Control Test	0.080	12:29
Air Blank	0.000	12:29
Control Test	0.080	12:30
Air Blank	0.000	12:30
Control Test	0.079	12:31
Air Blank	0.000	12:31
Control Test	0.076	12:32
Air Blank	0.000	12:32
Control Test	0.000	12:32
Air Blank	0.000	12:33
Control Test	0.000	12:33
Air Blank	0.000	12:34
Control Test Stats		
Average	0.0737	
Std Dev	0.0253	
Rel Std Dev(%)	34.2995	

RMB
Operator's Signature

10/5/17
ee

Stability Checks #80-005249 Hernando County S.O. 2/15/17 RWS
No indication of Ambient issues

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
02/15/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:35
Control Test	0.081	12:36
Air Blank	0.000	12:36
Control Test	0.081	12:36
Air Blank	0.000	12:37
Control Test	0.081	12:37
Air Blank	0.000	12:38
Control Test	0.081	12:38
Air Blank	0.000	12:39
Control Test	0.081	12:39
Air Blank	0.000	12:39
Control Test	0.081	12:40
Air Blank	0.000	12:40
Control Test	0.080	12:40
Air Blank	0.000	12:41
Control Test	0.080	12:41
Air Blank	0.000	12:42
Control Test	0.081	12:42
Air Blank	0.000	12:43
Control Test	0.081	12:43
Air Blank	0.000	12:43
Control Test	0.082	12:44
Air Blank	0.000	12:44
Control Test	0.080	12:45
Air Blank	0.000	12:45
Control Test	0.081	12:45
Air Blank	0.000	12:46
Control Test	0.081	12:46
Air Blank	0.000	12:47
Control Test	0.080	12:47
Air Blank	0.000	12:48
Control Test	0.081	12:48
Air Blank	0.000	12:48
Control Test	0.080	12:49
Air Blank	0.000	12:49
Control Test	0.080	12:50
Air Blank	0.000	12:50
Control Test	0.081	12:50
Air Blank	0.000	12:51
Control Test	0.081	12:51
Air Blank	0.000	12:52
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.6813	

RWS

Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
02/15/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:56
Control Test	0.193	12:57
Air Blank	0.000	12:57
Control Test	0.195	12:58
Air Blank	0.000	12:59
Control Test	0.197	12:59
Air Blank	0.000	13:00
Control Test	0.196	13:01
Air Blank	0.000	13:01
Control Test	0.196	13:02
Air Blank	0.000	13:03
Control Test	0.196	13:03
Air Blank	0.000	13:04
Control Test	0.195	13:04
Air Blank	0.000	13:05
Control Test	0.195	13:06
Air Blank	0.000	13:06
Control Test	0.195	13:07
Air Blank	0.000	13:08
Control Test	0.195	13:08
Air Blank	0.000	13:09
Control Test	0.195	13:10
Air Blank	0.000	13:10
Control Test	0.196	13:11
Air Blank	0.000	13:11
Control Test	0.195	13:12
Air Blank	0.000	13:13
Control Test	0.196	13:13
Air Blank	0.000	13:14
Control Test	0.196	13:15
Air Blank	0.000	13:15
Control Test	0.195	13:16
Air Blank	0.000	13:16
Control Test	0.196	13:17
Air Blank	0.000	13:18
Control Test	0.195	13:18
Air Blank	0.000	13:19
Control Test	0.195	13:20
Air Blank	0.000	13:20
Control Test	0.195	13:21
Air Blank	0.000	13:22
Control Test Stats		
Average	0.1954	
Std Dev	0.0008	
Rel Std Dev(%)	0.4160	

RWS

Operator's Signature

10/5/17
20