

INSTRUMENT PROCESSING SHEET

Agency Miami Dade Police Department S/N 80-005548
 Date In 01/30/2017 Date Out 01/30/2017 Ship P/U H/D CMI EE

Intake Performed By DELL

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other Replace Breath Tube

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other ANTI STATIC BAG

Notes: _____

Quality Checks Performed By DELL

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 194
 Flow Verification (L/s)
 Flow Column # ATP101
 32mm 160 (.139 - .169)
 36mm 179 (.156 - .190)
 53mm 246 (.228 - .278)
 103mm 515 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28663

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>SD 3967</u>	<u>201603D</u> <u>03/08/2018</u>
0.08	<u>SD 3968</u>	<u>201601F</u> <u>01/26/2018</u>
0.20	<u>SD 3969</u>	<u>201604C</u> <u>04/05/2018</u>
0.08 DGS	<u>N/A</u>	<u>A6605201</u> <u>02/22/2018</u>

Flow Calibration Performed By DELL

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By DELL

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 FEB 10 2017
 FDLL
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		<u>N/A</u>	<u>N/A</u>
0.040			
0.100			
0.200			
0.400			
0.080 DGS	<u>N/A</u>		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	<u>N/A</u>		

Department Inspection Performed By DELL

Barometric Pressure 1021 Gauge
 ID# 68639 1023 Instrument

Mouth Alcohol Solution Lot # 2016A
 Acetone Stock Solution Lot # 2016B

Simulator	Serial Number
0.00	<u>SD 3965</u>
Interferent	<u>SD 3966</u>
0.05	<u>SD 3967</u>
0.08	<u>SD 3968</u>
0.20	<u>SD 3969</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: E-MAILED APPROVED
01/30/2017
DD/OC OK GJM 2/8/2017

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Philbeck
 Quality Control Review

2/10/17
 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-005548	Miami Dade Police Department	01/30/2017	<i>WLL</i>

AK

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005548
01/30/2017
Software: 8100.27

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005548
01/30/2017
Software: 8100.27

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005548
01/30/2017
Software: 8100.27

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005548
01/30/2017
Software: 8100.27

Test g/210L Time

Air Blank 0.000 10:13
Control Test 0.049 10:13
Air Blank 0.000 10:14
Control Test 0.049 10:15
Air Blank 0.000 10:15
Control Test 0.049 10:16
Air Blank 0.000 10:16
Control Test Stats
Average 0.0490
Std Dev 0.0000
Rel Std Dev(%) 0.0000

Test g/210L Time

Air Blank 0.000 10:18
Control Test 0.079 10:19
Air Blank 0.000 10:20
Control Test 0.080 10:20
Air Blank 0.000 10:21
Control Test 0.080 10:22
Air Blank 0.000 10:22
Control Test Stats
Average 0.0797
Std Dev 0.0006
Rel Std Dev(%) 0.7247

Test g/210L Time

Air Blank 0.000 10:24
Control Test 0.196 10:25
Air Blank 0.000 10:26
Control Test 0.197 10:26
Air Blank 0.000 10:27
Control Test 0.197 10:28
Air Blank 0.000 10:28
Control Test Stats
Average 0.1967
Std Dev 0.0006
Rel Std Dev(%) 0.2936

Test g/210L Time

Air Blank 0.000 10:30
Control Test 0.080 10:30
Air Blank 0.000 10:31
Control Test 0.081 10:31
Air Blank 0.000 10:32
Control Test 0.080 10:32
Air Blank 0.000 10:33
Control Test Stats
Average 0.0803
Std Dev 0.0006
Rel Std Dev(%) 0.7187

DWA
Operator's Signature

DWA
Operator's Signature

WLL
Operator's Signature

WLL
Operator's Signature

Open