



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Fair Myers PD

S/N 80-005655

Date In 10/4/17

Date Out 10/9/17

Ship P/U H/D CMI EE

Intake Performed By JD

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By SP

Lab Temp °C 22.1
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 199
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 144 (.139 - .169)
 36mm 160 (.156 - .190)
 53mm 230 (.228 - .278)
 103mm 503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3962	201707D 7-25-19
0.08	SD1013	201707E 7-25-19
0.20	DR3856	201707C 7-24-19
0.08 DGS	N/A	AG020604 9-22-18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # 103mm
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By SP

Simulator Temperatures °C
 External Digital Therm. ID#: 30504
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD1013
 34°C +/- .2 Serial #: DR3856

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure 1012 Gauge
 ID# 26932 1012 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD3962
0.08	SD1013
0.20	DR3856

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:
PA/PC OK GSDM 10/9/2017

JJ Phelan 10/9/17
 Quality Control Review Date

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

STABILITY CHECKS # 86-005655

FORT MYERS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005655
10/09/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:56
Control Test	0.050	09:57
Air Blank	0.000	09:58
Control Test	0.050	09:58
Air Blank	0.000	09:59
Control Test	0.050	10:00
Air Blank	0.000	10:00
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

FORT MYERS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005655
10/09/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:49
Control Test	0.082	09:49
Air Blank	0.000	09:50
Control Test	0.081	09:51
Air Blank	0.000	09:51
Control Test	0.082	09:52
Air Blank	0.000	09:52
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

SP

Operator's Signature

FORT MYERS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005655
10/09/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:41
Control Test	0.200	09:42
Air Blank	0.000	09:42
Control Test	0.200	09:43
Air Blank	0.000	09:44
Control Test	0.200	09:44
Air Blank	0.000	09:44
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

FORT MYERS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005655
10/09/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:36
Control Test	0.079	09:36
Air Blank	0.000	09:37
Control Test	0.079	09:37
Air Blank	0.000	09:37
Control Test	0.079	09:38
Air Blank	0.000	09:38
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DAS

SP

SP

Operator's Signature

10/9/17
JD