

INSTRUMENT PROCESSING SHEET

Agency Palm Beach County SO S/N 80-006028

Date In 3/6/17 Date Out 3/6/17 3/8/17 Ship P/U H/D CMI EE

- Intake Performed By SP
- Registration
 - Annual
 - Return from CMI
 - Return from Enforcement
 - Electronics
 - Other _____

- Visual Inspection:
- OK Case OK Handle
 - OK Dry Gas Holder OK Feet
 - OK Keyboard/Plug OK Back/Plugs
 - OK Screws tight OK Breath Hose

Other Equipment:

- Power cord
- Printer Cable
- Other _____

Notes: _____

- Quality Checks Performed By SP
- Breath Tube Screen
 - Replace O-Rings
 - Instrument Set Up Verified
 - R-Value 214
 - Flow Verification (L/s)

Flow Column # ATP101

32mm	<u>.164</u>	(.139 - .169)
36mm	<u>.179</u>	(.156 - .190)
53mm	<u>.238</u>	(.228 - .278)
103mm	<u>.484</u>	(.447 - .547)

- Barometric Pressure Check
- Gauge ID # 68639

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>SD1014</u>	<u>201603D</u> <u>3-8-18</u>
0.08	<u>SD1015</u>	<u>201601F</u> <u>1-26-18</u>
0.20	<u>SD1017</u>	<u>201604C</u> <u>4-5-18</u>
0.08 DGS	<u>N/A</u>	<u>AG605301</u> <u>2-22-18</u>

- Flow Calibration Performed By SP
- Flow Calibration N/A
 - Flow Calibration Complete
 - Flow Column # _____
 - 5L/min - 17mm
 - 15L/min - 53mm
 - 30L/min - 103mm
 - R-Value _____
 - Post Calibration Verification (L/s)
 - Flow Column # _____
 - 32mm _____ (.139 - .169)
 - 36mm _____ (.156 - .190)
 - 53mm _____ (.228 - .278)
 - 103mm _____ (.447 - .547)

- Maintenance Performed By _____
- Battery Replacement
 - Dry Gas Regulator Replacement
 - Breath Tube Replacement
 - Other _____

Suggested Service _____

RECEIVED
MAR 08 2017
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By SP

- Optical Bench Calibration N/A
- Optical Bench Calibration Complete
- Barometric Pressure Gauge 1025 ID # 26932

Simulator	Serial Number	Lot Number	Expiration
0.000	<u>SD1016</u>	<u>N/A</u>	<u>N/A</u>
0.040	<u>SD1024</u>	<u>16101</u>	<u>2-2-18</u>
0.100	<u>SD1022</u>	<u>16001</u>	<u>5-8-18</u>
0.200	<u>DR3855</u>	<u>16103</u>	<u>6-14-18</u>
0.400	<u>G2407</u>	<u>16102</u>	<u>3-22-18</u>
0.080 DGS	<u>N/A</u>	<u>15615080A2</u>	<u>7-5-17</u>

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05	<u>G11739</u>	<u>201603D</u>	<u>3-8-18</u>
0.08	<u>G8149</u>	<u>201601F</u>	<u>1-26-18</u>
0.20	<u>G11621</u>	<u>201604C</u>	<u>4-5-18</u>
0.08 DGS	<u>N/A</u>	<u>AG626605</u>	<u>9-22-18</u>

Notes: **E-MAILED**

QA/QC OK DEER

PERFORMED OPTICAL BENCH CAL TO

BRING VALUES CLOSER TO NOMINAL SP

QC: ADAS

Brett Kvildand

Department Inspection Performed By SP

- Barometric Pressure 1025 Gauge -1023
- ID# 68639 1024 Instrument -1025
- 28427
- Mouth Alcohol Solution Lot # 2016-A - 2016-A
- Acetone Stock Solution Lot # 2016-B - 2016-B

Simulator	Serial Number
0.00	<u>G8147</u> / <u>G2880</u>
Interferent	<u>G12100</u> / <u>G2834</u>
0.05	<u>SD1014</u> / <u>G3709</u>
0.08	<u>SD1015</u> / <u>SD3964</u>
0.20	<u>SD1017</u> / <u>DR3856</u>

Attachments

- Form 41 x2
- Pre-Stability Tests
- Flow Calibration
- Optical Bench Cal
- Post-Stability Tests
- Other _____

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC
- Return to/Place into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

3/8/17 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006028	Palm Beach County SO	3/16/17	SP

KR

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p>PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006028 03/06/2017 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:20 Control Test 0.051 09:20 Air Blank 0.000 09:21 Control Test 0.052 09:22 Air Blank 0.000 09:22 Control Test 0.052 09:23 Air Blank 0.000 09:23 Control Test Stats Average 0.0517 Std Dev 0.0006 Rel Std Dev(%) 1.1175</p> <p>Operator's Signature <i>SP</i></p>	<p>PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006028 03/06/2017 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:14 Control Test 0.082 09:15 Air Blank 0.000 09:16 Control Test 0.083 09:16 Air Blank 0.000 09:17 Control Test 0.083 09:18 Air Blank 0.000 09:18 Control Test Stats Average 0.0827 Std Dev 0.0006 Rel Std Dev(%) 0.6984</p> <p>Operator's Signature <i>SP</i></p>	<p>PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006028 03/06/2017 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:05 Control Test 0.199 09:06 Air Blank 0.000 09:07 Control Test 0.202 09:07 Air Blank 0.000 09:08 Control Test 0.202 09:08 Air Blank 0.000 09:09 Control Test Stats Average 0.2010 Std Dev 0.0017 Rel Std Dev(%) 0.8617</p> <p>Operator's Signature <i>SP</i></p>	<p>PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006028 03/06/2017 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:11 Control Test 0.082 09:11 Air Blank 0.000 09:12 Control Test 0.082 09:12 Air Blank 0.000 09:13 Control Test 0.082 09:13 Air Blank 0.000 09:14 Control Test Stats Average 0.0820 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <i>SP</i></p> <p>DGS <i>DGS</i></p>

Auto Range Res Value = 20

Soil Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum 10 = 12624, Sum 10 = 12776

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1020 (-0.0150)
Sample #2 = 0.1010 (-0.0220)
Sample #3 = 0.0660 (-0.0600)
Sample #4 = 0.0910 (-0.0910)
Avg % Abs = 0.0860 (-0.0577)
STD DEU = 0.0180 (-0.0346)
REL STD DEU = 20.963 (59.929)

Soil Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum 10 = 12583, Sum 10 = 12755

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.7860 (-0.0140)
Sample #2 = 1.8090 (-0.0230)
Sample #3 = 1.7980 (-0.0530)
Sample #4 = 1.7550 (-0.0800)
Avg % Abs = 1.7873 (-0.0520)
STD DEU = 0.0285 (-0.0285)
REL STD DEU = 1.597 (54.833)

Soil Value = 0.400 g/210L ***
Fit value = 1.9048 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum 10 = 12330, Sum 10 = 12724

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.6210 (-0.0000)
Sample #2 = 6.5240 (-0.0400)
Sample #3 = 6.6210 (-0.0660)
Sample #4 = 6.6670 (-0.0440)
Avg % Abs = 6.6373 (-0.0500)
STD DEU = 0.0257 (-0.0140)
REL STD DEU = 0.388 (28.000)

Channel 2 >>>>
Soil Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.143
Std Dev = 0.00 Rel Std Dev = 3.15
Soil Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.365
Std Dev = 0.04 Rel Std Dev = 2.97
Soil Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.481
Std Dev = 0.01 Rel Std Dev = 0.24
Soil Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.678
Std Dev = 0.01 Rel Std Dev = 0.22
Soil Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 12.554
Std Dev = 0.02 Rel Std Dev = 0.16
Zero Order Coef = -104.92
First Order Coef = 1357.33
Second Order Coef = 13.34
Standard Deviation = 82.317741

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.002	-0.0017
0.040	0.037	0.0027
0.100	0.101	-0.0009
0.200	0.200	-0.0002
0.400	0.400	0.0001

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.002	-0.0019
0.040	0.037	0.0028
0.100	0.100	-0.0004
0.200	0.201	-0.0006
0.400	0.400	0.0002

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1410 (-0.0060)
Sample #2 = 0.1430 (-0.0050)
Sample #3 = 0.1480 (-0.0180)
Sample #4 = 0.1390 (-0.0270)
Avg % Abs = 0.1433 (-0.0167)
STD DEU = 0.0045 (-0.0111)
REL STD DEU = 3.146 (66.363)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.4770 (-0.0050)
Sample #2 = 3.4810 (-0.0230)
Sample #3 = 3.4900 (-0.0220)
Sample #4 = 3.4730 (-0.0360)
Avg % Abs = 3.4813 (-0.0270)
STD DEU = 0.0085 (-0.0078)
REL STD DEU = 0.244 (28.927)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 12.5620 (-0.0070)
Sample #2 = 12.5610 (-0.0380)
Sample #3 = 12.5320 (-0.0520)
Sample #4 = 12.5710 (-0.0420)
Avg % Abs = 12.5543 (-0.0440)
STD DEU = 0.0201 (-0.0072)
REL STD DEU = 0.160 (16.389)

Channel 2 >>>>
Soil Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.086
Std Dev = 0.02 Rel Std Dev = 20.96
Soil Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.702
Std Dev = 0.03 Rel Std Dev = 4.30
Soil Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.787
Std Dev = 0.03 Rel Std Dev = 1.60
Soil Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.440
Std Dev = 0.02 Rel Std Dev = 0.72
Soil Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 6.637
Std Dev = 0.03 Rel Std Dev = 0.39
Zero Order Coef = -156.43
First Order Coef = 2734.40
Second Order Coef = 23.84
Standard Deviation = 79.057556

Channel 2 >>>>
Soil Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

Soil Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum 10 = 12602, Sum 10 = 12755

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.6110 (-0.0000)
Sample #2 = 0.6880 (-0.0070)
Sample #3 = 0.7370 (-0.0420)
Sample #4 = 0.6820 (-0.0870)
Avg % Abs = 0.7023 (-0.0453)
STD DEU = 0.0302 (-0.0401)
REL STD DEU = 4.296 (88.455)

Soil Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum 10 = 12540, Sum 10 = 12730

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.4490 (-0.0120)
Sample #2 = 3.4360 (-0.0320)
Sample #3 = 3.4180 (-0.0520)
Sample #4 = 3.4670 (-0.0380)
Avg % Abs = 3.4403 (-0.0407)
STD DEU = 0.0248 (-0.0103)
REL STD DEU = 0.720 (25.237)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.6620 (-0.0240)
Sample #2 = 6.6800 (-0.0150)
Sample #3 = 6.6550 (-0.0280)
Sample #4 = 6.6720 (-0.0223)
Avg % Abs = 6.6783 (-0.0223)
STD DEU = 0.0146 (-0.0067)
REL STD DEU = 0.218 (29.813)

Channel 2 >>>>
Soil Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

SWB

OPTICAL BENCH
CALIBRATION
INSTRUMENT #
SD-006028

Dry Gas H2O Adjust Result 5 *****
Barometric Pressure = 1023
3 um H2O Adjust (mg/l*10,000) = 481
9 um H2O Adjust (mg/l*10,000) = 406
***** AUTO CAL PASS

SLT

POST CALIBRATION STABILITY CHECKS - INSTRUMENT # 80-006028

PALM BEACH CO SO
 IntoxiLizer - Alcohol Analyzer
 Model 8000 SN 80-006028
 03/07/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:52
Control Test	0.051	14:53
Air Blank	0.000	14:53
Control Test	0.052	14:54
Air Blank	0.000	14:55
Control Test	0.051	14:55
Air Blank	0.000	14:56
Control Test Stats		
Average	0.0513	
Std Dev	0.0006	
Rel Std Dev(%)	1.1247	

SP
 Operator's Signature

PALM BEACH CO SO
 IntoxiLizer - Alcohol Analyzer
 Model 8000 SN 80-006028
 03/07/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:44
Control Test	0.080	14:45
Air Blank	0.000	14:46
Control Test	0.081	14:46
Air Blank	0.000	14:47
Control Test	0.082	14:48
Air Blank	0.000	14:48
Control Test Stats		
Average	0.0810	
Std Dev	0.0010	
Rel Std Dev(%)	1.2346	

SP
 Operator's Signature

PALM BEACH CO SO
 IntoxiLizer - Alcohol Analyzer
 Model 8000 SN 80-006028
 03/07/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:40
Control Test	0.198	14:40
Air Blank	0.000	14:41
Control Test	0.200	14:42
Air Blank	0.000	14:42
Control Test	0.199	14:43
Air Blank	0.000	14:43
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

SP
 Operator's Signature

PALM BEACH CO SO
 IntoxiLizer - Alcohol Analyzer
 Model 8000 SN 80-006028
 03/07/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:49
Control Test	0.080	14:49
Air Blank	0.000	14:50
Control Test	0.079	14:50
Air Blank	0.000	14:51
Control Test	0.080	14:51
Air Blank	0.000	14:51
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SP
 Operator's Signature

DK

SP

DGS