

Agency Volusia County

S/N 80-006262

MAY 01 2017

Date In 4/13/17

Date Out 4/27/17

Ship P/U H/D CMI DEF

FDLE
Alcohol Testing Program

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 222
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .156 (.139 - .169)
 36mm .171 (.156 - .190)
 53mm .238 (.228 - .278)
 103mm .503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>AG626604</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure
 ID# 28421 Gauge 1009
 Instrument 1009

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD3964</u>
0.20	<u>SD1025</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC @ 4/28/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

[Signature]
Quality Control Review

5/1/17
Date

UOLUSIA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006262
04/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:38
Control Test	0.049	09:39
Air Blank	0.000	09:39
Control Test	0.048	09:40
Air Blank	0.000	09:40
Control Test	0.049	09:41
Air Blank	0.000	09:42
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

P Murphy
Operator's Signature

UOLUSIA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006262
04/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:33
Control Test	0.079	09:34
Air Blank	0.000	09:34
Control Test	0.079	09:35
Air Blank	0.000	09:35
Control Test	0.079	09:36
Air Blank	0.000	09:37
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
Operator's Signature

UOLUSIA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006262
04/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:44
Control Test	0.198	09:44
Air Blank	0.000	09:45
Control Test	0.197	09:46
Air Blank	0.000	09:46
Control Test	0.197	09:47
Air Blank	0.000	09:48
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

P Murphy
Operator's Signature

UOLUSIA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006262
04/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:54
Control Test	0.080	09:55
Air Blank	0.000	09:55
Control Test	0.079	09:56
Air Blank	0.000	09:56
Control Test	0.079	09:56
Air Blank	0.000	09:57
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

P Murphy
Operator's Signature

JP

JP