

INSTRUMENT PROCESSING SHEET

Agency Flagler County SO S/N 80-006616
 Date In 8/24/17 Date Out 8/31/17 Ship P/U H/D CMI EE

RECEIVED
 AUG 31 2017
 FDLE
 Alcohol Testing Program

Intake Performed By JD

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case OK Handle OK
 Dry Gas Holder OK Feet OK
 Keyboard/Plug OK Back/Plugs OK
 Screws tight OK Breath Hose OK

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By JD

Lab Temp °C 21.5
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 223
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 104 (.139 - .169)
 36mm 179 (.156 - .190)
 53mm 253 (.228 - .278)
 103mm 515 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3962	201707D 7/25/19
0.08	SD1013	201707E 7/25/19
0.20	SD3933	201707C 7/24/19
0.08 DGS	N/A	AG619605 7/14/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 170mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By JD
Simulator Temperatures °C
 External Digital Therm. ID#: 300505
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD1013
 34°C +/- .2 Serial #: SD3933

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By JD

Barometric Pressure 1014 Gauge
 ID# 26932 1014 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD3962
0.08	SD1013
0.20	SD3933

Attachments

Form 41 Calibration Adjustment
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes/Suggested Service:
QA/QC OK PGM 8/31/2017

[Signature] 8/31/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review _____ Date _____

80-006616
 Stability Checks
 8/31/17

INTOXILYZER 8000
 Instrument Initialization
 08:09 08/31/2017

FLAGLER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-006616
 08/31/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:44
Control Test	0.079	08:44
Air Blank	0.000	08:45
Control Test	0.080	08:45
Air Blank	0.000	08:45
Control Test	0.080	08:46
Air Blank	0.000	08:46
Control Test Stats		
Average	0.0797	
Std Dev	0.0016	
Rel Std Dev(%)	0.7247	

Operator's Signature

8/31/17
 JCB

FLAGLER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-006616
 08/31/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.049	08:51
Air Blank	0.000	08:52
Control Test	0.050	08:53
Air Blank	0.000	08:53
Control Test	0.049	08:54
Air Blank	0.000	08:54
Control Test Stats		
Average	0.0493	
Std Dev	0.0016	
Rel Std Dev(%)	1.1703	

Operator's Signature

FLAGLER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-006616
 08/31/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:56
Control Test	0.196	08:57
Air Blank	0.000	08:57
Control Test	0.196	08:58
Air Blank	0.000	08:59
Control Test	0.195	08:59
Air Blank	0.000	09:00
Control Test Stats		
Average	0.1957	
Std Dev	0.0016	
Rel Std Dev(%)	0.2951	

Operator's Signature

FLAGLER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-006616
 08/31/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:02
Control Test	0.079	09:02
Air Blank	0.000	09:03
Control Test	0.079	09:04
Air Blank	0.000	09:04
Control Test	0.080	09:05
Air Blank	0.000	09:05
Control Test Stats		
Average	0.0793	
Std Dev	0.0016	
Rel Std Dev(%)	0.7277	

Operator's Signature

Alan