



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Kennedy Space Center S/N 80-006621
Date In 6/9/17 Date Out 6/20/17 Ship P/U H/D CMI EE

Intake Performed By [Signature]
Registration
Annual
Return from CMI
Return from Enforcement Electronics
Other
Visual Inspection:
Case Handle
Dry Gas Holder Feet
Keyboard/Plug Back/Plugs
Screws tight Breath Hose
Other Equipment:
Power cord
Printer Cable
Other

Quality Checks Performed By SP
Lab Temp 22.5
Breath Tube Screen
Replace External O-Rings
Instrument Set Up Verified
R-Value 225
Flow Verification (L/s)
Flow Column # ATP103
32mm .160 (.139 - .169)
36mm .171 (.156 - .190)
53mm .238 (.228 - .278)
103mm .492 (.447 - .547)
Barometric Pressure Check
Gauge ID # 28427
Stability Checks
Table with Simulator, Serial #, Lot #/Exp

Flow Calibration Performed By [Signature]
Flow Calibration N/A
Flow Calibration Complete
Flow Column #
5L/min 17mm
15L/min 53mm
30L/min 103mm
R-Value
Post Calibration Verification (L/s)
Flow Column #
32mm (.139 - .169)
36mm (.156 - .190)
53mm (.228 - .278)
103mm (.447 - .547)
Maintenance Performed By
Battery Replacement
Dry Gas Regulator Replacement
Breath Tube Replacement
Other
Quality Checks Cont. Performed By SP
Simulator Temperatures C
External Digital Therm. ID#: 300504
34C +/- .2 Serial #: SD3962
34C +/- .2 Serial #: SD3964
34C +/- .2 Serial #: SD3933

RECEIVED JUN 20 2017 FDLE Alcohol Testing Program

Calibration Adjustment Performed By
Calibration Adjustment N/A
Calibration Adjustment Complete
Barometric Pressure Gauge ID #
Table with Simulator, Serial Number, Lot Number, Expiration
Post Calibration Adjustment Stability Checks
Table with Simulator, Serial Number, Lot Number, Expiration

Department Inspection Performed By SP
Barometric Pressure 1015 Gauge
ID# 28427 1015 Instrument
Mouth Alcohol Solution Lot # 2016-C
Acetone Stock Solution Lot # 2017-A
Table with Simulator, Serial Number
Attachments
Form 41
Pre-Stability Tests
Flow Calibration
Calibration Adjustment
Post-Stability Tests
Other

Notes/Suggested Service:
QA/BC OK GSPM 6/20/2017
Quality Control Review [Signature] Date 6/20/17

Instrument Complies with Chapter 11D-8, FAC
Instrument Does Not Comply with Chapter 11D-8, FAC
Return to/Place into Evidentiary Use
Remain Out of Evidentiary Use
Conduct an Agency Inspection Before Evidentiary Use

STABILITY CHECKS - # 80-006621 - KENNEDY SPACE CENTER

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006621
06/20/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:41
Control Test	0.046	09:41
Air Blank	0.000	09:42
Control Test	0.048	09:42
Air Blank	0.000	09:43
Control Test	0.048	09:44
Air Blank	0.000	09:44
Control Test Stats		
Average	0.0473	
Std Dev	0.0012	
Rel Std Dev(%)	2.4395	

SP
Operator's Signature

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006621
06/20/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:45
Control Test	0.048	09:46
Air Blank	0.000	09:46
Control Test	0.048	09:47
Air Blank	0.000	09:48
Control Test	0.048	09:48
Air Blank	0.000	09:49
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
Operator's Signature

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006621
06/20/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:09
Control Test	0.078	10:10
Air Blank	0.000	10:10
Control Test	0.078	10:11
Air Blank	0.000	10:11
Control Test	0.078	10:12
Air Blank	0.000	10:13
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
Operator's Signature

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006621
06/20/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:54
Control Test	0.197	09:54
Air Blank	0.000	09:55
Control Test	0.197	09:56
Air Blank	0.000	09:56
Control Test	0.197	09:57
Air Blank	0.000	09:58
Control Test Stats		
Average	0.1970	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
Operator's Signature

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006621
06/20/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:14
Control Test	0.077	10:14
Air Blank	0.000	10:14
Control Test	0.077	10:15
Air Blank	0.000	10:15
Control Test	0.077	10:16
Air Blank	0.000	10:16
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

SP
Operator's Signature

GSM

6/20/17
SP

Tightened
Sim Head
& Retested SP