

INSTRUMENT PROCESSING SHEET

Agency Kennedy Space Center S/N 80-006759
 Date In 2/27/17 Date Out 3/2/17 Ship P/U H/D CMI EE

Intake Performed By SP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By SP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 264
 Flow Verification (L/s)
 Flow Column # ATP105
 32mm .167 (.139 - .169)
 36mm .183 (.156 - .190)
 53mm .253 (.228 - .278)
 103mm .523 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3-8-18</u>
0.08	<u>SD1011</u>	<u>201601F</u> <u>1-26-18</u>
0.20	<u>SD1025</u>	<u>201604C</u> <u>4-5-18</u>
0.08 DGS	N/A	<u>AG626604</u> <u>9-22-18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (DGS)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 MAR 08 2017
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure 1019 Gauge
 ID# 28421 1019 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD1011</u>
0.20	<u>SD1025</u>

Attachments

<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Optical Bench Cal
<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Post-Stability Tests
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____

Notes: Dev DGS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Hankland

Quality Control Review

3/8/17
Date

STABILITY CHECKS - INSTRUMENT # 80-006759

KENNEDY SPACE CENTER
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006759
 03/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:24
Control Test	0.049	10:25
Air Blank	0.000	10:25
Control Test	0.049	10:26
Air Blank	0.000	10:26
Control Test	0.049	10:27
Air Blank	0.000	10:28
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

KENNEDY SPACE CENTER
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006759
 03/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:12
Control Test	0.079	10:13
Air Blank	0.000	10:13
Control Test	0.080	10:14
Air Blank	0.000	10:15
Control Test	0.081	10:15
Air Blank	0.000	10:16
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

SP

Operator's Signature

KENNEDY SPACE CENTER
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006759
 03/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:17
Control Test	0.197	10:17
Air Blank	0.000	10:18
Control Test	0.197	10:19
Air Blank	0.000	10:19
Control Test	0.197	10:20
Air Blank	0.000	10:20
Control Test Stats		
Average	0.1970	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

KENNEDY SPACE CENTER
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006759
 03/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:21
Control Test	0.080	10:21
Air Blank	0.000	10:22
Control Test	0.080	10:22
Air Blank	0.000	10:23
Control Test	0.080	10:23
Air Blank	0.000	10:23
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DAS

BK

SP

SP

Operator's Signature