FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FHP - TROOP F S/N 80-0067661

Date In 4/20/17 Date Out 4/26/17 ☒Ship ☐P/U ☐H/D ☐CMI ☐EE

RECEIVED MAY 01 2017 FDLE Alcohol Testing Program

Intake Performed By [illegible]

- ☒ Registration
- ☒ Annual
- ☒ Return from CMI
- ☐ Return from Enforcement Electronics
- ☐ Other

Visual Inspection:

OK Case	OK Handle
OK Dry Gas Holder	OK Feet
OK Keyboard/Plug	OK Back/Plugs
OK Screws tight	OK Breath Hose

Other Equipment:

- ☒ Power cord
- ☒ Printer Cable
- ☒ Other STATIC BAG

Notes:

Quality Checks Performed By SP

- ☒ Breath Tube Screen
- ☒ Replace O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 239
- ☒ Flow Verification (L/s)

Flow Column # ATP102

32mm .152 (.139 - .169)
36mm .167 (.156 - .190)
53mm .234 (.228 - .278)
103mm .496 (.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 28427

- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G3709	20160 3D 3-8-18
0.08	DR1279	20161 1B 11-15-18
0.20	DR3856	20160 4C 4-5-18
0.08 DGS	N/A	AG1612405 5-3-18

Flow Calibration Performed By

- ☒ Flow Calibration N/A
- ☐ Flow Calibration Complete

Flow Column # N/A

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)

Flow Column #

32mm (.139 - .169)
36mm (.156 - .190)
53mm (.228 - .278)
103mm (.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Suggested Service

Optical Bench Calibration Performed By

- ☒ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

- ☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

- ☒ Barometric Pressure ID# 28427 — 1009 Gauge / 1008 Instrument

Mouth Alcohol Solution Lot # 2016-A
Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	G3709
0.08	DR1279
0.20	DR3856

Attachments

- ☒ Form 41
- ☒ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other

Notes: Please change level 2 password to something unique SP

QA/QC OK BJM 4/26/17

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

BJ Graham
Quality Control Review

5/1/17
Date

Revised November 2012

STABILITY CHECKS - INSTRUMENT # 80-0067761 - FHP

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006761
04/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:21
Control Test	0.047	09:22
Air Blank	0.000	09:22
Control Test	0.047	09:23
Air Blank	0.000	09:23
Control Test	0.047	09:24
Air Blank	0.000	09:25
Control Test Stats		
Average	0.0470	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RGM

SP
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006761
04/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.077	09:33
Air Blank	0.000	09:33
Control Test	0.077	09:34
Air Blank	0.000	09:34
Control Test	0.077	09:35
Air Blank	0.000	09:36
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006761
04/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:26
Control Test	0.194	09:27
Air Blank	0.000	09:28
Control Test	0.194	09:28
Air Blank	0.000	09:29
Control Test	0.195	09:30
Air Blank	0.000	09:30
Control Test Stats		
Average	0.1943	
Std Dev	0.0006	
Rel Std Dev(%)	0.2971	

SP
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006761
04/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:36
Control Test	0.079	09:37
Air Blank	0.000	09:37
Control Test	0.078	09:37
Air Blank	0.000	09:38
Control Test	0.078	09:38
Air Blank	0.000	09:39
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

SP
Operator's Signature

FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FHP S/N 80-006761

Date In 3/2/17 Date Out 3/7/17 ☐Ship ☐P/U ☐H/D ☒CMI ☐EE

RECEIVED
MAY 01 2017
FDLE
Alcohol Testing Program

Intake Performed By [illegible]

- ☒ Registration
- ☒ Annual
- ☐ Return from CMI
- ☐ Return from Enforcement
- ☐ Other _____

Visual Inspection:

☒ Case	☒ Handle
☒ Dry Gas Holder	☒ Feet
☒ Keyboard/Plug	☒ Back/Plugs
☒ Screws tight	☒ Breath Hose

Other Equipment:

- ☐ Power cord
- ☐ Printer Cable
- ☐ Other _____

Notes: _____

Quality Checks Performed By SP

- ☒ Breath Tube Screen
- ☒ Replace O-Rings
- ☐ Instrument Set Up Verified
- ☒ R-Value 234
- ☒ Flow Verification (L/s)

Flow Column # ATP105

32mm	.164	(.139 - .169)
36mm	.175	(.156 - .190)
53mm	.242	(.228 - .278)
103mm	.511	(.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 28421

- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	DR2035	20160 3D 3-8-18
0.08	SD1011	20160 1F 1-26-18
0.20	SD1025	20160 4C 4-5-18
0.08 DGS	N/A	AG6026604 9-22-18

Flow Calibration Performed By _____

- ☒ Flow Calibration N/A
- ☐ Flow Calibration Complete

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32mm	_____	(.139 - .169)
36mm	_____	(.156 - .190)
53mm	_____	(.228 - .278)
103mm	_____	(.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Suggested Service

Optical Bench Calibration Performed By _____

- ☐ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By _____

- ☐ Barometric Pressure _____ Gauge
  ID# _____ _____ Instrument

Mouth Alcohol Solution Lot # _____

Acetone Stock Solution Lot # _____

Simulator	Serial Number
0.00	
Interferent	
0.05	
0.08	
0.20	

Attachments

- ☐ Form 41
- ☐ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other _____

Notes: _____

- ☐ **Instrument Complies with Chapter 11D-8, FAC**
- ☒ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☐ **Return to/Place into Evidentiary Use**
- ☒ **Remain Out of Evidentiary Use**
- ☐ **Conduct an Agency Inspection Before Evidentiary Use**

JO

Quality Control Review _____ Date _____

STABILITY CHECKS - INSTRUMENT # 80-006761

11:08 03/02/2017

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006761
03/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:35
Control Test	0.046	12:36
Air Blank	0.000	12:36
Control Test	0.047	12:37
Air Blank	0.000	12:37
Control Test	0.047	12:38
Air Blank	0.000	12:39
Control Test Stats		
Average	0.0467	
Std Dev	0.0006	
Rel Std Dev(%)	1.2372	

SP
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006761
03/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:40
Control Test	0.076	12:41
Air Blank	0.000	12:41
Control Test	0.077	12:42
Air Blank	0.000	12:42
Control Test	0.077	12:43
Air Blank	0.000	12:44
Control Test Stats		
Average	0.0767	
Std Dev	0.0006	
Rel Std Dev(%)	0.7531	

SP
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006761
03/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:51
Control Test	0.191	12:52
Air Blank	0.000	12:53
Control Test	0.193	12:53
Air Blank	0.000	12:54
Control Test	0.194	12:55
Air Blank	0.000	12:55
Control Test Stats		
Average	0.1927	
Std Dev	0.0015	
Rel Std Dev(%)	0.7928	

SP
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006761
03/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:45
Control Test	0.078	12:45
Air Blank	0.000	12:45
Control Test	0.077	12:46
Air Blank	0.000	12:46
Control Test	0.077	12:47
Air Blank	0.000	12:47
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

DGS

SP
Operator's Signature

SP