



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Miami Dade Police Department S/N 80-006895

Date In 6/15/2017 Date Out 7/3/2017 Ship P/U H/D CMI EE

Intake Performed By <u>DERL</u> <input checked="" type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>ANTI STATIC BAG</u> Notes: _____ _____ _____ _____	Quality Checks Performed By <u>DERL</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.90</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>144</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATA 101</u> 32mm <u>.167</u> (.139 - .169) 36mm <u>.187</u> (.156 - .190) 53mm <u>.253</u> (.228 - .278) 103mm <u>.519</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD3967</u></td> <td><u>201603D</u> <u>03/08/2018</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3968</u></td> <td><u>201611B</u> <u>11/15/2018</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3969</u></td> <td><u>201604C</u> <u>04/05/2018</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>A6626604</u> <u>09/22/2018</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>SD3967</u>	<u>201603D</u> <u>03/08/2018</u>	0.08	<u>SD3968</u>	<u>201611B</u> <u>11/15/2018</u>	0.20	<u>SD3969</u>	<u>201604C</u> <u>04/05/2018</u>	0.08 DGS	N/A	<u>A6626604</u> <u>09/22/2018</u>	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	<u>SD3967</u>	<u>201603D</u> <u>03/08/2018</u>															
0.08	<u>SD3968</u>	<u>201611B</u> <u>11/15/2018</u>															
0.20	<u>SD3969</u>	<u>201604C</u> <u>04/05/2018</u>															
0.08 DGS	N/A	<u>A6626604</u> <u>09/22/2018</u>															
		Maintenance Performed By <u>DERL</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Quality Checks Cont. Performed By <u>DERL</u> Simulator Temperatures °C External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3969</u>															

RECEIVED
JUL 19 2017
FDLE
Alcohol Testing Program

Calibration Adjustment Performed By _____ <input checked="" type="checkbox"/> Calibration Adjustment N/A <input type="checkbox"/> Calibration Adjustment Complete Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>DERL</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1017</u> Gauge ID# <u>68635</u> <u>1018</u> Instrument Mouth Alcohol Solution Lot # <u>2016 A</u> Acetone Stock Solution Lot # <u>2016 B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td><u>SD 3965</u></td> </tr> <tr> <td>Interferent</td> <td><u>SD3966</u></td> </tr> <tr> <td>0.05</td> <td><u>SD 3967</u></td> </tr> <tr> <td>0.08</td> <td><u>SD 3968</u></td> </tr> <tr> <td>0.20</td> <td><u>SD 3969</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	<u>SD 3965</u>	Interferent	<u>SD3966</u>	0.05	<u>SD 3967</u>	0.08	<u>SD 3968</u>	0.20	<u>SD 3969</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	<u>SD 3965</u>																																																												
Interferent	<u>SD3966</u>																																																												
0.05	<u>SD 3967</u>																																																												
0.08	<u>SD 3968</u>																																																												
0.20	<u>SD 3969</u>																																																												
Notes/Suggested Service: <u>07/03/2017</u> <input checked="" type="checkbox"/> APPROVED <u>QA /pc OK PSM 7/13/2017</u> <u>JJ Dehen</u> <u>7/19/17</u> Quality Control Review Date	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Tests <input checked="" type="checkbox"/> Other <u>FORM 40</u> <u>REGISTRATION</u>																																																												

Notes/Suggested Service: 07/03/2017
 APPROVED
QA /pc OK PSM 7/13/2017
JJ Dehen 7/19/17
Quality Control Review Date

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006895	Miami-Dade Police Department	7/3/2017	<i>MLL</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.07c	SN: SD3968 Temp: 34.09c	SN: SD3969 Temp: 34.06c	Lot AG626604
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

<p>MIAMI POLICE DEPARTM Intoxilyzer - Alcohol Analyzer Model 8000 07/03/2017 SN 80-006895 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 06:55 Control Test 0.050 06:55 Air Blank 0.000 06:56 Control Test 0.050 06:57 Air Blank 0.000 06:57 Control Test 0.050 06:58 Air Blank 0.000 06:58 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <i>MLL</i></p>	<p>MIAMI POLICE DEPARTM Intoxilyzer - Alcohol Analyzer Model 8000 07/03/2017 SN 80-006895 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:00 Control Test 0.080 07:01 Air Blank 0.000 07:02 Control Test 0.080 07:02 Air Blank 0.000 07:03 Control Test 0.079 07:03 Air Blank 0.000 07:04 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247</p> <p>Operator's Signature <i>MLL</i></p>	<p>MIAMI POLICE DEPARTM Intoxilyzer - Alcohol Analyzer Model 8000 07/03/2017 SN 80-006895 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:06 Control Test 0.198 07:06 Air Blank 0.000 07:07 Control Test 0.196 07:07 Air Blank 0.000 07:08 Control Test 0.196 07:09 Air Blank 0.000 07:09 Control Test Stats Average 0.1967 Std Dev 0.0012 Rel Std Dev(%) 0.5871</p> <p>Operator's Signature <i>MLL</i></p>	<p>MIAMI POLICE DEPARTM Intoxilyzer - Alcohol Analyzer Model 8000 07/03/2017 SN 80-006895 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:11 Control Test 0.080 07:11 Air Blank 0.000 07:12 Control Test 0.078 07:12 Air Blank 0.000 07:13 Control Test 0.079 07:13 Air Blank 0.000 07:13 Control Test Stats Average 0.0790 Std Dev 0.0010 Rel Std Dev(%) 1.2658</p> <p>Operator's Signature <i>MLL</i></p>
---	---	---	---

MLL

JD