

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO  
Time of Inspection: 17:23

Date of Inspection: 07/25/2018

Serial Number: 80-000795  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A2 Exp: 08/05/2019
0.000	0.051	0.080	0.201	0.073 / 0.073
0.000	0.051	0.080	0.202	0.073 / 0.073
0.000	0.051	0.080	0.202	0.073 / 0.074

Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: UNKNOWN PROBLEM WITH INTRUMENT.  
ENT.

*PB*

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*James P Baxley* JAMES P BAXLEY  
Signature and Printed Name

07/25/2018  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO  
Time of Inspection: 16:42

Date of Inspection: 07/25/2018

Serial Number: 80-000795  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1781708A2 Exp: 08/05/2019
0.000	0.051	0.080	0.201	0.073 / 0.075
0.000	0.051	0.080	0.203	0.075 / 0.074
0.000	0.051	0.081	0.202	0.075 / 0.074

Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: UNKNOWN PROBLEM WITH INSTRUMENT.

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAMES P BAXLEY  
Signature and Printed Name

07/25/2018  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO

Time of Inspection: 15:25

Date of Inspection: 07/25/2018

Serial Number: 80-000795  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1781708A2 Exp: 08/05/2019
0.000	0.051	0.079	0.199	0.075 / 0.075
0.000	0.051	0.080	0.201	0.074 / 0.074
0.000	0.051	0.081	0.202	0.074 / 0.075

Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance.UN

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*James P. Baxley*

\_\_\_\_\_  
Signature and Printed Name

JAMES P BAXLEY

07/25/2018  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO  
Time of Inspection: 11:26

Date of Inspection: 06/27/2018

Serial Number: 80-000795  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/21/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A2 Exp: 08/05/2019
0.000	0.050	0.079	0.200	0.076
0.000	0.051	0.080	0.201	0.075
0.000	0.051	0.081	0.202	0.075

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAMES P BAXLEY  
Signature and Printed Name

06/27/2018  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO  
Time of Inspection: 10:49

Date of Inspection: 05/25/2018

Serial Number: 80-000795  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A2 Exp: 08/05/2019
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.203	0.079


Number of Simulators Used: 5

Remarks:

**RECEIVED**  
MAY 31 2018  
FDLE  
Alcohol Testing Program

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAMES P BAXLEY

Signature and Printed Name

05/25/2018  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO  
Time of Inspection: 12:08

Date of Inspection: 04/24/2018

Serial Number: 80-000795  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A2 Exp: 08/05/2019
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.051	0.080	0.202	0.079

Number of Simulators Used: 5

**Remarks:**

A F / M A: TOO MUCH ALCOHOL.


**RECEIVED**

APR 30 2018

**FDLE  
Alcohol Testing Program**

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


JAMES P BAXLEY  
 \_\_\_\_\_  
 Signature and Printed Name

04/24/2018  
 Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

MAR 26 2018

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO  
Time of Inspection: 15:46

Date of Inspection: 03/21/2018

Serial Number: 80-000795  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A2 Exp: 08/05/2019
0.000	0.051	0.079	0.195	0.079
0.000	0.051	0.080	0.197	0.079
0.000	0.051	0.080	0.199	0.079


Number of Simulators Used: 5

**Remarks:**

A F / M A: TOO MUCH ALCOHOL.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAMES P BAXLEY

Signature and Printed Name

03/21/2018  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

FEB 28 2018

FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: GULF COUNTY SO  
Time of Inspection: 12:36

Date of Inspection: 02/23/2018

Serial Number: 80-000795  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A2 Exp: 08/05/2019
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.080	0.199	0.078
0.000	0.051	0.081	0.200	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*James P. Baxley*

JAMES P BAXLEY

Signature and Printed Name

02/23/2018  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
JAN 29 2018

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

FDLE  
Alcohol Testing Program

Agency: GULF COUNTY SO  
Time of Inspection: 11:59

Date of Inspection: 01/12/2018

Serial Number: 800-000795  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17817080A2 Exp: 08/05/2019
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.080	0.199	0.078
0.000	0.050	0.080	0.200	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



\_\_\_\_\_  
Signature and Printed Name

01/12/2018  
Date