

9038P

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FWCC SOUTH REGION
Time of Inspection: 15:26

Date of Inspection: 02/23/2018

Serial Number: 80-000903
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:840875 Exp: 09/19/2020
0.000	0.050	0.081	0.184 / 0.199	0.059 / 0.078
0.000	0.051	0.081	0.192 / 0.200	0.078 / 0.077
0.000	0.051	0.081	0.196 / 0.200	0.077 / 0.077

Number of Simulators Used: 4


Remarks:
20: Control Outside Tolerance. 08: Control Outside Tolerance.

Amended Note

20: Simulator sample jar loose. Jar tightened and repeated test.
08: Dry gas hose not connected. Connected hose and repeated test.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



NATHAN A BROCK

Signature and Printed Name

02/23/2018
Date