

914  
20

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT MYERS PD  
Time of Inspection: 11:58

Date of Inspection: 08/01/2018

Serial Number: 80-000914  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/20/2029	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:29017080A2 Exp: 12/05/2019
0.000	0.050	0.081	0.201	0.078
0.000	0.050	0.081	0.201	0.077
0.000	0.050	0.081	0.201	0.077

Number of Simulators Used: 5

Remarks:


RECEIVED  
AUG 23 2018  
FDLE  
Alcohol Testing Program

8/23/18  
This is to Amend the  
Solution DATE for .20  
The EXPIRE DATE  
Should BE 02/20/2020  
NOT 02/20/2029



The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



03-043  
Signature and Printed Name DAVID S CLAFLIN

08/01/2018  
Date

**COPY**