

Florida Department of Law Enforcement Alcohol Testing Program

1043
SP

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 08:30

Date of Inspection: 11/27/2018

Serial Number: 80-001043
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 13518080A6 Exp: 08/05/2020
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.078	0.199	0.080

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184

SHON W ABELN

Signature and Printed Name

11/27/2018
Date

10438P

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 13:16

Date of Inspection: 10/20/2018

Serial Number: 80-001043
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 13518080A6 Exp: 08/05/2020
0.000	0.049	0.080	0.199	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.080

Number of Simulators Used: 5

Remarks:

**A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.**

RECEIVED

OCT 25 2018

FDLE
Alcohol Testing Program

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184

SHON W ABELN

Signature and Printed Name

10/20/2018
Date

10438P

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 23:32

Date of Inspection: 09/26/2018

Serial Number: 80-001043
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22817080A5 Exp: 10/05/2019
0.000	0.047	0.077	0.194	0.079
0.000	0.047	0.078	0.197	0.079
0.000	0.047	0.078	0.198	0.079

Number of Simulators Used: 5

Remarks:

RECEIVED
OCT 01 2018
FDLE
Alcohol Testing Program

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO

Signature and Printed Name

09/26/2018
Date

104B CR

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 23:14

Date of Inspection: 08/26/2018

Serial Number: 80-001043
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22817080A5 Exp: 10/05/2019
0.000	0.047	0.078	0.195	0.078 / 0.079
0.000	0.047	0.078	0.198	/ 0.078
0.000	0.048	0.078	0.198	/ 0.078

Number of Simulators Used: 5

Remarks:

08: RFI Detect.

P/S Car ver was present with radio on which caused the RFI. Asked him to leave and repeated with no issue.

**A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY _____ D.S.**

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO
Signature and Printed Name

08/26/2018
Date

1043 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 09:13

Date of Inspection: 06/29/2018

Serial Number: 80-001043
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32316080A3 Exp: 01/05/2019
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.078	0.198	0.079

Number of Simulators Used: 5

Remarks:

RECEIVED
JUL 05 2018
FDLE
Alcohol Testing Program

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184

SHON W ABELN

Signature and Printed Name

06/29/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

1043
SP

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 16:01

Date of Inspection: 05/27/2018

Serial Number: 80-001043
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32316080A3 Exp: 01/05/2019
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.049	0.078	0.197	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184

SHON W ABELN

Signature and Printed Name

05/27/2018
Date

1043
8P

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 08:21

Date of Inspection: 04/25/2018

Serial Number: 80-001043
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32316080A3 Exp: 01/05/2019
0.000	0.050	0.078	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.078	0.199	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY
 PAUL BLACKMAN, SHERIFF
 HIGHLANDS COUNTY, FLORIDA
 BY Shon W. Abeln 2184 D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184
 Signature and Printed Name

SHON W ABELN

Signature and Printed Name

04/25/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

1043 SP

RECEIVED
FEB 01 2018
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 04:44

Date of Inspection: 01/27/2018

Serial Number: 80-001043
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.045	0.075	0.192	0.078
0.000	0.046	0.076	0.195	0.078
0.000	0.046	0.076	0.197	0.078


Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY _____ D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



GREGORY S GLASCO

Signature and Printed Name

01/27/2018
Date