

SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO  
Time of Inspection: 02:28

Date of Inspection: 08/01/2018

Serial Number: 80-001077  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:915310 Exp: 03/07/2021
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.049	0.079	0.199	0.079

Number of Simulators Used: 5


Remarks:

AF/MA: . on first Alcohol Free test is read Slope Not Met repeated test and past second time.

RECEIVED  
AUG 01 2018  
FDLE  
Alcohol Testing Program

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name GARY W CARTWRIGHT JR

08/01/2018  
Date

SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Hardee county Sheriff's office  
Time of Inspection: 13:30pm

Serial Number: 80-001077  
Software:

Date of Inspection: 7/29/2018

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: 5

Remarks: During the finale diagnostics test this instrument froze on the analytical check. It said ok for this but never continued. I waited a couple of minutes and nothing changed. I then powered the instrument down waited a couple of minutes and then powered it back on this time on the first diagnostics check it again froze in the same spot. I tried this multiple times and had the same result during the diagnostics check. The instrument was within range for all test that were conducted.


RECEIVED

AUG 01 2018

FDLE  
Alcohol Testing Program

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

Date

1077  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO  
Time of Inspection: 16:46

Date of Inspection: 06/26/2018

Serial Number: 80-001077  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:915310 Exp: 03/07/2021
/ 0.000	0.048	0.079	0.200	0.079
/ 0.000	0.048	0.079	0.200	0.079
/ 0.000	0.048	0.080	0.200	0.079

Number of Simulators Used: 5

Remarks:

00: Ambient Fail. *Did not have Alcohol Free simulator conducted*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GARY W CARTWRIGHT JR

Signature and Printed Name

06/26/2018  
Date

1077 SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO  
Time of Inspection: 15:50

Date of Inspection: 06/26/2018

Serial Number: 80-001077  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:203498 Exp: 12/10/2018
0.000	0.049	0.078	0.198	0.074 / 0.074
0.000	0.050	0.079	0.199	0.073 / 0.074
0.000	0.050	0.080	0.198	0.073 / 0.075


Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: *Changed Dry gas Std and completed another inspection*

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name

GARY W CARTWRIGHT JR

06/26/2018  
Date