

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
NOV 14 2018
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LADY LAKE P.D.
Time of Inspection: 07:45

Date of Inspection: 11/14/2018

Serial Number: 80-001177
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000				
0.000				
0.000				

Number of Simulators Used: 4

Remarks:

A F / M A: OPERATOR ERROR. 00: RFI Detect RADIO MIC KEYED. Non-compliance: TWO OPERATOR ERRORS DETECTED.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Matthew Duryea *257 MATTHEW DURYEA
Signature and Printed Name

11/14/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
APR 16 2018
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Lady Lake Police Department

Serial Number: 80-001177

Time of Inspection: 06:15

Date of Inspection: 04/16/2018

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		X
Diagnostic Check (Pre-Inspection): OK	X	
Alcohol Free Subject Test: 0.000	X	
Mouth Alcohol Test: Slope Not Met	X	
Interferent Detect Test: Interferent Detect	X	
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: 4

Remarks: While conducting the inspection, I accidentally unplugged the instrument after completing the simulator test at the 0.20g/210L level. There were no errors or issues noted to that point in the inspection and the inspection was restarted and completed with no errors or issues

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Andrew C. Crogan III

Signature and Printed Name

04/16/2018
Date