

30-001237

Amended Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 02 2018

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK COUNTY SO
Time of Inspection: 11:17

Date of Inspection: 03/26/2018

Serial Number: 80-001237
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		NO
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22817/84A4 Exp: 10/08/2017
0.000	0.047	0.080	0.195	0.080
0.000	0.047	0.080	0.197	0.080
0.000	0.048	0.080	0.197	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Refus91

Signature and Printed Name

KIMBRA L WINGERS

03/26/2018

Date

