

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SEMINOLE COUNTY S.O.  
Time of Inspection: 01:32

Date of Inspection: 12/28/2018

Serial Number: 80-001271  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:20802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:20709 Exp: 09/22/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:282638 Exp: 02/15/2019
0.000	0.048	0.078	0.201	0.080
0.000	0.048	0.078	0.201	0.079
0.000	0.049	0.078	0.202	0.080

Number of Simulators Used: 5

Remarks: 005 lot# 201803B corrected 12/28/18 KB  
008 lot# 201709A corrected 12/28/18 KB

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KEITH D BETHAM

Signature and Printed Name

12/28/2018  
Date