

SP 1647

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CAPE CORAL PD
Time of Inspection: 09:01

Date of Inspection: 08/01/2018

Serial Number: 80-001647
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/20/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23016080A3 Exp: 09/05/2018
0.000	0.049	0.079	0.199	0.074 / 0.077
0.000	0.050	0.080	0.200	0.074 / 0.076
0.000	0.050	0.080	0.199	0.074 / 0.076

Number of Simulators Used: 5

Remarks:
08: Control Outside Tolerance USED NEW TANK.

RECEIVED
AUG 02 2018
FDLE
Alcohol Testing Program

2nd Test of gas Std Test used:
LOT: 10818080AH
Exp: 06/05/2020

COPY

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David S. Claflin 03.043

DAVID S CLAFLIN

Signature and Printed Name

08/01/2018
Date

SP
1647

Platt, Shayla

From: Claflin, David <dclaflin@sheriffleefl.org>
Sent: Thursday, August 02, 2018 7:58 AM
To: Platt, Shayla
Subject: attachment
Attachments: 20180802071902640.pdf

MS Platt,

Attached is a corrected copy of an inspection to correct the gas lot and expire date.

David

David Claflin | Intox Inspector
Desk: 239-477-1344 | Cell: 239-410-3753
Traffic
Lee County Sheriff's Office
<mailto:dclaflin@sheriffleefl.org> | <http://www.sheriffleefl.org> -----Original Message-----
From: Traffic, Panasonic
Sent: Thursday, August 2, 2018 7:19 AM
To: Claflin, David <dclaflin@sheriffleefl.org>
Subject: Message from "RNP002673E4E0FA"

This E-mail was sent from "RNP002673E4E0FA" (MP C3004ex).

Scan Date: 08.02.2018 07:19:02 (-0400)
Queries to: trafficpan@sheriffleefl.org

RECEIVED
AUG 02 2018
FDLE
Alcohol Testing Program

|
Desk: 239-477-1000

Lee County Sheriff's Office
<mailto:TrafficPan@sheriffleefl.org> | <http://www.sheriffleefl.org>

IMPORTANT MESSAGE

This message is intended for the use of the person or entity to whom it is addressed and may contain information that is privileged and confidential, the disclosure of which is governed by applicable law. If the reader of this email is not the intended recipient or the employee or agent responsible to deliver it to the intended recipient, you are hereby notified that any dissemination, distribution or copying of this information is STRICTLY PROHIBITED. If you have received this email by error, please notify us immediately and destroy the related message. This footnote also confirms that this email message has been swept for the presence of computer viruses, worms, hostile scripts and other email-borne network threats. PLEASE NOTE: Florida has a very broad public records law. Most written communications to or from government officials are public records available to the public and media upon request. Your email communications may be subject to public disclosure per Sec. 119 F.S.