Florida Department of Law Enforcement
Alcohol Testing Program

1/24/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 13:50
Date of Inspection: 12/19/2018
Serial Number: 80-005810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 *	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.029			
0.000	0.034			
0.000	0.044			

Number of Simulators Used: 4

Remarks:
05: Control Outside Tolerance, Ambient Fail. Non-compliance: .
* .05 SIMULATOR: ERROR CODE 4 (TEMP. ABOVE 34.2C)
.05 SIMULATOR (SERIAL # DR 2078) P.M./CALIBRATED ON 01/04/2019

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L FRANZ
Signature and Printed Name

12/19/2018
Date

80-005810

RECEIVED
NOV 07 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 11:21
Date of Inspection: 11/06/2018
Serial Number: 80-005810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.048	0.077	0.191 / 0.186	
0.000	0.048	0.077	0.190 / 0.187	
0.000	0.047	0.076	0.189 / 0.185	

Number of Simulators Used: 5

Remarks:
20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

* .20 SIMULATOR DRAWING AMBIENT AIR - POSSIBLE SEAL/HOSE?

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L FRANZ
Signature and Printed Name

11/06/2018
Date

80-005810

RECEIVED
OCT 17 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 13:59
Date of Inspection: 10/15/2018
Serial Number: 80-005810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A3 Exp: 08/05/2019
0.000	0.046	0.076	0.192	0.078
0.000	0.047	0.077	0.193	0.078
0.000	0.047	0.077	0.193	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L FRANZ
Signature and Printed Name

10/15/2018
Date

FDLE/ATP Form 40 – March 2004

80-005810

RECEIVED
SEP 24 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 10:24 Date of Inspection: 09/24/2018

Serial Number: 80-005810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A3 Exp: 08/05/2019
0.000	0.046	0.076	0.192	0.078
0.000	0.047	0.076	0.193	0.078
0.000	0.047	0.076	0.193	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L FRANZ
Signature and Printed Name

09/24/2018
Date

80-005810

RECEIVED
AUG 21 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

COPY

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 16:36 Date of Inspection: 08/20/2018

Serial Number: 80-005810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A3 Exp: 08/05/2019
0.000	0.047	0.076	0.193	0.078
0.000	0.048	0.077	0.195	0.077
0.000	0.048	0.078	0.196	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L FRANZ
Signature and Printed Name

08/20/2018
Date

80-005810

RECEIVED
AUG 21 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

COPY

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Serial Number: 80-005810
Time of Inspection: 08:59
Date of Inspection: 08/20/2018
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.047	0.070 / 0.074		
0.000	0.047	0.071 / 0.075		
0.000	0.047	0.074 / 0.075		

Number of Simulators Used: 4

Remarks:
08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .
CRACKED SIMULATOR JAR -- G Franz

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L. FRANZ
Signature and Printed Name

08/20/2018
Date

80-0058 10 DB

Florida Department of Law Enforcement
Alcohol Testing Program

RECEIVED
JUL 16 2018
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 08:56 Date of Inspection: 07/16/2018 Serial Number: [illegible] Software: [illegible]

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000	[illegible]	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	[illegible]	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Test (g/210L)	0.05g/210L Test (g/210L) Lot#: [illegible] Exp: 02/07/2019	0.08g/210L Test (g/210L) Lot#: [illegible] Exp: 11/15/20[illegible]	0.20g/210L Test (g/210L) Lot#: 01704 Exp: 02/13/2019	0.05 g/210L Dry Gas Slot Test (g/210L) Lot#: [illegible] Exp: [illegible]
0.000	0.048	0.077	0.195 .195	0.077 .077
0.000	0.049	0.077	0.196 .196	[illegible] .077
0.000	0.048	0.077	0.196 .196	[illegible] .078

[illegible] of [illegible] Blasters [illegible]

Remarks:

[illegible] complies (X) does not comply ( ) with Chapter 11D-8, [illegible]

I certify that I hold a valid Florida Department of Law Enforcement [illegible] performed this inspection in accordance with the provisions of Chapter 11D-8, [illegible]

Gordon L. Franz
Signature and Printed Name [illegible]

7/16/20[illegible]
Date

FDLE/ATP Form 30 - March 2004

80-005810

RECEIVED
JUL 16 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 08:56 Date of Inspection: 07/16/2018 Serial Number: 80-005810
Software: [illegible]

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: [illegible] Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: [illegible] Exp: 11/15/2019	0.20g/210L Test (g/210L) Lot#: [illegible] Exp: [illegible]	0.05 g/210L Dry Gas Blk Test (g/210L) Lot#: [illegible] Exp: [illegible]
0.000	0.048	0.077	0.195	0.077
0.000	0.049	0.077	0.196	[illegible]
0.000	0.048	0.077	0.196	[illegible]

Number of Violations Found: [illegible]

Remarks:

I find the instrument complies (X) does not comply ( ) with Chapter [illegible]

I certify that I hold a valid Florida Department of Law Enforcement [illegible] and that this inspection in accordance with the provisions of Chapter 11D-[illegible]

Gordon L. Franz [illegible]
Signature and Printed Name

07/16/2018
Date

FDLE/ATP Form 30 - March 2004

80-005810

RECEIVED
JUN 25 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 10:56
Date of Inspection: 06/25/2018
Serial Number: 80-005810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703[illegible] Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:2016[illegible] Exp: 11/15/20[illegible]	0.20g/210L Test (g/210L) Lot#:20[illegible]702[illegible] Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17[illegible] Exp: 08/05/2[illegible]
0.000	0.047	0.077	0.191	0.07[illegible]
0.000	0.047	0.076	0.192	0.07[illegible]
0.000	0.046	0.076	0.192	0.07[illegible]

Number of Simulator Used: 3

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, F.A.C.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, F.A.C.

Gordon L. Franz
Signature and Printed Name
GORDON L. FRANZ

06/25/2018
Date

FDLE/ATP Form 40 – March 2001

80-005810

RECEIVED
JUN 06 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 10:42
Date of Inspection: 05/30/2018
Serial Number: 80-005810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A3 Exp: 08/05/2019
0.000	0.047	0.077	0.194	0.077
0.000	0.048	0.078	0.195	0.077
0.000	0.048	0.077	0.195	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
Signature and Printed Name
GORDON L FRANZ

05/30/2018
Date

FDLE ATP Form 40 – March 2004

80-005810 [illegible]

RECEIVED
APR 30 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 11:43
Date of Inspection: 04/30/2018
Serial Number: 80-005810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A3 Exp: 08/05/2019
0.000	0.047	0.079	0.194	0.077
0.000	0.047	0.078	0.195	0.077
0.000	0.048	0.079	0.196	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L Franz
GORDON L FRANZ
Signature and Printed Name

04/30/2018
Date

80-005810

RECEIVED
JAN 23 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 08:35 Date of Inspection: 01/23/2018 Serial Number: 80-005810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A3 Exp: 08/05/2019
0.000	0.044 / 0.047	0.076	0.191	0.078
0.000	0.047 / 0.049	0.077	0.192	0.077
0.000	0.046 / 0.049	0.078	0.193	0.078

Number of Simulators Used: 5

Remarks:
A F / M A: .05: Control Outside Tolerance.
.05 Alcohol REFERENCE Solution READING LOW (0.044) ON FIRST TEST. 2ND TEST (0.047) PASSED.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz
GORDON L FRANZ
Signature and Printed Name

01/23/2018
Date

FDLE/ATP Form 40 – March 2004