

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

1/2/19
JO 80-006235

Agency: ASTATULA PD

Time of Inspection: 13:47

Date of Inspection: 12/30/2018

Serial Number: 80-006235

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 09417080A5 Exp: 05/05/2019
0.000	0.047	0.078	0.194	0.079
0.000	0.048	0.078	0.196	0.079
0.000	0.048	0.078	0.196	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-3, FAC.



JOHN A. CALHOUN
Signature and Printed Name

12/30/2018
Date

* AMENDED

Florida Department of Law Enforcement

Alcohol Testing Program

BK

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Agency: ASTATULA PD

Time of Inspection: 20:15

Date of Inspection: 11/30/2018

Serial Number: 80-006235

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/05/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 05417080A5 Exp: 05/05/2019
/ 0.000	0.048	0.079	0.200	0.078
/ 0.000	0.048	0.079	0.200	0.078
/ 0.000	0.049	0.030	0.199	0.079

Number of Simulators Used: 5

Remarks:

00: Ambient Fail. TIGHTENED LID ON ALCOHOL FREE TEST (Jc)

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-3, FAC.

JOHN A. CALHOUN
Signature and Printed Name11/30/2018
Date

80-006235 RDS

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

OCT 30 2018

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOKILYZER 8000

Agency: ASTATULA PD

Time of Inspection: 17:09

Date of Inspection: 10/30/2018

Serial Number: 80-006235

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 09417030A5 Exp: 05/05/2019
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.203	0.078
0.000	0.050	0.079	0.204	0.078

Number of Simulators Used: 5

Remarks:
NONEThe above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JOHN A CALHOUN

Signature and Printed Name

10/30/2018
Date

80-006235

RECEIVED

SEP 03 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ASTATULA PD

Time of Inspection: 17:47

Date of Inspection: 08/30/2018

Serial Number: 80-006235

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802C Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 09417030A5 Exp: 05/05/2019
0.000	0.046	0.077	0.194	0.079
0.000	0.047	0.078	0.194	0.079
0.000	0.047	0.079	0.194	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name JOHN A CALHOUN

08/30/2018
Date

80-006235 RDS

RECEIVED

JUN 30 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ASTATULA PD

Time of Inspection: 22:11

Date of Inspection: 06/30/2018

Serial Number: 80-006235

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:09417080A5 Exp: 05/05/2019
0.000	0.047	0.078	0.195	0.077
0.000	0.047	0.079	0.198	0.077
0.000	0.047	0.080	0.198	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JOHN A CALHOUN

Signature and Printed Name

06/30/2018
Date

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

RECEIVED
MAY 30 2018
FDLE
Alcohol Testing Program

Agency: ASTATULA PD

Time of Inspection: 01:51

Date of Inspection: 05/30/2018

Serial Number: 80-006235

Software: 8100.27

Check or Test	YES	NO
Data and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:09417080A5 Exp: 05/05/2019
0.000	0.046	0.077	0.196	0.079
0.000	0.046	0.078	0.196	0.077
0.000	0.047	0.078	0.195	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JOHN A. CALHOUN

Signature and Printed Name

05/30/2018
Date

* AMENDED

80-006235 DB

RECEIVED

MAY 02 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ASTATULA PD

Serial Number: 80-006235

Time of Inspection: 16:38

Date of Inspection: 04/26/2018

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 09417080A5 Exp: 05/05/2019
0.000	0.000 / 0.048	0.077	0.196	0.078
0.000	0.000 / 0.049	0.078	0.198	0.078
0.000	0.000 / 0.049	0.078	0.199	0.078

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance. TIGHTENED SIMULATOR HEAD

JAC

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-3, FAC.

JOHN A. CALHOUN

Signature and Printed Name

04/26/2018
Date

80-006235 MB

RECEIVED

MAR 23 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ASTATULA PD

Time of Inspection: 19:23

Date of Inspection: 03/23/2018

Serial Number: 80-006235

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:09417080A5 Exp: 05/05/2019
0.000	0.050	0.081	0.201	0.080
0.000	0.051	0.082	0.202	0.080
0.000	0.050	0.082	0.203	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



141

DEBORAH REASONER

Signature and Printed Name

03/23/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

80-006235 *MS*
RECEIVED
FEB 19 2018
FDLE
Alcohol Testing Program

Agency: ASTATULA PD
Time of Inspection: 15:03

Date of Inspection: 02/19/2018

Serial Number: 80-006235
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:09417080A5 Exp: 05/05/2019
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.080	0.200	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] 141

DEBORAH REASONER

Signature and Printed Name

02/19/2018
Date

80-006235

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
JAN 24 2018

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ASTATULA PD

Time of Inspection: 14:34

Date of Inspection: 01/24/2018

Serial Number: 80-006235

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 09417080A5 Exp: 05/05/2019
0.000	0.048	0.079	0.200	0.078
0.000	0.049	0.081	0.201	0.079
0.000	0.050	0.080	0.202	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] 141

DEBORAH REASONER
Signature and Printed Name

01/24/2018
Date