



# INSTRUMENT PROCESSING SHEET

Agency Sumter County SOS/N 80-000816

Florida Department of Law Enforcement

Date In 02/08/2018 DI Completion Date 02/08/2018 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>TG</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ <b>Final Release Date</b> <b>FDLE</b> <b>FEB 09 2018</b> <b>Alcohol Testing Program</b>	<b>Quality Checks</b> Performed By <u>SMB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>217</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>0.082</u> (.139 - .169) 36 mm <u>0.105</u> (.156 - .190) 53 mm <u>0.199</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>G2835</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD1013</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG708807 03/29/2019</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	G2835	201707D 07/25/2019	0.080	SD1013	201707E 07/25/2019	0.200	SD1025	201707C 07/24/2019	0.080 DGS	N/A	AG708807 03/29/2019	<b>Flow Calibration</b> Performed By <u>SMB</u> Flow Column # <u>ATP 102</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>215</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>SMB</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.8</u> External Digital Therm. ID#: <u>300502</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2835</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1013</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1025</u>
Simulator	Serial #	Lot #/Exp															
0.050	G2835	201707D 07/25/2019															
0.080	SD1013	201707E 07/25/2019															
0.200	SD1025	201707C 07/24/2019															
0.080 DGS	N/A	AG708807 03/29/2019															

<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> Performed By <u>SMB</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1023</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>G2880</td></tr> <tr><td>Interferent</td><td>G2882</td></tr> <tr><td>0.050</td><td>G2835</td></tr> <tr><td>0.080</td><td>SD1013</td></tr> <tr><td>0.200</td><td>SD1025</td></tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____	Simulator	Serial Number	0.000	G2880	Interferent	G2882	0.050	G2835	0.080	SD1013	0.200	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	G2880																																																												
Interferent	G2882																																																												
0.050	G2835																																																												
0.080	SD1013																																																												
0.200	SD1025																																																												

Notes/Suggested Service: _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  <div style="display: flex; justify-content: space-between;"> <div style="text-align: center;"> <u>P. P. 2/9/2018</u>              Tech Review / Date           </div> <div style="text-align: center;"> <u>J. J. 2/9/18</u>              Admin Review / Date           </div> </div>
--	---

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO  
Time of Inspection: 16:44

Date of Inspection: 02/08/2018

Serial Number: 80-000816  
Software: 8100.27

*BCD*

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG708807 Exp: 03/29/2019
0.000	0.049	0.080	0.197	0.081
0.000	0.049	0.081	0.199	0.080
0.000	0.049	0.081	0.200	0.080
0.000	0.049	0.081	0.200	0.080
0.000	0.050	0.081	0.199	0.081
0.000	0.050	0.081	0.199	0.080
0.000	0.049	0.081	0.200	0.080
0.000	0.050	0.081	0.200	0.080
0.000	0.050	0.082	0.200	0.081
0.000	0.050	0.081	0.200	0.080
Standard Deviations	0.0005	0.0004	0.0009	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

*Pgm*

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Danielle M Bell*

DANIELLE M BELL

Signature and Printed Name

02/08/2018  
Date

*2/9/18  
DS*

# Stability Checks #80-000816 Sumter County S.O. 2/8/18 ~~2/8/18~~

067

SUMTER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000816  
02/08/2018  
Software: 8100.27

SUMTER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000816  
02/08/2018  
Software: 8100.27

SUMTER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000816  
02/08/2018  
Software: 8100.27

SUMTER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000816  
02/08/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:52
Control Test	0.081	14:52
Air Blank	0.000	14:53
Control Test	0.080	14:53
Air Blank	0.000	14:54
Control Test	0.080	14:54
Air Blank	0.000	14:55
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Test	g/210L	Time
Air Blank	0.000	14:56
Control Test	0.196	14:56
Air Blank	0.000	14:57
Control Test	0.198	14:57
Air Blank	0.000	14:58
Control Test	0.199	14:58
Air Blank	0.000	14:59
Control Test Stats		
Average	0.1977	
Std Dev	0.0015	
Rel Std Dev(%)	0.7728	

Test	g/210L	Time
Air Blank	0.000	14:39
Control Test	0.080	14:39
Air Blank	0.000	14:40
Control Test	0.081	14:41
Air Blank	0.000	14:41
Control Test	0.080	14:42
Air Blank	0.000	14:42
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Test	g/210L	Time
Air Blank	0.000	14:26
Control Test	0.050	14:27
Air Blank	0.000	14:27
Control Test	0.050	14:28
Air Blank	0.000	14:29
Control Test	0.049	14:29
Air Blank	0.000	14:30
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

*[Signature]*  
Operator's Signature

2/8/18  
*[Signature]*

*[Signature]*  
Operator's Signature

*[Signature]*  
Operator's Signature

*[Signature]*  
Operator's Signature

*[Signature]*



Florida Department of Law Enforcement  
 Alcohol Testing Program  
 2729 Fort Knox Blvd.  
 Bldg. 2, Suite 1300  
 Tallahassee, FL 32308

## Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000816, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000816</u>	UNCERTAINTY* ±
Owning Agency:	<u>SUMTER COUNTY SO</u>	0.05 g/ 210 L 0.004
Calibration Date:	<u>02/08/2018</u>	0.08 g/ 210 L 0.005
Calibration Time:	<u>16:44</u>	0.20 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.  
 Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Target Alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/08/2018

Date

*Danielle M Bell*

**DANIELLE M BELL,**  
**Department Inspector**

FDLE/ATP Form 69 January 2018  
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*2/9/18*

*PJAN*

# Flow Calibration Adjustment Data

#80 - 000816

Sumter County S.O.

2/8/18

~~RB~~

SUMTER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000816  
02/08/2018  
Software: 8100.27

## Flow Rate Calibration\*\*\*\*\*

1: Rate (Liters/min) = 5

SQRT(Diff) = 7.070

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.871

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.711

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 709

Rounded Intercept = -733448

Correlation = 0.99852

RB

2/9/18  
JA