



INSTRUMENT PROCESSING SHEET

Agency Hillsborough County Sheriff's Office s/N 80-000830

Florida Department of Law Enforcement Date In 04/11/2018 DI Completion Date 04/11/2018 Ship P/U H/D CMI EE

Intake Performed By TG Quality Checks Performed By [Signature] Flow Calibration Performed By [Signature]

Maintenance Performed By [Signature] Temperature Checks Performed By [Signature]

Calibration Adjustment Performed By [Signature] Department Inspection Performed By [Signature]

Notes/Suggested Service: Instrument Complies with Chapter 11D-8, FAC Return to/Place into Evidentiary Use

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 16:26

Date of Inspection: 04/11/2018

Serial Number: 80-000830
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.080	0.200	0.078
0.000	0.050	0.081	0.200	0.079
0.000	0.050	0.082	0.201	0.078
0.000	0.051	0.081	0.201	0.079
0.000	0.051	0.081	0.201	0.079
0.000	0.050	0.081	0.200	0.078
0.000	0.051	0.081	0.201	0.078
0.000	0.050	0.081	0.201	0.078
0.000	0.050	0.080	0.201	0.078
0.000	0.050	0.081	0.201	0.078

Standard Deviations	0.0006	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

10

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danelle M Bell

DANIELLE M BELL

Signature and Printed Name

04/11/2018
Date

4/12/18
JD

Stability Checks # 80-000830 Hillsborough County S.O. 4/11/18 ~~RB~~

DES

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000830
04/11/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:32
Control Test	0.051	12:32
Air Blank	0.000	12:33
Control Test	0.049	12:34
Air Blank	0.000	12:34
Control Test	0.049	12:35
Air Blank	0.000	12:35
Control Test Stats		
Average	0.0497	
Std Dev	0.0012	
Rel Std Dev(%)	2.3249	

DES

Operator's Signature

4/12/18
DES

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000830
04/11/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:37
Control Test	0.080	12:38
Air Blank	0.000	12:38
Control Test	0.081	12:39
Air Blank	0.000	12:39
Control Test	0.081	12:40
Air Blank	0.000	12:41
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

DES

Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000830
04/11/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:41
Control Test	0.200	12:42
Air Blank	0.000	12:43
Control Test	0.200	12:43
Air Blank	0.000	12:44
Control Test	0.200	12:45
Air Blank	0.000	12:45
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DES

Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000830
04/11/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:51
Control Test	0.078	12:51
Air Blank	0.000	12:51
Control Test	0.078	12:52
Air Blank	0.000	12:52
Control Test	0.078	12:52
Air Blank	0.000	12:53
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DES

Operator's Signature



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000830, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000830</u>	UNCERTAINTY* ±	
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/11/2018</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>16:26</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/11/2018 _____
 Date Danielle M Bell
 DANIELLE M BELL,
 Department Inspector

FDLE/ATP Form 69 March 2018
 Issuing Authority: Alcohol Testing Program
 Service • Integrity • Respect • Quality

4/12/18
DB