



INSTRUMENT PROCESSING SHEET

Agency Fish & Wildlife CCS/N 80-000904

Florida Department of Law Enforcement

Date In 1/17/2018 DI Completion Date 1/18/2018 Ship P/U H/D CMI EE

Intake Performed By <u>DELL</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input checked="" type="checkbox"/> 12V DC Cable Notes: <u>Pelican 1620 case with lock. DGS and mouth pieces</u>	Quality Checks Performed By <u>DELL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>170</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3963</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG626604 09/22/2018</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3963	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG626604 09/22/2018	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	SD3967	201707D 07/25/2019																																																											
0.080	SD3963	201707E 07/25/2019																																																											
0.200	SD3969	201707C 07/24/2019																																																											
0.080 DGS	N/A	AG626604 09/22/2018																																																											
Final Release Date 	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	Temperature Checks Performed By <u>DELL</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.41c</u> External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3963</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3969</u>																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>DELL</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1027</u> Instrument <u>1026</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3963</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3963	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD3965																																																												
Interferent	SD3966																																																												
0.050	SD3967																																																												
0.080	SD3963																																																												
0.200	SD3969																																																												
Notes/Suggested Service: <u>E-mailed</u> <input checked="" type="checkbox"/> APPROVED	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																												
 	<u>Pgm 1/29/18</u> / Tech Review / Date <u>JJ [Signature] 1/30/18</u> / Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FWCC SOUTH REGION
Time of Inspection: 10:17

Date of Inspection: 01/18/2018

Serial Number: 80-000904
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.050	0.081	0.198	0.081
0.000	0.051	0.082	0.201	0.081
0.000	0.050	0.083	0.202	0.081
0.000	0.051	0.082	0.203	0.081
0.000	0.051	0.082	0.203	0.080
0.000	0.050	0.082	0.203	0.080
0.000	0.050	0.082	0.202	0.081
0.000	0.051	0.082	0.202	0.081
0.000	0.051	0.082	0.202	0.080
0.000	0.051	0.081	0.202	0.081

Standard Deviations	0.0005	0.0005	0.0014	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Pgm

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

01/18/2018
Date

1/30/18
22

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000904	FFWCC SOUTH REGION	01/18/2018	<i>DCA</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.06c	SN: SD3963 Temp: 34.08c	SN: SD3969 Temp: 34.09c	Lot AG626604
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

FFWCC SOUTH REGION Intoxilyzer - Alcohol Analyzer Model: 8000 01/18/2018 Software: 8100.27	FFWCC SOUTH REGION Intoxilyzer - Alcohol Analyzer Model: 8000 01/18/2018 Software: 8100.27	FFWCC SOUTH REGION Intoxilyzer - Alcohol Analyzer Model: 8000 01/18/2018 Software: 8100.27	FFWCC SOUTH REGION Intoxilyzer - Alcohol Analyzer Model: 8000 01/18/2018 Software: 8100.27
Test	Test	Test	Test
Air Blank 0.000 06:27	Air Blank 0.000 06:33	Air Blank 0.000 06:36	Air Blank 0.000 06:43
Control Test 0.050 06:28	Control Test 0.081 06:34	Control Test 0.200 06:39	Control Test 0.082 06:45
Air Blank 0.000 06:28	Air Blank 0.000 06:34	Air Blank 0.000 06:39	Air Blank 0.000 06:44
Control Test 0.050 06:29	Control Test 0.082 06:35	Control Test 0.202 06:40	Control Test 0.081 06:44
Air Blank 0.000 06:30	Air Blank 0.000 06:35	Air Blank 0.000 06:40	Air Blank 0.000 06:44
Control Test 0.050 06:30	Control Test 0.082 06:36	Control Test 0.201 06:41	Control Test 0.082 06:45
Air Blank 0.000 06:31	Air Blank 0.000 06:37	Air Blank 0.000 06:42	Air Blank 0.000 06:45
Control Test Stats	Control Test Stats	Control Test Stats	Control Test Stats
Average 0.0500	Average 0.0617	Average 0.2010	Average 0.0817
Std Dev 0.0000	Std Dev 0.0006	Std Dev 0.0010	Std Dev 0.0006
Rel Std Dev(%) 0.0000	Rel Std Dev(%) 0.7077	Rel Std Dev(%) 0.4975	Rel Std Dev(%) 0.7070
<i>DCA</i> Operator's Signature	<i>DM</i> Operator's Signature	<i>DM</i> Operator's Signature	<i>DM</i> Operator's Signature

8/10/11

DM



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000904, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000904</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FFWCC SOUTH REGION</u>	0.05 g/ 210 L	0.004
Calibration Date:	<u>01/18/2018</u>	0.08 g/ 210 L	0.005
Calibration Time:	<u>10:17</u>	0.20 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Target Alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



 Date 01/18/2018

 DAVID E REYES-RIVERA,
 Department Inspector

FDLE/ATP Form 69 January 2018
Issuing Authority: Alcohol Testing Program

PRM

1/30/18
JD