



INSTRUMENT PROCESSING SHEET

Agency SANIBEL ISLAND PD S/N 80-000937

Florida Department of Law Enforcement

Date In 4-25-18 DI Completion Date 5-15-18 Ship P/U H/D CMI EE

Intake Performed By SP Quality Checks Performed By SP Flow Calibration Performed By
Annual Registration Return from CMI / EE
Visual Inspection: Case Handle Keyboard Dry Gas Shelf Feet Breath Tube Ports Screws Tight
Other Equipment/ Accessories: Power cord Printer Cable Static Bag 12V DC Cable
Notes:
Final Release Date FDLE MAY 15 2018 Alcohol Testing Program

Quality Checks Performed By SP
Breath Tube Screen
Replace External O-Rings
Instrument Set Up Verified
R-Value 181
Flow Verification (L/s)
Flow Column # ATP10X5 SP
32 mm 144 (.139 - .169)
36 mm 160 (.156 - .190)
53 mm 238 (.228 - .278)
103 mm 492 (.447 - .547)
Barometric Pressure Check
Gauge ID # 48639
Stability Checks

Flow Calibration Performed By
Flow Column #
5L/min - 17mm
15L/min - 53mm
30L/min - 103mm
R-Value
Post Calibration Verification (L/s)
Flow Column #
32 mm (.139 - .169)
36 mm (.156 - .190)
53 mm (.228 - .278)
103 mm (.447 - .547)

Table with 3 columns: Simulator, Serial #, Lot #/Exp. Rows include 0.050, 0.080, 0.200, and 0.080 DGS.

Maintenance Performed By
Battery Replacement
Dry Gas Regulator Replacement
Breath Tube Replacement
Other

Temperature Checks Performed By SP
Lab Temp °C 21.5
External Digital Therm. ID#: 300503
34°C +/- .2 Serial #: SD1018
34°C +/- .2 Serial #: SD3962
34°C +/- .2 Serial #: G2078

Calibration Adjustment Performed By SP
Barometric Pressure Gauge 1011 ID # 28427
Table with 4 columns: Simulator, Serial Number, Lot Number, Expiration. Rows include 0.000, 0.040, 0.100, 0.200, 0.300, and 0.080 DGS.

Department Inspection Performed By SP
Barometric Pressure ID# 28662
Gauge 1012 Instrument 1011
Mouth Alcohol Solution Lot # 2016-C
Acetone Stock Solution Lot # 2018-A
Table with 2 columns: Simulator, Serial Number. Rows include 0.000, Interferent, 0.050, 0.080, and 0.200.

Post Calibration Adjustment Stability Checks
Table with 4 columns: Simulator, Serial Number, Lot Number, Expiration. Rows include 0.050, 0.080, 0.200, and 0.080 DGS.

Attachments
Form 41
Stability Checks
Calibration Certificate
Calibration Adjustment
Post-Stability Checks
Flow Calibration
Form 40
Other

Notes/Suggested Service: Performed cal adjustment to bring values closer to nominal. SP

Instrument Complies with Chapter 11D-8, FAC
Instrument Does Not Comply with Chapter 11D-8, FAC
Return to/Place into Evidentiary Use
Remain Out of Evidentiary Use
Conduct an Agency Inspection Before Evidentiary Use
Tech Review / Date 5/15/18 Admin Review / Date 5/15/18

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SANIBEL ISLAND PD  
Time of Inspection: 13:08

Date of Inspection: 05/15/2018

Serial Number: 80-000937  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.080	0.200	0.079
0.000	0.050	0.081	0.202	0.077
0.000	0.050	0.081	0.202	0.077
0.000	0.050	0.081	0.203	0.077
0.000	0.051	0.081	0.202	0.077
0.000	0.050	0.082	0.202	0.077
0.000	0.050	0.082	0.201	0.077
0.000	0.050	0.082	0.201	0.076
0.000	0.050	0.082	0.201	0.077
0.000	0.050	0.082	0.201	0.076

Standard Deviations	0.0004	0.0006	0.0008	0.0008
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

*AMS*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

05/15/2018  
Date

*5/15/18*  
*JP*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000937	SANIBEL ISLAND PD	4-25-18	SP

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
<p>SANIBEL ISLAND PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000937 04/25/2018 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:14</td></tr> <tr><td>Control Test</td><td>0.051</td><td>09:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:15</td></tr> <tr><td>Control Test</td><td>0.051</td><td>09:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:16</td></tr> <tr><td>Control Test</td><td>0.051</td><td>09:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0510</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature _____</p>	Test	g/210L	Time	Air Blank	0.000	09:14	Control Test	0.051	09:14	Air Blank	0.000	09:15	Control Test	0.051	09:16	Air Blank	0.000	09:16	Control Test	0.051	09:17	Air Blank	0.000	09:17	Control Test Stats			Average	0.0510		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>SANIBEL ISLAND PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000937 04/25/2018 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:22</td></tr> <tr><td>Control Test</td><td>0.082</td><td>09:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:23</td></tr> <tr><td>Control Test</td><td>0.083</td><td>09:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:24</td></tr> <tr><td>Control Test</td><td>0.083</td><td>09:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0827</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.6984</td><td></td></tr> </tbody> </table> <p>Operator's Signature _____</p>	Test	g/210L	Time	Air Blank	0.000	09:22	Control Test	0.082	09:22	Air Blank	0.000	09:23	Control Test	0.083	09:24	Air Blank	0.000	09:24	Control Test	0.083	09:25	Air Blank	0.000	09:25	Control Test Stats			Average	0.0827		Std Dev	0.0006		Rel Std Dev(%)	0.6984		<p>SANIBEL ISLAND PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000937 04/25/2018 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:28</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:30</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:31</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:32</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2906</td><td></td></tr> </tbody> </table> <p>Operator's Signature _____</p>	Test	g/210L	Time	Air Blank	0.000	09:28	Control Test	0.199	09:29	Air Blank	0.000	09:30	Control Test	0.199	09:30	Air Blank	0.000	09:31	Control Test	0.198	09:32	Air Blank	0.000	09:32	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel Std Dev(%)	0.2906		<p>SANIBEL ISLAND PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000937 04/25/2018 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:04</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:05</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:06</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> <p>Operator's Signature _____</p>	Test	g/210L	Time	Air Blank	0.000	09:04	Control Test	0.079	09:05	Air Blank	0.000	09:05	Control Test	0.078	09:05	Air Blank	0.000	09:06	Control Test	0.078	09:06	Air Blank	0.000	09:07	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:14																																																																																																																																																	
Control Test	0.051	09:14																																																																																																																																																	
Air Blank	0.000	09:15																																																																																																																																																	
Control Test	0.051	09:16																																																																																																																																																	
Air Blank	0.000	09:16																																																																																																																																																	
Control Test	0.051	09:17																																																																																																																																																	
Air Blank	0.000	09:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0510																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:22																																																																																																																																																	
Control Test	0.082	09:22																																																																																																																																																	
Air Blank	0.000	09:23																																																																																																																																																	
Control Test	0.083	09:24																																																																																																																																																	
Air Blank	0.000	09:24																																																																																																																																																	
Control Test	0.083	09:25																																																																																																																																																	
Air Blank	0.000	09:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0827																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.6984																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:28																																																																																																																																																	
Control Test	0.199	09:29																																																																																																																																																	
Air Blank	0.000	09:30																																																																																																																																																	
Control Test	0.199	09:30																																																																																																																																																	
Air Blank	0.000	09:31																																																																																																																																																	
Control Test	0.198	09:32																																																																																																																																																	
Air Blank	0.000	09:32																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2906																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:04																																																																																																																																																	
Control Test	0.079	09:05																																																																																																																																																	
Air Blank	0.000	09:05																																																																																																																																																	
Control Test	0.078	09:05																																																																																																																																																	
Air Blank	0.000	09:06																																																																																																																																																	
Control Test	0.078	09:06																																																																																																																																																	
Air Blank	0.000	09:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		

5/15/18  
SP

DGS

DGS



Florida Department of Law Enforcement  
 Alcohol Testing Program  
 2729 Fort Knox Blvd.  
 Bldg. 2, Suite 1300  
 Tallahassee, FL 32308

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000937, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000937</u>	UNCERTAINTY* ±	
Owning Agency:	<u>SANIBEL ISLAND PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/15/2018</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:08</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
 \*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 March 2018  
 Issuing Authority: Alcohol Testing Program

05/15/2018

Date

*Shayla D Platt*

SHAYLA D PLATT,  
 Department Inspector

Service • Integrity • Respect • Quality

*5/15/18*  
*JD*

*DS*

SANIBEL ISLAND PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
05/15/2018  
09:27:37  
SN 80-000937

Auto Calibration  
Max Power Res Value = 35  
Auto Range Res Value = 24  
Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12653, 9um Io = 14243

<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1070 (-0.0130)  
Sample #2 = 0.1030 (0.0230)  
Sample #3 = 0.0840 (0.0040)  
Sample #4 = 0.0960 (0.0060)  
Aug % Abs = 0.0943 (0.0110)  
STD DEV = 0.0096 (0.0104)  
REL STD DEV = 10.186 (94.912)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1510 (-0.0040)  
Sample #2 = 0.1820 (0.0000)  
Sample #3 = 0.1510 (-0.0020)  
Sample #4 = 0.1680 (-0.0070)  
Aug % Abs = 0.1670 (-0.0030)  
STD DEV = 0.0155 (0.0036)  
REL STD DEV = 9.296 (120.185)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12647, 9um Io = 14243  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8020 (-0.0120)  
Sample #2 = 0.8010 (-0.0030)  
Sample #3 = 0.8030 (-0.0210)  
Sample #4 = 0.8160 (-0.0020)  
Aug % Abs = 0.8067 (-0.0087)  
STD DEV = 0.0081 (0.0107)  
REL STD DEV = 1.010 (123.377)

5/15/18  
JQ

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.6700 (-0.0070)  
Sample #2 = 1.6440 (0.0190)  
Sample #3 = 1.6380 (0.0130)  
Sample #4 = 1.6420 (0.0230)  
Aug % Abs = 1.6413 (0.0183)  
STD DEV = 0.0031 (0.0050)  
REL STD DEV = 0.186 (27.454)

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12644, 9um Io = 14239  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.7990 (-0.0120)  
Sample #2 = 1.8050 (-0.0050)  
Sample #3 = 1.8070 (-0.0040)  
Sample #4 = 1.8300 (0.0040)  
Aug % Abs = 1.8140 (-0.0017)  
STD DEV = 0.0139 (0.0049)  
REL STD DEV = 0.766 (295.973)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.7990 (-0.0260)  
Sample #2 = 3.7680 (0.0140)  
Sample #3 = 3.7750 (0.0280)  
Sample #4 = 3.7720 (0.0270)  
Aug % Abs = 3.7717 (0.0231)  
STD DEV = 0.0035 (0.0078)  
REL STD DEV = 0.093 (33.958)

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12641, 9um Io = 14238  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.4450 (-0.0200)  
Sample #2 = 3.4620 (-0.0270)  
Sample #3 = 3.4300 (-0.0160)  
Sample #4 = 3.4300 (-0.0130)  
Aug % Abs = 3.4407 (-0.0187)  
STD DEV = 0.0185 (0.0074)  
REL STD DEV = 0.537 (39.488)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 7.1130 (0.0080)  
Sample #2 = 7.0700 (0.0370)  
Sample #3 = 7.0580 (0.0370)  
Sample #4 = 7.0820 (0.0320)  
Aug % Abs = 7.0700 (0.0353)  
STD DEV = 0.0120 (0.0029)  
REL STD DEV = 0.170 (8.170)

Sol Value = 0.300 g/210L \*\*\*  
Fit value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12639, 9um Io = 14237  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 5.0960 (-0.0360)  
Sample #2 = 5.0430 (-0.0040)  
Sample #3 = 5.0370 (-0.0030)  
Sample #4 = 5.0490 (0.0300)  
Aug % Abs = 5.0430 (0.0077)  
STD DEV = 0.0060 (0.0193)  
REL STD DEV = 0.119 (252.361)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 10.3560 (0.0010)  
Sample #2 = 10.2970 (0.0560)  
Sample #3 = 10.2900 (0.0650)  
Sample #4 = 10.3010 (0.0800)  
Aug % Abs = 10.2960 (0.0670)  
STD DEV = 0.0056 (0.0121)  
REL STD DEV = 0.054 (18.096)

Sol Value = 0.000 mg/l or 0.000 g/210L  
Fit value = 0.3610 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 3047.00  
Sample #2 = 3039.00  
Sample #3 = 3086.00  
Sample #4 = 3135.00  
Average Result = 3086.6667  
STD DEV = 48.0035  
REL STD DEV = 1.555  
\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sample #1 = 3257.00  
Sample #2 = 3264.00  
Sample #3 = 3278.00  
Sample #4 = 3289.00  
Average Result = 3277.0000  
STD DEV = 12.5300  
REL STD DEV = 0.382  
\*\*\*\*\*  
Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1011  
3 um H2O Adjust (mg/l\*10,000) = 723  
9 um H2O Adjust (mg/l\*10,000) = 532  
\*\*\*\*\* AUTO CAL PASS \*\*\*\*\*

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
<<<<< CHANNEL 1 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.094  
Std Dev = 0.01 Rel Std Dev = 10.19  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.807  
Std Dev = 0.01 Rel Std Dev = 1.01  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.814  
Std Dev = 0.01 Rel Std Dev = 0.77  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.441  
Std Dev = 0.02 Rel Std Dev = 0.54  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.043  
Std Dev = 0.01 Rel Std Dev = 0.12  
Zero Order Coef = -285.57  
First Order Coef = 2732.58  
Second Order Coef = 31.60  
Standard Deviation = 29.180443  
-----  
<<<<< CHANNEL 2 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.167  
Std Dev = 0.02 Rel Std Dev = 9.30  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.641  
Std Dev = 0.00 Rel Std Dev = 0.19  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.772  
Std Dev = 0.00 Rel Std Dev = 0.09  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 7.070  
Std Dev = 0.01 Rel Std Dev = 0.17  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 10.296  
Std Dev = 0.01 Rel Std Dev = 0.05  
Zero Order Coef = -234.38  
First Order Coef = 1290.17  
Second Order Coef = 11.81  
Standard Deviation = 32.665943  
-----

Solution Stats Quadratic Fit Chan 1  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 -0.001 0.0006  
0.040 0.041 -0.0007  
0.100 0.101 -0.0003  
0.200 0.199 0.0007  
0.300 0.300 -0.0003

Solution Stats Quadratic Fit Chan 2  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 -0.000 0.0004  
0.040 0.040 -0.0002  
0.100 0.101 -0.0008  
0.200 0.199 0.0010  
0.300 0.300 -0.0003

CAL ADJUSTMENT  
#80-000937 SP

POST CAL ADJUST STABILITY CHECKS # 80-000937

SANIBEL ISLAND PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000937  
05/15/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:48
Control Test	0.050	10:49
Air Blank	0.000	10:49
Control Test	0.050	10:50
Air Blank	0.000	10:50
Control Test	0.050	10:51
Air Blank	0.000	10:52
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

SANIBEL ISLAND PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000937  
05/15/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:18
Control Test	0.080	11:19
Air Blank	0.000	11:20
Control Test	0.082	11:20
Air Blank	0.000	11:21
Control Test	0.081	11:22
Air Blank	0.000	11:22
Control Test Stats		
Average	0.0810	
Std Dev	0.0010	
Rel Std Dev(%)	1.2346	

SP

Operator's Signature

SANIBEL ISLAND PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000937  
05/15/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:23
Control Test	0.199	11:24
Air Blank	0.000	11:25
Control Test	0.201	11:25
Air Blank	0.000	11:26
Control Test	0.200	11:27
Air Blank	0.000	11:27
Control Test Stats		
Average	0.2000	
Std Dev	0.0010	
Rel Std Dev(%)	0.5000	

SP

Operator's Signature

SANIBEL ISLAND PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000937  
05/15/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:43
Control Test	0.079	10:43
Air Blank	0.000	10:44
Control Test	0.079	10:44
Air Blank	0.000	10:44
Control Test	0.079	10:45
Air Blank	0.000	10:45
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

SP

SP

Operator's Signature

5/15/18  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SANIBEL ISLAND PD  
Time of Inspection: 08:46

Date of Inspection: 05/15/2018

Serial Number: 80-000937  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT

N/A COMPLIANCE NOT DETERMINED.

DMS

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

05/15/2018  
Date

5/15/18  
JP