



# INSTRUMENT PROCESSING SHEET

Agency Seminole County SOS/N 80-000998

Florida Department of Law Enforcement

Date In 10/22/18DI Completion Date 10/23/18 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>SGC</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>104</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>A7P102</u> 32 mm <u>142</u> (.139 - .169) 36 mm <u>160</u> (.156 - .190) 53 mm <u>234</u> (.228 - .278) 103 mm <u>496</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
---	--	--

**Final Release Date**  
**FDLE**  
 OCT 23 2018  
 Alcohol Testing Program

Simulator	Serial #	Lot #/Exp
0.050	SD1021	201707D 7/25/19
0.080	DR1275	201707E 7/25/19
0.200	SD1013	201707C 7/24/19
0.080 DGS	N/A	AG805701 2/26/20

**Maintenance** Performed By \_\_\_\_\_  
 Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By [Signature]  
 Lab Temp °C 21.4  
 External Digital Therm. ID#: 300503  
 34°C +/-2 Serial #: SD1021  
 34°C +/-2 Serial #: DR1275  
 34°C +/-2 Serial #: SD1013

**Calibration Adjustment** Performed By \_\_\_\_\_

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure ID# 28427  
 Gauge 1018 Instrument 1018  
 Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2018-A 2018-A

Simulator	Serial Number
0.000	G11621
Interferent	DR3855
0.050	SD1021
0.080	DR1275
0.200	SD1013

**Attachments**

<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration
<input checked="" type="checkbox"/> Calibration Certificate	<input checked="" type="checkbox"/> Form 40
<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____

Notes/Suggested Service: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

[Signature] 10/23/18      [Signature] 10/23/18  
 Tech Review / Date      Admin Review / Date

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SEMINOLE COUNTY S.O.  
Time of Inspection: 09:09

Date of Inspection: 10/23/2018

Serial Number: 80-000998  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.050	0.080	0.200	0.079
0.000	0.049	0.081	0.201	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.051	0.081	0.200	0.079

Standard Deviations	0.0004	0.0005	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

*ADB*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

10/23/2018  
Date

*10/23/18  
JD*

80-000998  
 Std. by sheets  
 10/23/18

INTOXILYZER 8000  
 Instrument Initialization  
 06:37 10/23/2018

SEMINOLE COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000998  
 10/23/2018  
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	07:09
Control Test	0.048	07:10
Air Blank	0.000	07:11
Control Test	0.048	07:11
Air Blank	0.000	07:12
Control Test	0.048	07:12
Air Blank	0.000	07:13
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

SEMINOLE COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000998  
 10/23/2018  
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	07:15
Control Test	0.080	07:15
Air Blank	0.000	07:16
Control Test	0.080	07:17
Air Blank	0.000	07:17
Control Test	0.080	07:18
Air Blank	0.000	07:18
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

SEMINOLE COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000998  
 10/23/2018  
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	07:21
Control Test	0.197	07:22
Air Blank	0.000	07:22
Control Test	0.197	07:23
Air Blank	0.000	07:24
Control Test	0.198	07:24
Air Blank	0.000	07:25
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

Operator's Signature

SEMINOLE COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000998  
 10/23/2018  
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	07:06
Control Test	0.080	07:06
Air Blank	0.000	07:06
Control Test	0.079	07:07
Air Blank	0.000	07:07
Control Test	0.079	07:08
Air Blank	0.000	07:08
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

JCS

Operator's Signature

10/23/18

*[Handwritten mark]*



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000998, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000998</u>	UNCERTAINTY* ±	
Owning Agency:	<u>SEMINOLE COUNTY S.O.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/23/2018</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>09:09</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 July 2018  
Issuing Authority: Alcohol Testing Program

10/23/2018

Date

JAKE L SHANAHAN,  
Department Inspector

Service • Integrity • Respect • Quality

*10/23/18*  
*[Signature]*

*[Signature]*

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SEMINOLE COUNTY S.O.  
Time of Inspection: 07:05

Date of Inspection: 10/23/2018

Serial Number: 80-000998  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_


Remarks:  
BYPASS TO OPERATE

NOT a compliance check

KUB

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name

10/23/2018  
Date

10/23/18  
JL