



INSTRUMENT PROCESSING SHEET

Agency Satellite Beach PD

S/N 80-001001

Florida Department of Law Enforcement

Date In 12/22/17

DI Completion Date 1/10/18

Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>191</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.160</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>286602</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
---	--	--

Final Release Date  
**FDLE**  
**JAN 12 2018**  
**Alcohol Testing Program**

Simulator	Serial #	Lot #/Exp
0.050	SD3962	201707D 7-25-19
0.080	SD3964	201707E 7-25-19
0.200	DR3856	201707C 7-24-19
0.080 DGS	N/A	AG6266004 9-22-18

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By SP

Lab Temp °C 22.5  
 External Digital Therm. ID#: 300505  
 34°C +/- .2 Serial #: SD3962  
 34°C +/- .2 Serial #: SD3964  
 34°C +/- .2 Serial #: DR3856

**Calibration Adjustment** Performed By \_\_\_\_\_

Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

**Department Inspection** Performed By SP

Barometric Pressure ID# 286602  
 Gauge 1019 Instrument 1020  
 Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	SD1019
Interferent	SD1021
0.050	SD3962
0.080	SD3964
0.200	DR3856

**Attachments**

Form 41  Flow Calibration  
 Stability Checks  Form 40  
 Calibration Adjustment  Other Calibration Certificate  
 Post-Stability Checks

Notes/Suggested Service: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Tech Review / Date 1/10/18 J. Durham Admin Review / Date 1/12/18



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SATELLITE BEACH P.D.  
Time of Inspection: 12:01

Date of Inspection: 01/10/2018

Serial Number: 80-001001  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

*See*

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.049	0.080	0.201	0.000 / 0.080
0.000	0.049	0.080	0.201	0.000 / 0.080
0.000	0.049	0.080	0.201	0.000 / 0.079
0.000	0.049	0.080	0.201	0.000 / 0.079
0.000	0.050	0.080	0.201	0.000 / 0.079
0.000	0.049	0.080	0.201	0.000 / 0.079
0.000	0.050	0.080	0.201	0.000 / 0.080
0.000	0.049	0.080	0.201	0.000 / 0.079
0.000	0.049	0.080	0.201	0.000 / 0.079
0.000	0.049	0.081	0.201	0.000 / 0.079
Standard Deviations	0.0004	0.0003	0.0000	0.0000 / 0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance- TANK NOT SCREWED IN PROPERLY.

*10*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla D Platt*  
*SP*

SHAYLA D PLATT

Signature and Printed Name

01/10/2018  
Date

*1/12/18*  
*JD*

STABILITY CHECKS - # 80-001001

SATELLITE BEACH P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001001  
 01/10/2018  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:17
Control Test	0.049	09:18
Air Blank	0.000	09:18
Control Test	0.049	09:19
Air Blank	0.000	09:19
Control Test	0.048	09:20
Air Blank	0.000	09:21
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

SP

Operator's Signature

8/12/11  
 DS

SATELLITE BEACH P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001001  
 01/10/2018  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:12
Control Test	0.081	09:13
Air Blank	0.000	09:13
Control Test	0.080	09:14
Air Blank	0.000	09:14
Control Test	0.080	09:15
Air Blank	0.000	09:16
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

SP

Operator's Signature

SATELLITE BEACH P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001001  
 01/10/2018  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:06
Control Test	0.202	09:07
Air Blank	0.000	09:07
Control Test	0.201	09:08
Air Blank	0.000	09:09
Control Test	0.200	09:09
Air Blank	0.000	09:10
Control Test Stats		
Average	0.2010	
Std Dev	0.0010	
Rel Std Dev(%)	0.4975	

SP

Operator's Signature

SATELLITE BEACH P.D.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001001  
 01/10/2018  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:48
Control Test	0.080	08:49
Air Blank	0.000	08:49
Control Test	0.079	08:49
Air Blank	0.000	08:50
Control Test	0.080	08:50
Air Blank	0.000	08:51
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

SP

Operator's Signature





Florida Department of Law Enforcement  
 Alcohol Testing Program  
 2729 Fort Knox Blvd.  
 Bldg. 2, Suite 1300  
 Tallahassee, FL 32308

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001001, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001001</u>	UNCERTAINTY* ±	
Owning Agency:	<u>SATELLITE BEACH P.D.</u>	0.05 g/ 210 L	0.004
Calibration Date:	<u>01/10/2018</u>	0.08 g/ 210 L	0.005
Calibration Time:	<u>12:01</u>	0.20 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
 Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Target Alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/10/2018 Date Shayla Platt

SHAYLA D PLATT,  
 Department Inspector

FDLE/ATP Form 69 January 2018  
 Issuing Authority: Alcohol Testing Program

1/12/18  
 JD