FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Winter Haven Police Department S/N 80-001060

Date In 03/09/2018 DI Completion Date 03/14/2018 ☑Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By TG

- ☑ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ☑ Case
- ☑ Handle
- ☑ Keyboard
- ☑ Dry Gas Shelf
- ☑ Feet
- ☑ Breath Tube
- ☑ Ports
- ☑ Screws Tight

Other Equipment/ Accessories:

- ☑ Power cord
- ☐ Printer Cable
- ☐ Static Bag
- ☐ 12V DC Cable

Notes:

Final Release Date

FDLE

MAR 15 2018

Alcohol Testing Program

Quality Checks Performed By DMB

- ☑ Breath Tube Screen
- ☑ Replace External O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 222
- ☑ Flow Verification (L/s)

Flow Column # ATP 103

32 mm 0.140 (.139 - .169)
36 mm 0.156 (.156 - .190)
53 mm 0.226 (.228 - .278)
103 mm 0.503 (.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 28662

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	G11739	201707D 07/25/2019
0.080	SD3964	201707E 07/25/2019
0.200	DR3856	201707C 07/24/2019
0.080 DGS	N/A	AG715202 06/01/2019

Flow Calibration Performed By DMB

Flow Column # ATP 102

- ☑ 5L/min – 17mm
- ☑ 15L/min – 53mm
- ☑ 30L/min – 103mm
- ☑ R-Value 220
- ☑ Post Calibration Verification (L/s)

Flow Column # ATP 103

32 mm 0.152 (.139 - .169)
36 mm 0.167 (.156 - .190)
53 mm 0.238 (.228 - .278)
103 mm 0.523 (.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Temperature Checks Performed By DMB

- ☑ Lab Temp °C 22.6

External Digital Therm. ID#: 300503

- ☑ 34°C +-.2 Serial #: G11739
- ☑ 34°C +-.2 Serial #: SD3964
- ☑ 34°C +-.2 Serial #: DR3856

Calibration Adjustment Performed By DMB

Barometric Pressure Gauge 1020 ID # 28427

Simulator	Serial Number	Lot Number	Expiration
0.000	G2879	N/A	N/A
0.040	SD1024	16320	10/21/2018
0.100	G2834	17280	09/11/2019
0.200	SD1011	17090	02/24/2019
0.300	DR1275	17140	05/15/2019
0.080 DGS	N/A	22817080A5	10/05/2019

☑ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050	G11739	201707D	07/25/2019
0.080	SD3964	201707E	07/25/2019
0.200	DR3856	201707C	07/24/2019
0.080 DGS	N/A	AG715202	06/01/2019

Department Inspection Performed By DMB

Barometric Pressure ID# 28662

Gauge 1021 Instrument 1020

Mouth Alcohol Solution Lot # 2016-C

Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	SD1019
Interferent	SD1021
0.050	G11739
0.080	SD3964
0.200	DR3856

Attachments

- ☑ Form 41
- ☑ Stability Checks
- ☑ Calibration Certificate
- ☑ Calibration Adjustment
- ☑ Post-Stability Checks
- ☑ Flow Calibration
- ☐ Form 40
- ☑ Other Addtl stabilities

Notes/Suggested Service: Stabililty checks were run to replicate the INT/Purge fail intermittent issue. I was able to replicate it after roughly 80+ checks. Add'tl checks were run after the Cal Adjust. Optical bench calibration adjustment performed to eliminate the INT and purge fail intermittent issue and bring values closer to nominal. 3/14/18 DMB

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

3/15/18 — Tech Review / Date

3/15/18 — Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WINTER HAVEN PD
Time of Inspection: 14:43
Date of Inspection: 03/14/2018
Serial Number: 80-001060
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG715202 Exp: 06/01/2019
0.000	0.049	0.081	0.200	0.080
0.000	0.050	0.082	0.201	0.079
0.000	0.050	0.082	0.202	0.079
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.082	0.202	0.079
0.000	0.051	0.081	0.202	0.079
0.000	0.050	0.081	0.202	0.079
0.000	0.051	0.081	0.202	0.079
Standard Deviations	0.0005	0.0004	0.0006	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIELLE M BELL
Signature and Printed Name

03/14/2018
Date

3/15/18

FDLE/ATP Form 41 –Revised August 2005

Stability Checks #80-001060 Winter Haven P.D. 3/13/18 DMB

0.05 g/210 L

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:38
Control Test	INT*	15:39
Air Blank	PUR**	15:40
Air Blank	PUR**	15:41

*Interferent Detect
**Purge Fail

DMB
Operator's Signature

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:41
Control Test	0.051	15:42
Air Blank	0.000	15:42
Control Test	0.051	15:43
Air Blank	0.000	15:44
Control Test	0.050	15:44
Air Blank	0.000	15:45
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

DMB
Operator's Signature

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:46
Control Test	0.081	15:47
Air Blank	0.000	15:47
Control Test	0.081	15:48
Air Blank	0.000	15:48
Control Test	0.082	15:49
Air Blank	0.000	15:50
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

DMB
Operator's Signature

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:51
Control Test	0.196	15:51
Air Blank	0.000	15:52
Control Test	0.199	15:53
Air Blank	0.000	15:53
Control Test	0.200	15:54
Air Blank	0.000	15:54
Control Test Stats		
Average	0.1983	
Std Dev	0.0021	
Rel Std Dev(%)	1.0496	

DMB
Operator's Signature

DGS

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:58
Control Test	0.084	15:59
Air Blank	0.000	15:59
Control Test	0.083	15:59
Air Blank	0.000	16:00
Control Test	0.084	16:00
Air Blank	0.000	16:01
Control Test Stats		
Average	0.0837	
Std Dev	0.0006	
Rel Std Dev(%)	0.6901	

DMB
Operator's Signature

3/15/18

FDLE

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001060, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001060		UNCERTAINTY* ±
Owning Agency:	WINTER HAVEN PD	0.050 g/ 210 L	0.004
Calibration Date:	03/14/2018	0.080 g/ 210 L	0.005
Calibration Time:	14:43	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/14/2018
Date

DANIELLE M BELL,
Department Inspector

3/15/18

Optical bench calibration adjustment data #80-001060 Winter Haven P.D. 3/14/18

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/14/2018 09:53:18

Auto Calibration
Max Power Res Value = 21
Auto Range Res Value = 10

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12948, 9um Io = 14410

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1270	(-0.0120)
Sample #2 =	0.1080	(-0.0080)
Sample #3 =	0.1130	(0.0020)
Sample #4 =	0.1340	(0.0280)

Avg % Abs = 0.1183 (0.0073)
STD DEV = 0.0138 (0.0186)
REL STD DEV = 11.659 (253.407)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1440	(-0.0200)
Sample #2 =	0.1230	(0.0010)
Sample #3 =	0.1210	(0.0030)
Sample #4 =	0.1180	(0.0150)

Avg % Abs = 0.1207 (0.0063)
STD DEV = 0.0025 (0.0076)
REL STD DEV = 2.086 (119.556)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12943, 9um Io = 14404

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.8220	(-0.0040)
Sample #2 =	0.8410	(0.0240)
Sample #3 =	0.8330	(0.0110)
Sample #4 =	0.8420	(0.0170)

Avg % Abs = 0.8387 (0.0173)
STD DEV = 0.0049 (0.0065)
REL STD DEV = 0.588 (37.537)

3/15/18 JP

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.5650	(-0.0170)
Sample #2 =	1.5720	(-0.0030)
Sample #3 =	1.5670	(-0.0100)
Sample #4 =	1.5760	(-0.0190)

Avg % Abs = 1.5717 (-0.0107)
STD DEV = 0.0045 (0.0080)
REL STD DEV = 0.287 (75.195)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12939, 9um Io = 14406

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.9250	(-0.0200)
Sample #2 =	1.9020	(0.0390)
Sample #3 =	1.9100	(0.0290)
Sample #4 =	1.9160	(0.0540)

Avg % Abs = 1.9093 (0.0407)
STD DEV = 0.0070 (0.0126)
REL STD DEV = 0.368 (30.942)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.7260	(0.0010)
Sample #2 =	3.7320	(0.0440)
Sample #3 =	3.7280	(0.0410)
Sample #4 =	3.7250	(0.0600)

Avg % Abs = 3.7283 (0.0483)
STD DEV = 0.0035 (0.0102)
REL STD DEV = 0.094 (21.133)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12932, 9um Io = 14402

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.6870	(-0.0140)
Sample #2 =	3.6670	(0.0060)
Sample #3 =	3.6720	(0.0410)
Sample #4 =	3.6790	(0.0370)

Avg % Abs = 3.6727 (0.0280)
STD DEV = 0.0060 (0.0192)
REL STD DEV = 0.164 (68.419)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	7.1890	(-0.0030)
Sample #2 =	7.1470	(0.0760)
Sample #3 =	7.1520	(0.0920)
Sample #4 =	7.1500	(0.0980)

Avg % Abs = 7.1497 (0.0887)
STD DEV = 0.0025 (0.0114)
REL STD DEV = 0.035 (12.826)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12930, 9um Io = 14397

<<<<< CHANNEL 1 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	5.3530	(-0.0250)
Sample #2 =	5.3370	(0.0220)
Sample #3 =	5.3170	(0.0340)
Sample #4 =	5.2910	(0.0620)

Avg % Abs = 5.3150 (0.0393)
STD DEV = 0.0231 (0.0205)
REL STD DEV = 0.434 (52.186)

<<<<< CHANNEL 2 >>>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	10.3530	(-0.0280)
Sample #2 =	10.3320	(0.0590)
Sample #3 =	10.3230	(0.0830)
Sample #4 =	10.2980	(0.1060)

Avg % Abs = 10.3177 (0.0827)
STD DEV = 0.0176 (0.0235)
REL STD DEV = 0.171 (28.430)

***** AUTO CAL DATA *****

<<<<< CHANNEL 1 >>>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.118
Std Dev = 0.01 Rel Std Dev = 11.66
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.839
Std Dev = 0.00 Rel Std Dev = 0.59
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.909
Std Dev = 0.01 Rel Std Dev = 0.37
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.673
Std Dev = 0.01 Rel Std Dev = 0.16
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.315
Std Dev = 0.02 Rel Std Dev = 0.43
Zero Order Coef = -280.59
First Order Coef = 2558.12
Second Order Coef = 33.73
Standard Deviation = 33.137852

<<<<< CHANNEL 2 >>>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.121
Std Dev = 0.00 Rel Std Dev = 2.09
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.572
Std Dev = 0.00 Rel Std Dev = 0.29
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.728
Std Dev = 0.00 Rel Std Dev = 0.09
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.150
Std Dev = 0.00 Rel Std Dev = 0.04
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.318
Std Dev = 0.02 Rel Std Dev = 0.17
Zero Order Coef = -131.76
First Order Coef = 1259.70
Second Order Coef = 13.23
Standard Deviation = 22.618065

Solution Stats Quadratic Fit Chan 1

Act g/210L	Fit g/210L	Residual g/210L
0.000	0.000	-0.0005
0.040	0.040	0.0003
0.100	0.099	0.0007
0.200	0.201	-0.0010
0.300	0.300	0.0004

Solution Stats Quadratic Fit Chan 2

Act g/210L	Fit g/210L	Residual g/210L
0.000	0.000	-0.0004
0.040	0.039	0.0005
0.100	0.100	0.0003
0.200	0.201	-0.0006
0.300	0.300	0.0003

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3211.00
Sample #2 = 3244.00
Sample #3 = 3333.00
Sample #4 = 3280.00
Average Result = 3285.6667
STD DEV = 44.7698
REL STD DEV = 1.363
****
***** CHANNEL 2
Sample #1 = 3441.00
Sample #2 = 3453.00
Sample #3 = 3499.00
Sample #4 = 3521.00
Average Result = 3491.0000
STD DEV = 34.6987
REL STD DEV = 0.994
****
Dry Gas H20 Adjust Results **********
Barometric Pressure = 1020
3 um H20 Adjust (mg/l*10,000) = 524
9 um H20 Adjust (mg/l*10,000) = 318
**** AUTO CAL PASS

Post Calibration Adjustment
Stability Checks #80-001060 Winter Haven P.D. 3/14/18 DGS

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:29
Control Test	0.050	11:30
Air Blank	0.000	11:30
Control Test	0.050	11:31
Air Blank	0.000	11:31
Control Test	0.050	11:32
Air Blank	0.000	11:33
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

3/15/18

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:34
Control Test	0.082	11:35
Air Blank	0.000	11:35
Control Test	0.082	11:36
Air Blank	0.000	11:37
Control Test	0.081	11:37
Air Blank	0.000	11:38
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

Operator's Signature

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:39
Control Test	0.201	11:40
Air Blank	0.000	11:40
Control Test	0.200	11:41
Air Blank	0.000	11:42
Control Test	0.200	11:42
Air Blank	0.000	11:43
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

Operator's Signature

DGS

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:44
Control Test	0.080	11:45
Air Blank	0.000	11:45
Control Test	0.080	11:45
Air Blank	0.000	11:46
Control Test	0.080	11:46
Air Blank	0.000	11:47
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Flow Calibration
Adjustment Data
#80-001060
Winter Haven P.D.
3/13/18
DMB

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/13/2018
Software: 8100.27

Flow Rate Calibration********
1: Rate (Liters/min) = 5
SQRT(Diff)) = 5.926
2: Rate (Liters/min) = 15
SQRT(Diff)) = 11.445
3: Rate (Liters/min) = 30
SQRT(Diff)) = 20.418
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 714
Rounded Intercept = -696028
Correlation = 0.99731

Test Breath Test
Looking for INT/Purge Fails
#80-001060
Winter Haven P.D.
3/13/18
DMB

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/13/2018
Software: 8100.27

LAI Date = 03/07/2018
Obs Period Began = 14:25
Sub Name =
TEST

TEST
Sub DOB = 01/01/1960
Sub Sex = M
Driv Lic = SAMPLE BREATH TEST /FL
Officer =
DANIELLE
BELL
Operator =
DANIELLE, M
BELL

Test	g/210L	Time
Diagnostics Check	OK	14:48
Air Blank	0.000	14:48
Control Test	0.081	14:49
Air Blank	0.000	14:49
Subject Sample #1	0.000	14:50
Air Blank	0.000	14:50
Air Blank	0.000	14:52
Subject Sample #2	0.000	14:53
Air Blank	0.000	14:53
Control Test	0.082	14:53
Air Blank	0.000	14:54
Diagnostics Check	OK	14:54

Operator's Signature

3/15/18
JD

Stability Checks
Looking for Purge Fails & INT Detect (when not introducing INT)
80-001060
Winter Haven Police Dept.
3/13/18 DMB

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:36
Control Test	0.200	12:36
Air Blank	0.000	12:37
Control Test	0.199	12:38
Air Blank	0.000	12:38
Control Test	0.199	12:39
Air Blank	0.000	12:39
Control Test	0.200	12:40
Air Blank	0.000	12:41
Control Test	0.198	12:41
Air Blank	0.000	12:42
Control Test	0.200	12:43
Air Blank	0.000	12:43
Control Test	0.200	12:44
Air Blank	0.000	12:44
Control Test	0.200	12:45
Air Blank	0.000	12:46
Control Test	0.199	12:46
Air Blank	0.000	12:47
Control Test	0.199	12:48
Air Blank	0.000	12:48

Control Test Stats

Average	0.1994
Std Dev	0.0007
Rel Std Dev(%)	0.3507

DMB
Operator's Signature

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:54
Control Test	0.000	12:55
Air Blank	0.000	12:55
Control Test	0.000	12:56
Air Blank	0.000	12:56
Control Test	0.000	12:57
Air Blank	0.000	12:58
Control Test	0.000	12:58
Air Blank	0.000	12:59
Control Test	0.000	12:59
Air Blank	0.000	13:00
Control Test	0.000	13:01
Air Blank	0.000	13:01
Control Test	0.000	13:02
Air Blank	0.000	13:02
Control Test	0.000	13:03
Air Blank	0.000	13:03
Control Test	0.000	13:04
Air Blank	0.000	13:05
Control Test	0.000	13:05
Air Blank	0.000	13:06
Control Test	0.000	13:06
Air Blank	0.000	13:07
Control Test	0.000	13:08
Air Blank	0.000	13:08
Control Test	0.000	13:09
Air Blank	0.000	13:09
Control Test	0.000	13:10
Air Blank	0.000	13:10
Control Test	0.000	13:11
Air Blank	0.000	13:11
Control Test	0.000	13:12
Air Blank	0.000	13:13
Control Test	0.000	13:13
Air Blank	0.000	13:14
Control Test	0.000	13:14
Air Blank	0.000	13:15
Control Test	0.000	13:16
Air Blank	0.000	13:16
Control Test	0.000	13:17
Air Blank	0.000	13:17

Control Test Stats

Average	0.0000
Std Dev	0.0000
Rel Std Dev(%)	0.0000

DMB
Operator's Signature

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:18
Control Test	0.000	13:19
Air Blank	0.000	13:19
Control Test	0.000	13:20
Air Blank	0.000	13:21
Control Test	0.000	13:21
Air Blank	0.000	13:22
Control Test	0.000	13:22
Air Blank	0.000	13:23
Control Test	0.000	13:24
Air Blank	0.000	13:24
Control Test	0.000	13:25
Air Blank	0.000	13:25
Control Test	0.000	13:26
Air Blank	0.000	13:26
Control Test	0.000	13:27
Air Blank	0.000	13:28
Control Test	0.000	13:28
Air Blank	0.000	13:29
Control Test	0.000	13:29
Air Blank	0.000	13:30
Control Test	0.000	13:31
Air Blank	0.000	13:31
Control Test	0.000	13:32
Air Blank	0.000	13:32
Control Test	0.000	13:33
Air Blank	0.000	13:33
Control Test	0.000	13:34
Air Blank	0.000	13:34
Control Test	0.000	13:35
Air Blank	0.000	13:36
Control Test	0.000	13:36
Air Blank	0.000	13:37
Control Test	0.000	13:37
Air Blank	0.000	13:38
Control Test	0.000	13:39
Air Blank	0.000	13:39
Control Test	0.000	13:40
Air Blank	0.000	13:40
Control Test	0.000	13:41
Air Blank	0.000	13:41

Control Test Stats

Average	0.0000
Std Dev	0.0000
Rel Std Dev(%)	0.0000

3/15/18 DB

DMB
Operator's Signature

Additional Stability Checks
#80-001060 Winter Haven 3/13/18 RWB

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:52
Control Test	INT*	13:53
Air Blank	0.000	13:54
Control Test	INT*	13:54
Air Blank	0.000	13:55
Control Test	INT*	13:55
Air Blank	0.000	13:56
Control Test	INT*	13:57
Air Blank	0.000	13:57
Control Test	INT*	13:58
Air Blank	0.000	13:59
Control Test	INT*	13:59
Air Blank	0.000	14:00
Control Test	INT*	14:01
Air Blank	0.000	14:01
Control Test	INT*	14:02
Air Blank	0.000	14:02
Control Test	INT*	14:03
Air Blank	0.000	14:04
Control Test	INT*	14:04
Air Blank	0.000	14:05
Control Test	INT*	14:06
Air Blank	0.000	14:06
Control Test	INT*	14:07
Air Blank	0.000	14:08
Control Test	INT*	14:08
Air Blank	0.000	14:09
Control Test	INT*	14:09
Air Blank	0.000	14:10
Control Test	INT*	14:11
Air Blank	0.000	14:11
Control Test	INT*	14:12
Air Blank	0.000	14:13
Control Test	INT*	14:13
Air Blank	0.000	14:14
Control Test	INT*	14:15
Air Blank	0.000	14:15
Control Test	INT*	14:16
Air Blank	0.000	14:16
Control Test	INT*	14:17
Air Blank	0.000	14:18

Control Test Stats

Average	0.0000
Std Dev	0.0000
Rel Std Dev(%)	0.0000

*Interferent Detect

RWB
Operator's Signature

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:21
Control Test	0.000	14:22
Air Blank	0.000	14:22
Control Test	0.000	14:23
Air Blank	0.000	14:24
Control Test	0.000	14:24
Air Blank	0.000	14:25
Control Test	0.000	14:25
Air Blank	0.000	14:26
Control Test	0.000	14:27
Air Blank	0.000	14:27
Control Test	0.000	14:28
Air Blank	0.000	14:28
Control Test	0.000	14:29
Air Blank	0.000	14:29
Control Test	0.000	14:30
Air Blank	0.000	14:31
Control Test	0.000	14:31
Air Blank	0.000	14:32
Control Test	0.000	14:32
Air Blank	0.000	14:33
Control Test	0.000	14:34
Air Blank	0.000	14:34
Control Test	0.000	14:35
Air Blank	0.000	14:35
Control Test	0.000	14:36
Air Blank	0.000	14:36
Control Test	0.000	14:37
Air Blank	0.000	14:37
Control Test	0.000	14:38
Air Blank	0.000	14:39
Control Test	0.000	14:39
Air Blank	0.000	14:40
Control Test	0.000	14:40
Air Blank	0.000	14:41
Control Test	0.000	14:42
Air Blank	0.000	14:42
Control Test	0.000	14:43
Air Blank	0.000	14:43
Control Test	0.000	14:44
Air Blank	0.000	14:44

Control Test Stats

Average	0.0000
Std Dev	0.0000
Rel Std Dev(%)	0.0000

3/15/18
JD

RWB
Operator's Signature

Additional Stabilities Post Calibration Adjustment
#80-001060 Winter Haven P.D. 3/14/18 DMB

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:46
Control Test	0.049	14:47
Air Blank	0.000	14:47
Control Test	0.050	14:48
Air Blank	0.000	14:48
Control Test	0.050	14:49
Air Blank	0.000	14:50
Control Test	0.050	14:50
Air Blank	0.000	14:51
Control Test	0.050	14:52
Air Blank	0.000	14:52
Control Test	0.050	14:53
Air Blank	0.000	14:53
Control Test	0.050	14:54
Air Blank	0.000	14:54
Control Test	0.050	14:55
Air Blank	0.000	14:56
Control Test	0.050	14:56
Air Blank	0.000	14:57
Control Test	0.050	14:58
Air Blank	0.000	14:58
Control Test	0.049	14:59
Air Blank	0.000	14:59
Control Test	0.049	15:00
Air Blank	0.000	15:01
Control Test	0.049	15:01
Air Blank	0.000	15:02
Control Test	0.049	15:03
Air Blank	0.000	15:03
Control Test	0.050	15:04
Air Blank	0.000	15:04
Control Test	0.050	15:05
Air Blank	0.000	15:06
Control Test	0.049	15:06
Air Blank	0.000	15:07
Control Test	0.050	15:07
Air Blank	0.000	15:08
Control Test	0.050	15:09
Air Blank	0.000	15:09
Control Test	0.050	15:10
Air Blank	0.000	15:10

Control Test Stats

Average	0.0497
Std Dev	0.0005
Rel Std Dev(%)	0.9460

Operator's Signature

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:23
Control Test	INT*	15:24
Air Blank	0.000	15:24
Control Test	INT*	15:25
Air Blank	0.000	15:26
Control Test	INT*	15:26
Air Blank	0.000	15:27
Control Test	INT*	15:28
Air Blank	0.000	15:28
Control Test	INT*	15:29
Air Blank	0.000	15:29
Control Test	INT*	15:30
Air Blank	0.000	15:31
Control Test	INT*	15:31
Air Blank	0.000	15:32
Control Test	INT*	15:32
Air Blank	0.000	15:33
Control Test	INT*	15:34
Air Blank	0.000	15:34
Control Test	INT*	15:35
Air Blank	0.000	15:35
Control Test	INT*	15:36
Air Blank	0.000	15:37
Control Test	INT*	15:37
Air Blank	0.000	15:38
Control Test	INT*	15:39
Air Blank	0.000	15:39
Control Test	INT*	15:40
Air Blank	0.000	15:40
Control Test	INT*	15:41
Air Blank	0.000	15:42
Control Test	INT*	15:42
Air Blank	0.000	15:43
Control Test	INT*	15:43
Air Blank	0.000	15:44
Control Test	INT*	15:45
Air Blank	0.000	15:45
Control Test	INT*	15:46
Air Blank	0.000	15:46
Control Test	INT*	15:47
Air Blank	0.000	15:48

Control Test Stats

Average	0.0000
Std Dev	0.0000
Rel Std Dev(%)	0.0000

*Interferent Detect

Operator's Signature

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:55
Control Test	0.000	15:55
Air Blank	0.000	15:56
Control Test	0.000	15:57
Air Blank	0.000	15:57
Control Test	0.000	15:58
Air Blank	0.000	15:58
Control Test	0.000	15:59
Air Blank	0.000	15:59
Control Test	0.000	16:00
Air Blank	0.000	16:00
Control Test	0.000	16:01
Air Blank	0.000	16:02
Control Test	0.000	16:02
Air Blank	0.000	16:03
Control Test	0.000	16:03
Air Blank	0.000	16:04
Control Test	0.000	16:05
Air Blank	0.000	16:05
Control Test	0.000	16:06
Air Blank	0.000	16:06
Control Test	0.000	16:07
Air Blank	0.000	16:07
Control Test	0.000	16:08
Air Blank	0.000	16:09
Control Test	0.000	16:09
Air Blank	0.000	16:10
Control Test	0.000	16:10
Air Blank	0.000	16:11
Control Test	0.000	16:12
Air Blank	0.000	16:12
Control Test	0.000	16:13
Air Blank	0.000	16:13
Control Test	0.000	16:14
Air Blank	0.000	16:14
Control Test	0.000	16:15
Air Blank	0.000	16:15
Control Test	0.000	16:16
Air Blank	0.000	16:17
Control Test	0.000	16:17
Air Blank	0.000	16:18

Control Test Stats

Average	0.0000
Std Dev	0.0000
Rel Std Dev(%)	0.0000

3/15/18
JO

Operator's Signature

Additional Stabilities to confirm no INT or purge fails unless an INT is introduced. #80-001060 Winter Haven P.D. 3/14/18 DMB

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:19
Control Test	0.000	16:20
Air Blank	0.000	16:20
Control Test	0.000	16:21
Air Blank	0.000	16:22
Control Test	0.000	16:22
Air Blank	0.000	16:23
Control Test	0.000	16:23
Air Blank	0.000	16:24
Control Test	0.000	16:25
Air Blank	0.000	16:25
Control Test	0.000	16:26
Air Blank	0.000	16:26
Control Test	0.000	16:27
Air Blank	0.000	16:27
Control Test	0.000	16:28
Air Blank	0.000	16:28
Control Test	0.000	16:29
Air Blank	0.000	16:30
Control Test	0.000	16:30
Air Blank	0.000	16:31
Control Test	0.000	16:31
Air Blank	0.000	16:32
Control Test	0.000	16:33
Air Blank	0.000	16:33
Control Test	0.000	16:34
Air Blank	0.000	16:34
Control Test	0.000	16:35
Air Blank	0.000	16:35
Control Test	0.000	16:36
Air Blank	0.000	16:37
Control Test	0.000	16:37
Air Blank	0.000	16:38
Control Test	0.000	16:38
Air Blank	0.000	16:39
Control Test	0.000	16:40
Air Blank	0.000	16:40
Control Test	0.000	16:41
Air Blank	0.000	16:41
Control Test	0.000	16:42
Air Blank	0.000	16:42

Control Test Stats

Average	0.0000
Std Dev	0.0000
Rel Std Dev(%)	0.0000

DMB
Operator's Signature

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:45
Control Test	0.000	16:46
Air Blank	0.000	16:46
Control Test	0.000	16:47
Air Blank	0.000	16:47
Control Test	0.000	16:48
Air Blank	0.000	16:49
Control Test	0.000	16:49
Air Blank	0.000	16:50
Control Test	0.000	16:50
Air Blank	0.000	16:51
Control Test	0.000	16:52
Air Blank	0.000	16:52
Control Test	0.000	16:53
Air Blank	0.000	16:53
Control Test	0.000	16:54
Air Blank	0.000	16:54
Control Test	0.000	16:55
Air Blank	0.000	16:56
Control Test	0.000	16:56
Air Blank	0.000	16:57
Control Test	0.000	16:57
Air Blank	0.000	16:58
Control Test	0.000	16:59
Air Blank	0.000	16:59
Control Test	0.000	17:00
Air Blank	0.000	17:00
Control Test	0.000	17:01
Air Blank	0.000	17:01
Control Test	0.000	17:02
Air Blank	0.000	17:02
Control Test	0.000	17:03
Air Blank	0.000	17:04
Control Test	0.000	17:04
Air Blank	0.000	17:05
Control Test	0.000	17:05
Air Blank	0.000	17:06
Control Test	0.000	17:07
Air Blank	0.000	17:07
Control Test	0.000	17:08
Air Blank	0.000	17:08

Control Test Stats

Average	0.0000
Std Dev	0.0000
Rel Std Dev(%)	0.0000

DMB
Operator's Signature

WINTER HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001060
03/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	17:09
Control Test	0.000	17:10
Air Blank	0.000	17:10
Control Test	0.000	17:11
Air Blank	0.000	17:11
Control Test	0.000	17:12
Air Blank	0.000	17:12
Control Test	0.000	17:13
Air Blank	0.000	17:13
Control Test	0.000	17:14
Air Blank	0.000	17:15
Control Test	0.000	17:15
Air Blank	0.000	17:16
Control Test	0.000	17:16
Air Blank	0.000	17:17
Control Test	0.000	17:18
Air Blank	0.000	17:18
Control Test	0.000	17:19
Air Blank	0.000	17:19
Control Test	0.000	17:20
Air Blank	0.000	17:20
Control Test	0.000	17:21
Air Blank	0.000	17:22
Control Test	0.000	17:22
Air Blank	0.000	17:23
Control Test	0.000	17:23
Air Blank	0.000	17:24
Control Test	0.000	17:25
Air Blank	0.000	17:25
Control Test	0.000	17:26
Air Blank	0.000	17:26
Control Test	0.000	17:27
Air Blank	0.000	17:27
Control Test	0.000	17:28
Air Blank	0.000	17:28
Control Test	0.000	17:29
Air Blank	0.000	17:30
Control Test	0.000	17:30
Air Blank	0.000	17:31
Control Test	0.000	17:31
Air Blank	0.000	17:32

Control Test Stats

Average	0.0000
Std Dev	0.0000
Rel Std Dev(%)	0.0000

3/15/18
JD

DMB
Operator's Signature