FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Deland Police Department S/N 80-001114

Date In 07/02/2018 DI Completion Date 7/6/2018 ☒Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By SDS

- ☒ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

☒ Case	☒ Handle
☒ Keyboard	☒ Dry Gas Shelf
☒ Feet	☒ Breath Tube
☒ Ports	☒ Screws Tight

Other Equipment/ Accessories:

☐ Power cord	☐ Printer Cable
☐ Static Bag	☐ 12V DC Cable

Notes: Note from AI - 1. Clock keeps getting off 2. Dry gas Control Tol is too low - Avg 0.076

Final Release Date

FDLE

JUL 09 2018

Alcohol Testing Program

Quality Checks Performed By PJM

- ☒ Breath Tube Screen
- ☒ Replace External O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 207
- ☒ Flow Verification (L/s)

Flow Column # ATP102

32 mm	.148	(.139 - .169)
36 mm	.162	(.156 - .190)
53 mm	.234	(.228 - .278)
103 mm	.492	(.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 28427

- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD1018	201707D
		07/25/2019
0.080	SD3962	201707E
		07/25/2019
0.200	G2078	201707C
		07/24/2019
0.080 DGS	N/A	A6805702
		2/26/20

Flow Calibration Performed By

Flow Column #

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)

Flow Column #

32 mm		(.139 - .169)
36 mm		(.156 - .190)
53 mm		(.228 - .278)
103 mm		(.447 - .547)

Maintenance Performed By SP

- ☒ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Temperature Checks Performed By PJM

- ☒ Lab Temp °C 21.4

External Digital Therm. ID#: 300503

- ☒ 34°C +-.2 Serial #: SD1021
- ☒ 34°C +-.2 Serial #: DR1275
- ☒ 34°C +-.2 Serial #: SD1019

Calibration Adjustment Performed By

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By PJM

Barometric Pressure ID# 28427

Gauge 1019 Instrument 1018

Mouth Alcohol Solution Lot # 2017-B

Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	G11621
Interferent	DR3855
0.050	SD1021
0.080	DR1275
0.200	SD1019

Attachments

☒ Form 41	☐ Post-Stability Checks
☒ Stability Checks	☐ Flow Calibration
☒ Calibration Certificate	☐ Form 40
☐ Calibration Adjustment	☐ Other

Notes/Suggested Service: AI, REQUESTED CAL ADJ. BUT VALUES WERE EXCELLENT. NO CAL ADJ. PERFORMED.

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

7/9/18 Tech Review / Date

7/9/18 Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DELAND P.D. Serial Number: 80-001114
Time of Inspection: 11:45 Date of Inspection: 07/06/2018 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.080	0.198	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.077
0.000	0.049	0.081	0.199	0.078
0.000	0.049	0.080	0.199	0.077
0.000	0.049	0.080	0.199	0.077
0.000	0.049	0.080	0.200	0.077
0.000	0.049	0.080	0.199	0.077
0.000	0.049	0.080	0.200	0.077

Standard Deviations	0.0004	0.0004	0.0008	0.0007

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy PATRICK J MURPHY
Signature and Printed Name

07/06/2018
Date

7/9/18

FDLE/ATP Form 41 –Revised August 2005

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
07/05/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:40
Control Test	0.049	12:40
Air Blank	0.000	12:41
Control Test	0.048	12:42
Air Blank	0.000	12:42
Control Test	0.048	12:43
Air Blank	0.000	12:43
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

P Murphy
Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
07/05/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:35
Control Test	0.079	12:36
Air Blank	0.000	12:36
Control Test	0.081	12:37
Air Blank	0.000	12:37
Control Test	0.080	12:38
Air Blank	0.000	12:39
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

P Murphy
Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
07/05/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:26
Control Test	0.197	12:27
Air Blank	0.000	12:27
Control Test	0.198	12:28
Air Blank	0.000	12:28
Control Test	0.199	12:29
Air Blank	0.000	12:30
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

P Murphy
Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
07/05/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:55
Control Test	0.081	12:55
Air Blank	0.000	12:56
Control Test	0.081	12:56
Air Blank	0.000	12:57
Control Test	0.082	12:57
Air Blank	0.000	12:57
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

DGS

P Murphy
Operator's Signature

7/9/18
JO

FDLE

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001114, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001114
Owning Agency:	DELAND P.D.
Calibration Date:	07/06/2018
Calibration Time:	11:45

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.005
0.200 g/ 210 L	0.008
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/06/2018
Date

Patrick J Murphy
PATRICK J MURPHY,
Department Inspector

7/6/18

- Clock keeps getting off

- Dry Gas control tolerance is too low. AVG .076

80-001114
JD

7/9/18
JD

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Deland PD S/N 80-001114

Date In 05/17/2018 DI Completion Date 5/23/2018 ☒Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By JG

- ☑ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord ☐ Printer Cable
- ☐ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date

FDLE

MAY 24 2018

Alcohol Testing Program

Quality Checks Performed By PSM

- ☒ Breath Tube Screen
- ☒ Replace External O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 205
- ☒ Flow Verification (L/s)

Flow Column # ATP103

32 mm	.152	(.139 - .169)
36 mm	.160	(.156 - .190)
53 mm	.242	(.228 - .278)
103 mm	.484	(.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 28427

- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD1021	201707D 07/25/2019
0.080	DR1275	201707E 07/25/2019
0.200	SD1019	201707C 07/24/2019
0.080 DGS	N/A	A6805702 2/26/20

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm		(.139 - .169)
36 mm		(.156 - .190)
53 mm		(.228 - .278)
103 mm		(.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By PSM

- ☒ Lab Temp °C 21.4
- External Digital Therm. ID#: 300502
- ☒ 34°C +-.2 Serial #: SD1021
- ☒ 34°C +-.2 Serial #: DR1275
- ☒ 34°C +-.2 Serial #: SD1019

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By PSM

Barometric Pressure ID# 28427

Gauge 1015 Instrument 1014

Mouth Alcohol Solution Lot # 2017-B

Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	G11621
Interferent	DR3855
0.050	SD1021
0.080	DR1275
0.200	SD1019

Attachments

- ☒ Form 41
- ☒ Stability Checks
- ☒ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☐ Other _____

Notes/Suggested Service: _____

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

SP 5/24/18 Tech Review / Date

JJ 5/24/18 Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DELAND P.D.
Time of Inspection: 11:35 Date of Inspection: 05/23/2018 Serial Number: 80-001114
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.048	0.079	0.198	0.078
0.000	0.048	0.080	0.200	0.077
0.000	0.048	0.080	0.200	0.076
0.000	0.049	0.080	0.200	0.076
0.000	0.049	0.080	0.200	0.077
0.000	0.048	0.080	0.200	0.076
0.000	0.048	0.080	0.200	0.076
0.000	0.048	0.080	0.200	0.076
0.000	0.049	0.080	0.199	0.076
0.000	0.048	0.080	0.200	0.076

Standard Deviations	0.0004	0.0003	0.0006	0.0006

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy
Signature and Printed Name

PATRICK J MURPHY

05/23/2018
Date

5/24/18
JD

FDLE/ATP Form 41 –Revised August 2005

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
05/23/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:45
Control Test	0.047	08:46
Air Blank	0.000	08:47
Control Test	0.048	08:47
Air Blank	0.000	08:48
Control Test	0.048	08:48
Air Blank	0.000	08:49

Control Test Stats

Average	0.0477
Std Dev	0.0006
Rel Std Dev(%)	1.2112

P Murphy
Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
05/23/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:00
Control Test	0.079	09:01
Air Blank	0.000	09:02
Control Test	0.079	09:02
Air Blank	0.000	09:03
Control Test	0.079	09:04
Air Blank	0.000	09:04

Control Test Stats

Average	0.0790
Std Dev	0.0000
Rel Std Dev(%)	0.0000

P Murphy
Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
05/23/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:55
Control Test	0.199	08:56
Air Blank	0.000	08:56
Control Test	0.198	08:57
Air Blank	0.000	08:58
Control Test	0.199	08:58
Air Blank	0.000	08:59

Control Test Stats

Average	0.1987
Std Dev	0.0006
Rel Std Dev(%)	0.2906

P Murphy
Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
05/23/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:09
Control Test	0.078	09:09
Air Blank	0.000	09:10
Control Test	0.079	09:10
Air Blank	0.000	09:11
Control Test	0.079	09:11
Air Blank	0.000	09:11

Control Test Stats

Average	0.0787
Std Dev	0.0006
Rel Std Dev(%)	0.7339

DGS

P Murphy
Operator's Signature

SP

5/24/18
JO

FDLE

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001114, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001114
Owning Agency:	DELAND P.D.
Calibration Date:	05/23/2018
Calibration Time:	11:35

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.005
0.200 g/ 210 L	0.008
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

SP

05/23/2018
Date

Patrick J Murphy
PATRICK J MURPHY,
Department Inspector

5/24/18 SP