

Return Material Authorization

Ship to: CMI, Inc.
 Enforcement Electronics

Shipment to repair facility authorized by: Debbie Reasoner on 05/10/2018

Items Returned: Instrument Supplies Other Describe: _____
Instrument Model: Intoxilyzer 8000 Serial Number: 80-001174

Bill To Address: <u>Clermont Police Department</u> <u>856 West Montrose Street</u> <u>Clermont, FL 34711</u>	Ship to Address: <u>FDLE Off Site Mail Facility</u> <u>c/o Florida Dept. of Law Enforcement</u> <u>Alcohol Testing Program / Danielle Bell</u> <u>813-B Lake Bradford Road</u> <u>Tallahassee, FL 32304</u>
--	---

Reason for Return:
Will not completely boot intermittently. Cycles power and prints start-up on internal printer then goes back from green to red. Also, speaker volume excessively low, suspect shifted speaker.
Please also check Dry Gas regulator for leaks - Agency indicated dry gas usage way above normal. DB

Please choose one of the following options:

1. I _____, authorize all repairs.

2. I _____, authorize repairs up to \$ _____.

3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Deborah Reasoner
Phone #: (352) 394-5588 ext 141 Email: DReasoner@clermontfl.org

ATP Contact Name: Danielle Bell ATP Email: DanielleBell@fdle.state.fl.us

5/10/18
DB

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.
Time of Inspection: 14:15

Date of Inspection: 05/10/2018

Serial Number: 80-001174
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.049	0.081	0.199	0.078
0.000	0.050	0.081	0.200	0.078
0.000	0.050	0.081	0.201	0.078
0.000	0.050	0.081	0.200	0.078
0.000	0.050	0.081	0.200	0.078
0.000	0.050	0.081	0.200	0.078
0.000	0.050	0.081	0.201	0.078
0.000	0.050	0.081	0.200	0.078
0.000	0.051	0.081	0.201	0.078
0.000	0.050	0.081	0.201	0.078

Standard Deviations	0.0004	0.0000	0.0006	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

(Handwritten mark)

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

(Handwritten Signature)

DANIELLE M BELL

Signature and Printed Name

05/10/2018
Date

(Handwritten: 5/10/18)

Stability Checks #80-001174 Clermont P.D. 5/10/18 ~~800~~

DGS

CLERMONT P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000
05/10/2018
Software: 8100.27
SN 80-001174

Test	g/210L	Time
Air Blank	0.000	11:50
Control Test	0.079	11:51
Air Blank	0.000	11:51
Control Test	0.079	11:52
Air Blank	0.000	11:52
Control Test	0.079	11:52
Air Blank	0.000	11:53
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

~~800~~
Operator's Signature

CLERMONT P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000
05/10/2018
Software: 8100.27
SN 80-001174

Test	g/210L	Time
Air Blank	0.000	11:45
Control Test	0.201	11:46
Air Blank	0.000	11:46
Control Test	0.200	11:47
Air Blank	0.000	11:48
Control Test	0.199	11:48
Air Blank	0.000	11:49
Control Test Stats		
Average	0.2000	
Std Dev	0.0010	
Rel Std Dev(%)	0.5000	

~~800~~
Operator's Signature

CLERMONT P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000
05/10/2018
Software: 8100.27
SN 80-001174

Test	g/210L	Time
Air Blank	0.000	11:40
Control Test	0.082	11:41
Air Blank	0.000	11:41
Control Test	0.082	11:42
Air Blank	0.000	11:43
Control Test	0.081	11:43
Air Blank	0.000	11:44
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

~~800~~
Operator's Signature

CLERMONT P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000
05/10/2018
Software: 8100.27
SN 80-001174

Test	g/210L	Time
Air Blank	0.000	11:35
Control Test	0.050	11:36
Air Blank	0.000	11:36
Control Test	0.050	11:37
Air Blank	0.000	11:37
Control Test	0.050	11:38
Air Blank	0.000	11:39
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

~~800~~
Operator's Signature

8/10/18
[Signature]



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001174, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001174</u>	UNCERTAINTY* ±	
Owning Agency:	<u>CLERMONT P.D.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/10/2018</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>14:15</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/10/2018 _____
 Date *Danielle M Bell*
 DANIELLE M BELL,
 Department Inspector

FDLE/ATP Form 69 March 2018
 Issuing Authority: Alcohol Testing Program
 Service • Integrity • Respect • Quality

5/10/18
[Signature]