



# INSTRUMENT PROCESSING SHEET

Agency Rockledge PDS/N 80-001261

Florida Department of Law Enforcement

Date In 2/2/2018DI Completion Date 3/5/18 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>JP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>200</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.246</u> (.228 - .278) 103 mm <u>.527</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28662</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>G2835</u></td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td><u>SD3933</u></td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG626604</u> <u>9-22-18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>G2835</u>	201707D 07/25/2019	0.080	<u>SD3933</u>	201707E 07/25/2019	0.200	<u>SD1025</u>	201707C 07/24/2019	0.080 DGS	N/A	<u>AG626604</u> <u>9-22-18</u>	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <b>Maintenance</b> Performed By <u>SP</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input checked="" type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.3</u> External Digital Therm. ID#: <u>300502</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2835</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3933</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1025</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	<u>G2835</u>	201707D 07/25/2019																																																												
0.080	<u>SD3933</u>	201707E 07/25/2019																																																												
0.200	<u>SD1025</u>	201707C 07/24/2019																																																												
0.080 DGS	N/A	<u>AG626604</u> <u>9-22-18</u>																																																												
<b>Final Release Date</b> <b>FDLE</b> <b>MAR 08 2018</b> <b>Alcohol Testing Program</b>																																																														
<b>Calibration Adjustment</b> Performed By <u>SP</u> Barometric Pressure Gauge <u>1017</u> ID # <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>DR3855</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>SD1024</u></td> <td><u>16320</u></td> <td><u>10-21-18</u></td> </tr> <tr> <td>0.100</td> <td><u>G2834</u></td> <td><u>17280</u></td> <td><u>9-11-19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1011</u></td> <td><u>17090</u></td> <td><u>2-24-19</u></td> </tr> <tr> <td>0.300</td> <td><u>DR1275</u></td> <td><u>17140</u></td> <td><u>5-15-19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>178170SDA2</u></td> <td><u>8-5-19</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>G2835</u></td> <td><u>201707D</u></td> <td><u>7-25-19</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3933</u></td> <td><u>201707E</u></td> <td><u>7-25-19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>201707C</u></td> <td><u>7-24-19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG626604</u></td> <td><u>9-22-18</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>DR3855</u>	N/A	N/A	0.040	<u>SD1024</u>	<u>16320</u>	<u>10-21-18</u>	0.100	<u>G2834</u>	<u>17280</u>	<u>9-11-19</u>	0.200	<u>SD1011</u>	<u>17090</u>	<u>2-24-19</u>	0.300	<u>DR1275</u>	<u>17140</u>	<u>5-15-19</u>	0.080 DGS	N/A	<u>178170SDA2</u>	<u>8-5-19</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>G2835</u>	<u>201707D</u>	<u>7-25-19</u>	0.080	<u>SD3933</u>	<u>201707E</u>	<u>7-25-19</u>	0.200	<u>SD1025</u>	<u>201707C</u>	<u>7-24-19</u>	0.080 DGS	N/A	<u>AG626604</u>	<u>9-22-18</u>	<b>Department Inspection</b> Performed By <u>SP</u> Barometric Pressure ID# <u>28662</u> Gauge <u>1016</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2830</u></td> </tr> <tr> <td>Interferent</td> <td><u>G8144</u></td> </tr> <tr> <td>0.050</td> <td><u>G2835</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3933</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between; margin-top: 10px;"> <div style="text-align: center;"> <u>3/8/18</u>              Tech Review / Date           </div> <div style="text-align: center;"> <u>[Signature]</u> <u>3/8/18</u>              Admin Review / Date           </div> </div>		Simulator	Serial Number	0.000	<u>G2830</u>	Interferent	<u>G8144</u>	0.050	<u>G2835</u>	0.080	<u>SD3933</u>	0.200	<u>SD1025</u>
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000	<u>DR3855</u>	N/A	N/A																																																											
0.040	<u>SD1024</u>	<u>16320</u>	<u>10-21-18</u>																																																											
0.100	<u>G2834</u>	<u>17280</u>	<u>9-11-19</u>																																																											
0.200	<u>SD1011</u>	<u>17090</u>	<u>2-24-19</u>																																																											
0.300	<u>DR1275</u>	<u>17140</u>	<u>5-15-19</u>																																																											
0.080 DGS	N/A	<u>178170SDA2</u>	<u>8-5-19</u>																																																											
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050	<u>G2835</u>	<u>201707D</u>	<u>7-25-19</u>																																																											
0.080	<u>SD3933</u>	<u>201707E</u>	<u>7-25-19</u>																																																											
0.200	<u>SD1025</u>	<u>201707C</u>	<u>7-24-19</u>																																																											
0.080 DGS	N/A	<u>AG626604</u>	<u>9-22-18</u>																																																											
Simulator	Serial Number																																																													
0.000	<u>G2830</u>																																																													
Interferent	<u>G8144</u>																																																													
0.050	<u>G2835</u>																																																													
0.080	<u>SD3933</u>																																																													
0.200	<u>SD1025</u>																																																													



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.

Serial Number: 80-001261

Time of Inspection: 15:39

Date of Inspection: 03/08/2018

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.050	0.082	0.199	0.079
0.000	0.052	0.082	0.201	0.079
0.000	0.050	0.082	0.201	0.079
0.000	0.050	0.082	0.201	0.079
0.000	0.051	0.082	0.202	0.079
0.000	0.050	0.082	0.201	0.079
0.000	0.050	0.081	0.201	0.079
0.000	0.051	0.082	0.201	0.079
0.000	0.050	0.082	0.201	0.080
0.000	0.050	0.082	0.201	0.079

Standard Deviations	0.0006	0.0003	0.0007	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

*(Handwritten mark)*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*(Handwritten Signature: Shayla Platt)*

SHAYLA D PLATT

Signature and Printed Name

03/08/2018  
Date

*(Handwritten: 3/8/18)*

# STABILITY CHECKS - #80-001261

ROCKLEDGE P.D.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001261  
03/08/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:30
Control Test	0.049	12:31
Air Blank	0.000	12:31
Control Test	0.050	12:32
Air Blank	0.000	12:32
Control Test	0.049	12:33
Air Blank	0.000	12:33
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Operator's Signature *SP* 3/8/18

ROCKLEDGE P.D.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001261  
03/08/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:35
Control Test	0.079	12:36
Air Blank	0.000	12:36
Control Test	0.079	12:37
Air Blank	0.000	12:37
Control Test	0.079	12:38
Air Blank	0.000	12:38
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature *SP*

ROCKLEDGE P.D.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001261  
03/08/2018  
Software: 8100.27

Test	g/210L	n	Time
Air Blank	0.000		12:46
Control Test	0.194		12:47
Air Blank	0.000		12:47
Control Test	0.193		12:48
Air Blank	0.000		12:49
Control Test	0.193		12:49
Air Blank	0.000		12:50
Control Test Stats			
Average	0.1933		
Std Dev	0.0006		
Rel Std Dev(%)	0.2986		

Operator's Signature *SP*

ROCKLEDGE P.D.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001261  
03/08/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:51
Control Test	0.079	12:51
Air Blank	0.000	12:51
Control Test	0.079	12:52
Air Blank	0.000	12:52
Control Test	0.080	12:52
Air Blank	0.000	12:53
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

Operator's Signature *SP*



# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001261, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001261</u>	UNCERTAINTY* ±	
Owning Agency:	<u>ROCKLEDGE P.D.</u>	0.05 g/ 210 L	0.004
Calibration Date:	<u>03/08/2018</u>	0.08 g/ 210 L	0.005
Calibration Time:	<u>15:39</u>	0.20 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Target Alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/08/2018

Date

*Shayla D Platt*

**SHAYLA D PLATT,**  
Department Inspector

FDLE/ATP Form 69 January 2018  
Issuing Authority: Alcohol Testing Program

*Service • Integrity • Respect • Quality*

*3/8/18*  
*[Signature]*



ROCKLEDGE P.O.  
Intoxilyzer - Alcohol Analyzer  
Model: 8000  
03/08/2018 13:11:38

Auto Calibration  
Max Power Res Value = 44  
Auto Range Res Value = 25

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12655, Sum Io = 13492

Sample % Abs (% Abs Ref)  
Sample #1 = 0.0690 (-0.0060)  
Sample #2 = 0.0640 (0.0590)  
Sample #3 = 0.0590 (0.0670)  
Sample #4 = 0.0620 (0.0850)  
Aug % Abs = 0.0617 (0.0703)  
STD DEV = 0.0025 (0.0133)  
REL STD DEV = 4.081 (18.934)

Sample % Abs (% Abs Ref)  
Sample #1 = 1.4640 (-0.0240)  
Sample #2 = 1.4590 (-0.0140)  
Sample #3 = 1.4480 (-0.0080)  
Sample #4 = 1.4490 (-0.0010)  
Aug % Abs = 1.4520 (-0.0077)  
STD DEV = 0.0061 (0.0065)  
REL STD DEV = 0.419 (84.866)

Sample % Abs (% Abs Ref)  
Sample #1 = 6.7120 (0.0200)  
Sample #2 = 6.7540 (0.0210)  
Sample #3 = 6.7750 (0.0500)  
Sample #4 = 6.7730 (0.0150)  
Aug % Abs = 6.7673 (0.0137)  
STD DEV = 0.0116 (0.0081)  
REL STD DEV = 0.171 (59.143)

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
<<<<< CHANNEL 1 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.062  
Std Dev = 0.00 Rel Std Dev = 4.08  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.747  
Std Dev = 0.02 Rel Std Dev = 2.41  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.845  
Std Dev = 0.03 Rel Std Dev = 1.47  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.568  
Std Dev = 0.01 Rel Std Dev = 0.17  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.225  
Std Dev = 0.04 Rel Std Dev = 0.74  
Zero Order Coef = -121.36  
First Order Coef = 2602.63  
Second Order Coef = 29.31  
Standard Deviation = 39.874956

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 3461.00  
Sample #2 = 3546.00  
Sample #3 = 3461.00  
Sample #4 = 3489.00  
Average Result = 3498.6667  
STD DEV = 43.3167  
REL STD DEV = 1.238  
\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sample #1 = 3638.00  
Sample #2 = 3664.00  
Sample #3 = 3623.00  
Sample #4 = 3617.00  
Average Result = 3634.6667  
STD DEV = 25.5799  
REL STD DEV = 0.704  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1016  
3 um H2O Adjust (mg/\*10,000) = 311  
9 um H2O Adjust (mg/\*10,000) = 175  
\*\*\*\* AUTO CAL PASS

Sol Value = 0.300 g/210L \*\*\*  
Fit value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12619, Sum Io = 13480

Sample % Abs (% Abs Ref)  
Sample #1 = 5.2010 (-0.0030)  
Sample #2 = 5.1880 (0.0360)  
Sample #3 = 5.2650 (0.0140)  
Sample #4 = 5.2220 (0.0550)  
Aug % Abs = 5.2250 (0.0350)  
STD DEV = 0.0386 (0.0205)  
REL STD DEV = 0.739 (58.624)

Sample % Abs (% Abs Ref)  
Sample #1 = 9.7600 (-0.0130)  
Sample #2 = 9.7320 (0.0200)  
Sample #3 = 9.8810 (0.0070)  
Sample #4 = 9.8340 (0.0220)  
Aug % Abs = 9.8223 (0.0163)  
STD DEV = 0.0653 (0.0081)  
REL STD DEV = 0.665 (49.864)

Sample % Abs (% Abs Ref)  
Sample #1 = 3.4750 (-0.0030)  
Sample #2 = 3.5070 (0.0020)  
Sample #3 = 3.5420 (0.0170)  
Sample #4 = 3.5330 (0.0110)  
Aug % Abs = 3.5273 (0.0100)  
STD DEV = 0.0182 (0.0075)  
REL STD DEV = 0.515 (75.498)

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 0.2000 mg/l or 0.000 g/210L  
% Abs = 0.098  
Std Dev = 0.02 Rel Std Dev = 19.46  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.452  
Std Dev = 0.01 Rel Std Dev = 0.42  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.527  
Std Dev = 0.02 Rel Std Dev = 0.52  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.767  
Std Dev = 0.01 Rel Std Dev = 0.17  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.822  
Std Dev = 0.07 Rel Std Dev = 0.66  
Zero Order Coef = -105.91  
First Order Coef = 1337.03  
Second Order Coef = 12.98  
Standard Deviation = 26.095860

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 0.2000 mg/l or 0.000 g/210L  
% Abs = 0.098  
Std Dev = 0.02 Rel Std Dev = 19.46  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.452  
Std Dev = 0.01 Rel Std Dev = 0.42  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.527  
Std Dev = 0.02 Rel Std Dev = 0.52  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.767  
Std Dev = 0.01 Rel Std Dev = 0.17  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.822  
Std Dev = 0.07 Rel Std Dev = 0.66  
Zero Order Coef = -105.91  
First Order Coef = 1337.03  
Second Order Coef = 12.98  
Standard Deviation = 26.095860

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12624, Sum Io = 13483

Sample % Abs (% Abs Ref)  
Sample #1 = 3.5440 (-0.0120)  
Sample #2 = 3.5690 (0.0110)  
Sample #3 = 3.5620 (0.0420)  
Sample #4 = 3.5740 (0.0270)  
Aug % Abs = 3.5683 (0.0267)  
STD DEV = 0.0060 (0.0155)  
REL STD DEV = 0.169 (58.135)

Sample % Abs (% Abs Ref)  
Sample #1 = 0.7910 (-0.0080)  
Sample #2 = 0.7680 (0.0120)  
Sample #3 = 0.7390 (0.0600)  
Sample #4 = 0.7350 (0.0800)  
Aug % Abs = 0.7473 (0.0507)  
STD DEV = 0.0180 (0.0349)  
REL STD DEV = 2.410 (68.975)

Sample % Abs (% Abs Ref)  
Sample #1 = 0.0400 (-0.0080)  
Sample #2 = 0.1905 (0.0120)  
Sample #3 = 0.7390 (0.0600)  
Sample #4 = 0.7350 (0.0800)  
Aug % Abs = 0.7473 (0.0507)  
STD DEV = 0.0180 (0.0349)  
REL STD DEV = 2.410 (68.975)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 0.0400 mg/l or 0.000 g/210L  
% Abs = 0.000  
Std Dev = 0.00 Rel Std Dev = 0.0000  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.040  
Std Dev = 0.00 Rel Std Dev = 0.0000  
Sol Val = 0.7390 mg/l or 0.160 g/210L  
% Abs = 0.100  
Std Dev = 0.00 Rel Std Dev = 0.0000  
Sol Val = 0.7350 mg/l or 0.160 g/210L  
% Abs = 0.200  
Std Dev = 0.00 Rel Std Dev = 0.0000  
Sol Val = 0.3000 mg/l or 0.060 g/210L  
% Abs = 0.300  
Std Dev = 0.00 Rel Std Dev = 0.0000

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 0.0400 mg/l or 0.000 g/210L  
% Abs = 0.000  
Std Dev = 0.00 Rel Std Dev = 0.0000  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.040  
Std Dev = 0.00 Rel Std Dev = 0.0000  
Sol Val = 0.7390 mg/l or 0.160 g/210L  
% Abs = 0.100  
Std Dev = 0.00 Rel Std Dev = 0.0000  
Sol Val = 0.7350 mg/l or 0.160 g/210L  
% Abs = 0.200  
Std Dev = 0.00 Rel Std Dev = 0.0000  
Sol Val = 0.3000 mg/l or 0.060 g/210L  
% Abs = 0.300  
Std Dev = 0.00 Rel Std Dev = 0.0000

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12638, Sum Io = 13484

Sample % Abs (% Abs Ref)  
Sample #1 = 0.7910 (-0.0080)  
Sample #2 = 0.7680 (0.0120)  
Sample #3 = 0.7390 (0.0600)  
Sample #4 = 0.7350 (0.0800)  
Aug % Abs = 0.7473 (0.0507)  
STD DEV = 0.0180 (0.0349)  
REL STD DEV = 2.410 (68.975)

Sample % Abs (% Abs Ref)  
Sample #1 = 0.7910 (-0.0080)  
Sample #2 = 0.7680 (0.0120)  
Sample #3 = 0.7390 (0.0600)  
Sample #4 = 0.7350 (0.0800)  
Aug % Abs = 0.7473 (0.0507)  
STD DEV = 0.0180 (0.0349)  
REL STD DEV = 2.410 (68.975)

Sample % Abs (% Abs Ref)  
Sample #1 = 0.7910 (-0.0080)  
Sample #2 = 0.7680 (0.0120)  
Sample #3 = 0.7390 (0.0600)  
Sample #4 = 0.7350 (0.0800)  
Aug % Abs = 0.7473 (0.0507)  
STD DEV = 0.0180 (0.0349)  
REL STD DEV = 2.410 (68.975)

CAL ADJUSTMENT  
#80-0018261 SP

3/8/18  
[Signature]

POST CAL ADJUST STABILITY CHECKS #80-001261

ROCKLEDGE P.D.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001261  
03/08/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:13
Control Test	0.050	14:13
Air Blank	0.000	14:14
Control Test	0.050	14:15
Air Blank	0.000	14:15
Control Test	0.051	14:16
Air Blank	0.000	14:16
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

*SP*  
Operator's Signature

3/8/18  
JD

ROCKLEDGE P.D.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001261  
03/08/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:05
Control Test	0.081	14:06
Air Blank	0.000	14:06
Control Test	0.081	14:07
Air Blank	0.000	14:08
Control Test	0.082	14:08
Air Blank	0.000	14:09
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

*SP*  
Operator's Signature

ROCKLEDGE P.D.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001261  
03/08/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:51
Control Test	0.201	13:52
Air Blank	0.000	13:53
Control Test	0.203	13:53
Air Blank	0.000	13:54
Control Test	0.202	13:54
Air Blank	0.000	13:55
Control Test Stats		
Average	0.2020	
Std Dev	0.0010	
Rel Std Dev(%)	0.4950	

*SP*  
Operator's Signature

ROCKLEDGE P.D.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001261  
03/08/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:10
Control Test	0.080	14:10
Air Blank	0.000	14:10
Control Test	0.081	14:11
Air Blank	0.000	14:11
Control Test	0.081	14:12
Air Blank	0.000	14:12
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

*SP*  
Operator's Signature

DGS



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.

Time of Inspection: 12:28

Date of Inspection: 03/08/2018

Serial Number: 80-001261

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT

10

3/8/18  
JL

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

03/08/2018  
Date