



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol (Brevard) S/N 80-001361Florida Department of Law Enforcement Date In 1/4/2018 DI Completion Date 1/10/18 Ship P/U H/D CMI EE

Intake Performed By <u>JE</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____	Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>1.07</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>286602</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
--	---	--

Final Release Date

FDLE

JAN 12 2018

Alcohol Testing Program

Simulator	Serial #	Lot #/Exp
0.050	<u>SD3962</u>	201707D 7/25/2019
0.080	<u>SD3964</u>	201707E 7/25/2019
0.200	<u>DR3850</u>	201707C 7/24/2019
0.080 DGS	N/A	<u>AC10246004</u> <u>9-22-18</u>

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By SP

Lab Temp °C 22.5
 External Digital Therm. ID#: 300505
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD3964
 34°C +/- .2 Serial #: DR3850

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure ID# 286602
 Gauge 1019 Instrument 1010
 Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.050	<u>SD3962</u>
0.080	<u>SD3964</u>
0.200	<u>DR3850</u>

Attachments

<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration
<input checked="" type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40
<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____

Notes/Suggested Service: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

1/10/18 J. Duhon 1/12/18
 Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:03

Date of Inspection: 01/10/2018

Serial Number: 80-001361
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.049	0.081	0.197	0.080
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.197	0.080
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.198	0.080
0.000	0.049	0.082	0.198	0.080

Standard Deviations	0.0004	0.0003	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

10

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

01/10/2018
Date

*1/10/18
JD*

STABILITY CHECKS - #80-001361

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001361
01/10/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:54
Control Test	0.049	08:55
Air Blank	0.000	08:55
Control Test	0.049	08:56
Air Blank	0.000	08:57
Control Test	0.049	08:57
Air Blank	0.000	08:58
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

SP
Operator's Signature

1/12/18
SP

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001361
01/10/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:42
Control Test	0.079	08:43
Air Blank	0.000	08:44
Control Test	0.079	08:44
Air Blank	0.000	08:45
Control Test	0.081	08:45
Air Blank	0.000	08:46
Control Test Stats		
Average	0.0797	
Std Dev	0.0012	
Rel. Std Dev(%)	1.4494	

SP
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001361
01/10/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:11
Control Test	0.197	09:12
Air Blank	0.000	09:12
Control Test	0.198	09:13
Air Blank	0.000	09:13
Control Test	0.197	09:14
Air Blank	0.000	09:15
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel. Std Dev(%)	0.2926	

SP
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001361
01/10/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:00
Control Test	0.080	09:00
Air Blank	0.000	09:01
Control Test	0.080	09:01
Air Blank	0.000	09:02
Control Test	0.080	09:02
Air Blank	0.000	09:02
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

DGS

SP
Operator's Signature

