FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Hillsborough County Sheriff's Office S/N 80-003387

Date In 03/26/2018 DI Completion Date 04/03/2018 ☑Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By TG

☑ Annual
☐ Registration
☑ Return from CMI / EE

Visual Inspection:
☑ Case ☑ Handle
☑ Keyboard ☑ Dry Gas Shelf
☑ Feet ☑ Breath Tube
☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:
☐ Power cord ☐ Printer Cable
☑ Static Bag ☐ 12V DC Cable

Notes:

Final Release Date

FDLE
APR 0 3 2018
Alcohol Testing Program

Quality Checks Performed By DMB

☑ Breath Tube Screen
☑ Replace External O-Rings
☑ Instrument Set Up Verified
☑ R-Value 232
☑ Flow Verification (L/s)
Flow Column # ATP 103
32 mm 0.152 (.139 - .169)
36 mm 0.167 (.156 - .190)
53 mm 0.238 (.228 - .278)
103 mm 0.496 (.447 - .547)
☑ Barometric Pressure Check
Gauge ID # 28662
☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	G11739	201707D 07/25/2019
0.080	SD3964	201707E 07/25/2019
0.200	DR3856	201707C 07/24/2019
0.080 DGS	N/A	AG805702 02/26/2020

Flow Calibration Performed By

Flow Column #
☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm
☐ R-Value
☐ Post Calibration Verification (L/s)
Flow Column #
32 mm (.139 - .169)
36 mm (.156 - .190)
53 mm (.228 - .278)
103 mm (.447 - .547)

Maintenance Performed By

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☐ Other

Temperature Checks Performed By DMB

☑ Lab Temp °C 22.3
External Digital Therm. ID#: 300503
☑ 34°C +-.2 Serial #: G11739
☑ 34°C +-.2 Serial #: SD3964
☑ 34°C +-.2 Serial #: DR3856

Calibration Adjustment Performed By

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By DMB

Barometric Pressure ID# 28662
Gauge 1016 Instrument 1015
Mouth Alcohol Solution Lot # 2017-B
Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	SD1019
Interferent	SD1021
0.050	G11739
0.080	SD3964
0.200	DR3856

Attachments

☑ Form 41
☑ Stability Checks
☑ Calibration Certificate
☐ Calibration Adjustment
☐ Post-Stability Checks
☐ Flow Calibration
☐ Form 40
☐ Other

Notes/Suggested Service: Please update level 2 menu password to something unique. 4/3/18 DMB

☑ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
☑ **Return to/Place into Evidentiary Use**
☐ **Remain Out of Evidentiary Use**
☑ **Conduct an Agency Inspection Before Evidentiary Use**

4/3/18 — Tech Review / Date
4/3/18 — Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 13:20
Date of Inspection: 04/03/2018
Serial Number: 80-003387
Software: 8100.27

1/3CC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.048	0.079	0.196	0.077
0.000	0.049	0.080	0.198	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.080	0.198	0.078
0.000	0.048	0.079	0.195	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.080	0.196	0.078
0.000	0.049	0.081	0.198	0.078
0.000	0.049	0.080	0.198	0.078
Standard Deviations	0.0004	0.0006	0.0013	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL
Signature and Printed Name

04/03/2018
Date

4/3/18

FDLE/ATP Form 41 –Revised August 2005

Stability Checks #80-003387 Hillsborough County S.O. 4/3/18 DDB

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003387
04/03/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:56
Control Test	0.049	10:57
Air Blank	0.000	10:57
Control Test	0.048	10:58
Air Blank	0.000	10:58
Control Test	0.049	10:59
Air Blank	0.000	11:00
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

DDB
Operator's Signature

4/3/18

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003387
04/03/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:02
Control Test	0.080	11:03
Air Blank	0.000	11:03
Control Test	0.079	11:04
Air Blank	0.000	11:04
Control Test	0.080	11:05
Air Blank	0.000	11:06
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DDB
Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003387
04/03/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:07
Control Test	0.194	11:08
Air Blank	0.000	11:08
Control Test	0.198	11:09
Air Blank	0.000	11:09
Control Test	0.197	11:10
Air Blank	0.000	11:11
Control Test Stats		
Average	0.1963	
Std Dev	0.0021	
Rel Std Dev(%)	1.0603	

DDB
Operator's Signature

DGS

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003387
04/03/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:12
Control Test	0.078	11:12
Air Blank	0.000	11:13
Control Test	0.078	11:13
Air Blank	0.000	11:14
Control Test	0.078	11:14
Air Blank	0.000	11:15
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DDB
Operator's Signature

FDLE

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003387, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-003387		UNCERTAINTY* ±
Owning Agency:	HILLSBOROUGH CO SO	0.050 g/ 210 L	0.004
Calibration Date:	04/03/2018	0.080 g/ 210 L	0.005
Calibration Time:	13:20	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/03/2018
Date

DANIELLE M BELL,
Department Inspector

4/3/18

Return Material Authorization

Ship to: ☐ CMI, Inc.

☑ Enforcement Electronics

Shipment to repair facility authorized by: Roger Skipper on 03/05/2018

Items Returned: Instrument ☑ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-003387

Bill To Address:	Ship to Address:
Hillsborough County Sheriff's Office	Attn: Danielle Bell
Central Breath Testing Unit	Florida Department of Law Enforcement
1201 Orient Rd	Alcohol Testing Program
Tampa, FL 33619	2331 Phillips Rd.
	Tallahassee, FL 32308

Reason for Return:

R-Value reading 101. Please replace.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to $_____.

☑ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Roger Skipper

Phone #: (813) 247-8361 Email: rskipper@hcso.tampa.fl.us

ATP Contact Name: Danielle Bell ATP Email: DanielleBell@fdle.state.fl.us

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Hillsborough County SO S/N 80-003387

Date In 02/16/2018 DI Completion Date 03/05/2018 ☐Ship ☐P/U ☐H/D ☐CMI ☑EE

Intake Performed By TG

☑ Annual
☐ Registration
☐ Return from CMI / EE

Visual Inspection:
☑ Case ☑ Handle
☑ Keyboard ☑ Dry Gas Shelf
☑ Feet ☑ Breath Tube
☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:
☐ Power cord ☐ Printer Cable
☑ Static Bag ☐ 12V DC Cable

Notes: See correspondence shipped with instrument.

Final Release Date

FDLE
MAR 06 2018
Alcohol Testing Program

Quality Checks Performed By DMB

☑ Breath Tube Screen
☑ Replace External O-Rings
☑ Instrument Set Up Verified
☑ R-Value 101
☑ Flow Verification (L/s)
Flow Column # ATP 102
32 mm 0.113 (.139 - .169)
36 mm 0.132 (.156 - .190)
53 mm 0.203 (.228 - .278)
103 mm 0.480 (.447 - .547)
☑ Barometric Pressure Check
Gauge ID # 28427
☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	G2835	201707D 07/25/2019
0.080 DMB	SD1013 ~~SD3964~~	201707E 07/25/2019
0.200	SD1025	201707C 07/24/2019
0.080 DGS	N/A	AG715202 06/01/2019

Flow Calibration Performed By DMB

Flow Column # ATP 103
☑ 5L/min – 17mm
☑ 15L/min – 53mm
☑ 30L/min – 103mm
☑ R-Value 101
☑ Post Calibration Verification (L/s)
Flow Column # ATP 102
32 mm 0.148 (.139 - .169)
36 mm 0.171 (.156 - .190)
53 mm 0.238 (.228 - .278)
103 mm 0.511 (.447 - .547)

Maintenance Performed By _____

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☐ Other _____

Temperature Checks Performed By DMB

☑ Lab Temp °C 21.7
External Digital Therm. ID#: 300502
☑ 34°C +-.2 Serial #: G2835
☑ 34°C +-.2 Serial #: SD3964
☑ 34°C +-.2 Serial #: SD1025

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By DMB

Barometric Pressure ID# 28427
Gauge 1016 Instrument 1012
Mouth Alcohol Solution Lot # 2016-C
Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	G2880
Interferent	G8144
0.050	G2835
0.080	SD3964
0.200	SD1025

Attachments

☑ Form 41
☑ Stability Checks
☑ Calibration Certificate
☐ Calibration Adjustment
☐ Post-Stability Checks
☑ Flow Calibration
☐ Form 40
☐ Other _____

Notes/Suggested Service: Sending to Enforcement Electronics for Flow Sensor replacement due to R-value reading low. 3/5/18 DMB

☑ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
☐ **Return to/Place into Evidentiary Use**
☑ **Remain Out of Evidentiary Use**
☐ **Conduct an Agency Inspection Before Evidentiary Use**

3/6/18 — Tech Review / Date
3/6/18 — Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Serial Number: 80-003387
Time of Inspection: 15:46
Date of Inspection: 03/05/2018
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG715202 Exp: 06/01/2019
0.000	0.049	0.081	0.197	0.080
0.000	0.050	0.081	0.198	0.079
0.000	0.049	0.081	0.200	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.081	0.199	0.080
0.000	0.049	0.082	0.197	0.080
0.000	0.050	0.080	0.200	0.080
Standard Deviations	0.0005	0.0006	0.0011	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIELLE M BELL

Signature and Printed Name

03/05/2018

Date

3/6/18

Stability Checks #80-003387 Hillsborough County S.O. 3/5/18

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003387
03/05/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:05
Control Test	0.050	12:06
Air Blank	0.000	12:07
Control Test	0.049	12:07
Air Blank	0.000	12:08
Control Test	0.051	12:09
Air Blank	0.000	12:09
Control Test Stats		
Average	0.0500	
Std Dev	0.0010	
Rel Std Dev(%)	2.0000	

Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003387
03/05/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:10
Control Test	0.082	12:11
Air Blank	0.000	12:12
Control Test	0.081	12:12
Air Blank	0.000	12:13
Control Test	0.082	12:13
Air Blank	0.000	12:14
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

Sim SD1013 determined to be running hot. Replaced simulator.

Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003387
03/05/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:19
Control Test	0.081	13:20
Air Blank	0.000	13:20
Control Test	0.080	13:21
Air Blank	0.000	13:21
Control Test	0.080	13:22
Air Blank	0.000	13:23
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

SD3964

Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003387
03/05/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:16
Control Test	0.196	12:16
Air Blank	0.000	12:17
Control Test	0.201	12:18
Air Blank	0.000	12:18
Control Test	0.200	12:19
Air Blank	0.000	12:19
Control Test Stats		
Average	0.1990	
Std Dev	0.0026	
Rel Std Dev(%)	1.3295	

Operator's Signature

D35

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003387
03/05/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:21
Control Test	0.078	12:21
Air Blank	0.000	12:22
Control Test	0.079	12:22
Air Blank	0.000	12:22
Control Test	0.080	12:23
Air Blank	0.000	12:23
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

Operator's Signature

3/5/18

FDLE

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003387, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-003387		UNCERTAINTY* ±
Owning Agency:	HILLSBOROUGH CO SO	0.05 g/ 210 L	0.004
Calibration Date:	03/05/2018	0.08 g/ 210 L	0.005
Calibration Time:	15:46	0.20 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Tartget Alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/05/2018
Date

DANIELLE M BELL,
Department Inspector

3/6/18

Flow Calibration Adjustment Data
#80-003387
Hillsborough County S.O.
3/5/18
RMB

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003387
03/05/2018
Software: 8100.27

Flow Rate Calibration********
1: Rate (Liters/min) = 5
SQRT(Diff)) = 5.000
2: Rate (Liters/min) = 15
SQRT(Diff)) = 9.898
3: Rate (Liters/min) = 30
SQRT(Diff)) = 19.645
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 658
Rounded Intercept = -271552
Correlation = 0.99727

3/6/18

RECEIVED
FEB 16 2018
FDLE
Alcohol Testing Program

CHAD CHRONISTER, SHERIFF
Hillsborough County Sheriff's Office
Tampa, FL

INTER-OFFICE
MEMORANDUM

DATE: January 23, 2018

TO: Master Sergeant Timothy R. Bell #4152
JAC/CBTU Sergeant

FROM: Supervisor Roger G. Skipper #246109
Central Breath Testing Unit Supervisor

RE: Unescorted Instrument Access

MESSAGE:

At 1051 hours on January 23, 2018, Alcohol Breath Analysis Technician Melanie Snyder #4113 informed me that Locksmith Samuel Ramirez, Jr. #242169 had accessed Room SI-01 (Breath Testing Room) as a pass-through to Room FR-01 (CBTU File Room) to finish a lock repair. Mr. Ramirez Jr. had been previously escorted by Technician Snyder and was told to call when he returned to finish the work order. It was clearly stated on the work order and a notice on the door that the area had restricted access and that Central Breath Testing personnel must be contacted.

Upon return from court, I advised Master Sergeant Bell #4152. I also reported the access violation to State Inspector Danielle Bell of the Florida Department of Law Enforcement's Alcohol Testing Program (FDLE/ATP). After consultation with Inspector Bell, I conducted a System Check breath test on the four instruments located in the testing room with no issues noted. I also conducted a dry gas standard stability check (10 repetitions each) on all four instruments involved with no issues noted.

The four instruments (80-000833, 80-003387, 80-003388 and 80-006567) were removed from use and will be shipped to FDLE/ATP after an agency inspection.

APPROVED	DISAPPROVED	ACK	DATE	NAME	ABN	TITLE
[✓]	[ ]	[ ]	1-23-18	M/Sgt. T.R. Bell	4152	M/Sergeant
[ ]	[ ]	[ ]				
[ ]	[ ]	[ ]				
[ ]	[ ]	[ ]				
[ ]	[ ]	[ ]				
[ ]	[ ]	[ ]				
[ ]	[ ]	[ ]				
[ ]	[ ]	[ ]				

RECEIVED
FEB 16 2018
FDLE
Alcohol Testing Program

HCSO Facilities
Support Ticket Center

01/22/18 20:04:18

Ticket #026263

Status	Open	**Name**	TIMOTHY BELL
Priority	Normal	**Email**	tbell@hcso.tampa.fl.us
Department	DMB ORJ	**Phone**	(813) 936-9099
Create Date	01/22/18 15:04:12	**Source**	Web

Ticket Details

Department: ORJ

Location: ORJ - Building 10 - CBT

CBT Staff Lockers Keys/Locks

01/22/18 15:04:12 TIMOTHY BELL

Keys to some of the staff lockers in CBT are not working. Example: locker #8. Also, need a spare key for locker #8 and #9. See Mr. Skipper as he can get you access to the restricted access area and has the set of locker keys.