



INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-003406

Florida Department of Law Enforcement

Date In 10/29/18DI Completion Date 11/14/2018 Ship P/U H/D CMI EE

Intake Performed By <u>SQC</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>missing 2 feet</u>	Quality Checks Performed By <u>WJM</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>160</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.140</u> (.139 - .169) 36 mm <u>.156</u> (.156 - .190) 53 mm <u>.230</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1021</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1275</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1013</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG805701</u> <u>2/26/20</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1021</u>	<u>201707D</u> <u>7/25/19</u>	0.080	<u>DR1275</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>SD1013</u>	<u>201707C</u> <u>7/24/19</u>	0.080 DGS	N/A	<u>AG805701</u> <u>2/26/20</u>	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>SD1021</u>	<u>201707D</u> <u>7/25/19</u>																																																											
0.080	<u>DR1275</u>	<u>201707E</u> <u>7/25/19</u>																																																											
0.200	<u>SD1013</u>	<u>201707C</u> <u>7/24/19</u>																																																											
0.080 DGS	N/A	<u>AG805701</u> <u>2/26/20</u>																																																											
Final Release Date FDLE NOV 21 2018 Alcohol Testing Program	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	Temperature Checks Performed By <u>WJM</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.4</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1021</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>DR1275</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1013</u>																																																											
Calibration Adjustment Performed By <u>WJM</u> Barometric Pressure Gauge <u>1015</u> ID # <u>28662</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2834</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>SD1022</u></td> <td><u>17410</u></td> <td><u>12/6/19</u></td> </tr> <tr> <td>0.100</td> <td><u>SD3964</u></td> <td><u>18070</u></td> <td><u>2/26/20</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>17340</u></td> <td><u>10/9/19</u></td> </tr> <tr> <td>0.300</td> <td><u>SD1024</u></td> <td><u>18110</u></td> <td><u>4/2/20</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>17817080A2</u></td> <td><u>8/5/19</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1021</u></td> <td><u>201707D</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1275</u></td> <td><u>201707E</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1013</u></td> <td><u>201707C</u></td> <td><u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG805701</u></td> <td><u>2/26/20</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>G2834</u>	N/A	N/A	0.040	<u>SD1022</u>	<u>17410</u>	<u>12/6/19</u>	0.100	<u>SD3964</u>	<u>18070</u>	<u>2/26/20</u>	0.200	<u>SD1025</u>	<u>17340</u>	<u>10/9/19</u>	0.300	<u>SD1024</u>	<u>18110</u>	<u>4/2/20</u>	0.080 DGS	N/A	<u>17817080A2</u>	<u>8/5/19</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>SD1021</u>	<u>201707D</u>	<u>7/25/19</u>	0.080	<u>DR1275</u>	<u>201707E</u>	<u>7/25/19</u>	0.200	<u>SD1013</u>	<u>201707C</u>	<u>7/24/19</u>	0.080 DGS	N/A	<u>AG805701</u>	<u>2/26/20</u>	Department Inspection Performed By <u>WJM</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1014/1020</u> Instrument <u>1001/1021</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G11621</u></td> </tr> <tr> <td>Interferent</td> <td><u>DR3855</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1021</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1275</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1013</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G11621</u>	Interferent	<u>DR3855</u>	0.050	<u>SD1021</u>	0.080	<u>DR1275</u>	0.200	<u>SD1013</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	<u>G2834</u>	N/A	N/A																																																										
0.040	<u>SD1022</u>	<u>17410</u>	<u>12/6/19</u>																																																										
0.100	<u>SD3964</u>	<u>18070</u>	<u>2/26/20</u>																																																										
0.200	<u>SD1025</u>	<u>17340</u>	<u>10/9/19</u>																																																										
0.300	<u>SD1024</u>	<u>18110</u>	<u>4/2/20</u>																																																										
0.080 DGS	N/A	<u>17817080A2</u>	<u>8/5/19</u>																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	<u>SD1021</u>	<u>201707D</u>	<u>7/25/19</u>																																																										
0.080	<u>DR1275</u>	<u>201707E</u>	<u>7/25/19</u>																																																										
0.200	<u>SD1013</u>	<u>201707C</u>	<u>7/24/19</u>																																																										
0.080 DGS	N/A	<u>AG805701</u>	<u>2/26/20</u>																																																										
Simulator	Serial Number																																																												
0.000	<u>G11621</u>																																																												
Interferent	<u>DR3855</u>																																																												
0.050	<u>SD1021</u>																																																												
0.080	<u>DR1275</u>																																																												
0.200	<u>SD1013</u>																																																												
Notes/Suggested Service: <u>CALIBRATION ADJUSTMENT performed to bring barometer closer to nominal</u>	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																																												
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	<table style="width:100%;"> <tr> <td style="text-align: center;"><u>11/14/18</u> Tech Review / Date</td> <td style="text-align: center;"><u>Brett Kirkland 11/21/18</u> Admin Review / Date</td> </tr> </table>		<u>11/14/18</u> Tech Review / Date	<u>Brett Kirkland 11/21/18</u> Admin Review / Date																																																									
<u>11/14/18</u> Tech Review / Date	<u>Brett Kirkland 11/21/18</u> Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 13:04

Date of Inspection: 11/14/2018

Serial Number: 80-003406
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.048	0.082	0.197	0.081
0.000	0.049	0.081	0.199	0.081
0.000	0.049	0.080	0.198	0.080
0.000	0.049	0.081	0.199	0.080
0.000	0.049	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.081	0.199	0.080
0.000	0.049	0.081	0.200	0.080
0.000	0.049	0.081	0.200	0.078

Standard Deviations	0.0004	0.0005	0.0008	0.0008
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

13K

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy PATRICK J MURPHY

Signature and Printed Name

11/14/2018
Date

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
11/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:10
Control Test	0.049	10:10
Air Blank	0.000	10:11
Control Test	0.049	10:12
Air Blank	0.000	10:12
Control Test	0.049	10:13
Air Blank	0.000	10:13
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
11/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.080	10:05
Air Blank	0.000	10:06
Control Test	0.080	10:07
Air Blank	0.000	10:07
Control Test	0.080	10:08
Air Blank	0.000	10:08
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
11/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:00
Control Test	0.200	10:01
Air Blank	0.000	10:01
Control Test	0.201	10:02
Air Blank	0.000	10:03
Control Test	0.201	10:03
Air Blank	0.000	10:04
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

P Murphy
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
11/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:15
Control Test	0.079	10:15
Air Blank	0.000	10:16
Control Test	0.079	10:16
Air Blank	0.000	10:17
Control Test	0.079	10:17
Air Blank	0.000	10:17
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

10 BK

P Murphy
Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003406
11/13/2018 12:19:34

Auto Calibration
Max Power Res Value = 19
Auto Range Res Value = 11

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12761, 9um Io = 13118

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0480 (-0.0040)
Sample #2 = 0.0470 (0.0150)
Sample #3 = 0.0380 (0.0400)
Sample #4 = 0.0850 (0.0400)
Avg % Abs = 0.0567 (0.0317)
STD DEU = 0.0249 (0.0144)
REL STD DEU = 44.023 (45.580)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0780 (-0.0100)
Sample #2 = 0.0940 (-0.0030)
Sample #3 = 0.0910 (-0.0080)
Sample #4 = 0.1180 (0.0000)
Avg % Abs = 0.1010 (-0.0037)
STD DEU = 0.0148 (0.0040)
REL STD DEU = 14.652 (110.221)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12751, 9um Io = 13117

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.7930 (-0.0090)
Sample #2 = 0.8100 (-0.0110)
Sample #3 = 0.7690 (0.0120)
Sample #4 = 0.8260 (-0.0030)
Avg % Abs = 0.8017 (-0.0007)
STD DEU = 0.0294 (0.0117)
REL STD DEU = 3.667 (1751.428)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.4910 (-0.0050)
Sample #2 = 1.5530 (-0.0210)
Sample #3 = 1.5050 (-0.0040)
Sample #4 = 1.5380 (-0.0140)
Avg % Abs = 1.5320 (-0.0130)
STD DEU = 0.0246 (0.0085)
REL STD DEU = 1.603 (65.723)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12746, 9um Io = 13116

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.8510 (-0.0160)
Sample #2 = 1.8050 (0.0150)
Sample #3 = 1.8070 (0.0300)
Sample #4 = 1.8770 (0.0020)
Avg % Abs = 1.8297 (0.0157)
STD DEU = 0.0410 (0.0140)
REL STD DEU = 2.241 (89.438)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.5350 (0.0040)
Sample #2 = 3.5380 (0.0220)
Sample #3 = 3.5590 (0.0090)
Sample #4 = 3.5920 (0.0110)
Avg % Abs = 3.5630 (0.0140)
STD DEU = 0.0272 (0.0070)
REL STD DEU = 0.764 (50.000)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12741, 9um Io = 13111

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.5010 (-0.0090)
Sample #2 = 3.5250 (-0.0090)
Sample #3 = 3.5590 (-0.0040)
Sample #4 = 3.5170 (0.0220)
Avg % Abs = 3.5337 (0.0030)
STD DEU = 0.0223 (0.0166)
REL STD DEU = 0.631 (554.777)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.7200 (-0.0040)
Sample #2 = 6.7630 (-0.0040)
Sample #3 = 6.8000 (0.0030)
Sample #4 = 6.7800 (0.0020)
Avg % Abs = 6.7810 (0.0003)
STD DEU = 0.0185 (0.0038)
REL STD DEU = 0.273 (1135.782)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12738, 9um Io = 13113

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.1840 (-0.0190)
Sample #2 = 5.1990 (0.0130)
Sample #3 = 5.1840 (0.0280)
Sample #4 = 5.1760 (0.0360)
Avg % Abs = 5.1863 (0.0257)
STD DEU = 0.0117 (0.0117)
REL STD DEU = 0.225 (45.492)

Channel 2 Data:

Sample % Abs (% Abs Ref)
Sample #1 = 9.8530 (-0.0080)
Sample #2 = 9.8610 (0.0260)
Sample #3 = 9.8650 (0.0330)
Sample #4 = 9.8660 (0.0320)
Avg % Abs = 9.8640 (0.0303)
STD DEU = 0.0026 (0.0038)
REL STD DEU = 0.027 (12.481)

Auto Calibration Data:

Channel 1 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.057
Std Dev = 0.02 Rel Std Dev = 44.02
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.802
Std Dev = 0.03 Rel Std Dev = 3.67
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.830
Std Dev = 0.04 Rel Std Dev = 2.24
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.534
Std Dev = 0.02 Rel Std Dev = 0.63
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.186
Std Dev = 0.01 Rel Std Dev = 0.23
Zero Order Coef = -178.82
First Order Coef = 2638.73
Second Order Coef = 29.22
Standard Deviation = 31.054647

Channel 2 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.101
Std Dev = 0.01 Rel Std Dev = 14.65
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.532
Std Dev = 0.02 Rel Std Dev = 1.60
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.563
Std Dev = 0.03 Rel Std Dev = 0.76
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.781
Std Dev = 0.02 Rel Std Dev = 0.27
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.864
Std Dev = 0.00 Rel Std Dev = 0.03
Zero Order Coef = -150.68
First Order Coef = 1334.20
Second Order Coef = 13.19
Standard Deviation = 17.263474

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.001	0.0006
0.040	0.041	-0.0011
0.100	0.100	0.0003
0.200	0.200	0.0003
0.300	0.300	-0.0001

BSK

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0003
0.040	0.040	-0.0004
0.100	0.100	-0.0002
0.200	0.200	0.0004
0.300	0.300	-0.0002

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1
 Sample #1 = 3397.00
 Sample #2 = 3388.00
 Sample #3 = 3366.00
 Sample #4 = 3442.00
 Average Result = 3398.6667
 STD DEV = 39.1067
 REL STD DEV = 1.151

***** CHANNEL 2
 Sample #1 = 3409.00
 Sample #2 = 3402.00
 Sample #3 = 3401.00
 Sample #4 = 3429.00
 Average Result = 3410.6667
 STD DEV = 15.8850
 REL STD DEV = 0.466

Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1014
 3 um H2O Adjust (mg/l*10,000) = 411
 9 um H2O Adjust (mg/l*10,000) = 399
 **** AUTO CAL PASS

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003406
 11/14/2018
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:22
Control Test	0.048	09:23
Air Blank	0.000	09:23
Control Test	0.049	09:24
Air Blank	0.000	09:24
Control Test	0.049	09:25
Air Blank	0.000	09:26
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

P Murphy
 Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003406
 11/14/2018
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:27
Control Test	0.079	09:28
Air Blank	0.000	09:28
Control Test	0.079	09:29
Air Blank	0.000	09:30
Control Test	0.080	09:30
Air Blank	0.000	09:31
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

P Murdy
 Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003406
 11/14/2018
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.198	09:33
Air Blank	0.000	09:33
Control Test	0.201	09:34
Air Blank	0.000	09:35
Control Test	0.202	09:35
Air Blank	0.000	09:36
Control Test Stats		
Average	0.2003	
Std Dev	0.0021	
Rel Std Dev(%)	1.0391	

P Murphy
 Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003406
 11/14/2018
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:37
Control Test	0.079	09:37
Air Blank	0.000	09:38
Control Test	0.080	09:38
Air Blank	0.000	09:39
Control Test	0.080	09:39
Air Blank	0.000	09:40
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

P Murphy
 Operator's Signature

POST CALIBRATION
 ADJUST STABILITIES

TSX



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003406, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003406</u>	UNCERTAINTY* ±
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L
Calibration Date:	<u>11/14/2018</u>	0.080 g/ 210 L
Calibration Time:	<u>13:04</u>	0.200 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/14/2018

Date

PATRICK J MURPHY,
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality