



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol MiamiS/N 80-003408Florida Department of
Law EnforcementDate In 09/04/2018 DI Completion Date 09/04/2018☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	Performed By <u>DELL</u>	Quality Checks	Performed By <u>DELL</u>	Flow Calibration	Performed By <u>DELL</u>																																																										
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>184</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks		Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date FDLE SEP 13 2018 Alcohol Testing Program		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>SD3967</td><td>201707D 07/25/2019</td></tr><tr><td>0.080</td><td>SD3968</td><td>201707E 07/25/2019</td></tr><tr><td>0.200</td><td>SD3969</td><td>201707C 07/24/2019</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG805701 02/26/2020</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3968	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG805701 02/26/2020	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>DELL</u> ☒ Lab Temp °C <u>22.36C</u> External Digital Therm. ID#: <u>300949</u> ☒ 34°C +/- .2 Serial #: <u>SD3967</u> ☒ 34°C +/- .2 Serial #: <u>SD3968</u> ☒ 34°C +/- .2 Serial #: <u>SD3969</u>																																												
Simulator	Serial #	Lot #/Exp																																																													
0.050	SD3967	201707D 07/25/2019																																																													
0.080	SD3968	201707E 07/25/2019																																																													
0.200	SD3969	201707C 07/24/2019																																																													
0.080 DGS	N/A	AG805701 02/26/2020																																																													
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>DELL</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1016</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2017-B</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>SD3965</td></tr><tr><td>Interferent</td><td>SD3966</td></tr><tr><td>0.050</td><td>SD3967</td></tr><tr><td>0.080</td><td>SD3968</td></tr><tr><td>0.200</td><td>SD3969</td></tr></tbody></table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____		Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	SD3965																																																														
Interferent	SD3966																																																														
0.050	SD3967																																																														
0.080	SD3968																																																														
0.200	SD3969																																																														
Notes/Suggested Service: <u>E-mailed</u> ☒ APPROVED _____ _____ _____ _____ _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>Poon 9/13/18</u> <u>J. DeLa</u> <u>9/13/18</u> Tech Review / Date Admin Review / Date																																																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP E MIAMI
Time of Inspection: 14:11

Date of Inspection: 09/04/2018

Serial Number: 80-003408
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805701 Exp: 02/26/2020
0.000	0.049	0.081	0.198	0.080
0.000	0.049	0.082	0.199	0.079
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.082	0.199	0.080
0.000	0.050	0.082	0.199	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.082	0.200	0.080
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.079

Standard Deviations	0.0004	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

Pgm

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

09/04/2018
Date

*9/13/18
JO*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-003408	Florida Highway Patrol Miami	09/04/2018	<i>PKR</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
SN: SD3967 Temp: 34.07c 0.047 to 0.053 ☒	SN: SD3968 Temp: 34.08c 0.077 to 0.083 ☒	SN: SD3969 Temp: 34.09c 0.194 to 0.206 ☒	Lot AG805701 0.077 to 0.083 ☒																																																																																																																																																
FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 09/04/2018 Software: 8100.27	FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 09/04/2018 Software: 8100.27	FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 09/04/2018 Software: 8100.27	FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 09/04/2018 Software: 8100.27																																																																																																																																																
<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:08</td></tr><tr><td>Control Test</td><td>0.049</td><td>12:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:09</td></tr><tr><td>Control Test</td><td>0.049</td><td>12:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:10</td></tr><tr><td>Control Test</td><td>0.050</td><td>12:11</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:11</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	12:08	Control Test	0.049	12:09	Air Blank	0.000	12:09	Control Test	0.049	12:10	Air Blank	0.000	12:10	Control Test	0.050	12:11	Air Blank	0.000	12:11	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:13</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:14</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:15</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:16</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	12:13	Control Test	0.079	12:14	Air Blank	0.000	12:14	Control Test	0.080	12:15	Air Blank	0.000	12:15	Control Test	0.080	12:16	Air Blank	0.000	12:16	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:18</td></tr><tr><td>Control Test</td><td>0.195</td><td>12:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:19</td></tr><tr><td>Control Test</td><td>0.198</td><td>12:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr><tr><td>Control Test</td><td>0.199</td><td>12:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1973</td><td></td></tr><tr><td>Std Dev</td><td>0.0021</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.0549</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	12:18	Control Test	0.195	12:19	Air Blank	0.000	12:19	Control Test	0.198	12:20	Air Blank	0.000	12:20	Control Test	0.199	12:21	Air Blank	0.000	12:21	Control Test Stats			Average	0.1973		Std Dev	0.0021		Rel Std Dev(%)	1.0549		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:24</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:24</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	12:24	Control Test	0.080	12:24	Air Blank	0.000	12:24	Control Test	0.080	12:25	Air Blank	0.000	12:25	Control Test	0.080	12:26	Air Blank	0.000	12:26	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:08																																																																																																																																																	
Control Test	0.049	12:09																																																																																																																																																	
Air Blank	0.000	12:09																																																																																																																																																	
Control Test	0.049	12:10																																																																																																																																																	
Air Blank	0.000	12:10																																																																																																																																																	
Control Test	0.050	12:11																																																																																																																																																	
Air Blank	0.000	12:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:13																																																																																																																																																	
Control Test	0.079	12:14																																																																																																																																																	
Air Blank	0.000	12:14																																																																																																																																																	
Control Test	0.080	12:15																																																																																																																																																	
Air Blank	0.000	12:15																																																																																																																																																	
Control Test	0.080	12:16																																																																																																																																																	
Air Blank	0.000	12:16																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:18																																																																																																																																																	
Control Test	0.195	12:19																																																																																																																																																	
Air Blank	0.000	12:19																																																																																																																																																	
Control Test	0.198	12:20																																																																																																																																																	
Air Blank	0.000	12:20																																																																																																																																																	
Control Test	0.199	12:21																																																																																																																																																	
Air Blank	0.000	12:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1973																																																																																																																																																		
Std Dev	0.0021																																																																																																																																																		
Rel Std Dev(%)	1.0549																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:24																																																																																																																																																	
Control Test	0.080	12:24																																																																																																																																																	
Air Blank	0.000	12:24																																																																																																																																																	
Control Test	0.080	12:25																																																																																																																																																	
Air Blank	0.000	12:25																																																																																																																																																	
Control Test	0.080	12:26																																																																																																																																																	
Air Blank	0.000	12:26																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
9/13/18 ja PKR Operator's Signature	PKR Operator's Signature	PKR Operator's Signature	PKR Operator's Signature																																																																																																																																																

PKR



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003408, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003408</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP TROOP E MIAMI</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>09/04/2018</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>14:11</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/04/2018

Date

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

David E Reyes-Rivera
DAVID E REYES-RIVERA,
Department Inspector

Service • Integrity • Respect • Quality

9/13/18
20