



INSTRUMENT PROCESSING SHEET

Agency Osceola County Sheriff's Office S/N 80-003937Florida Department of Law Enforcement Date In 02/23/2018 DI Completion Date 03/14/2018 Ship P/U H/D CMI EE

Intake Performed By <u>PJM</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DMB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>217</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 103</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.156</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.492</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28662</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>G11739</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3964</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG715202 06/01/2019</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	G11739	201707D 07/25/2019	0.080	SD3964	201707E 07/25/2019	0.200	DR3856	201707C 07/24/2019	0.080 DGS	N/A	AG715202 06/01/2019	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	G11739	201707D 07/25/2019																																																											
0.080	SD3964	201707E 07/25/2019																																																											
0.200	DR3856	201707C 07/24/2019																																																											
0.080 DGS	N/A	AG715202 06/01/2019																																																											
Final Release Date <div style="text-align: center; font-size: 1.2em; font-weight: bold;">FDLE</div> <div style="text-align: center; font-size: 1.1em; font-weight: bold;">MAR 14 2018</div> <div style="text-align: center;">Alcohol Testing Program</div>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>DMB</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.6</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>G11739</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>SD3964</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>DR3856</u>																																																												
Calibration Adjustment Performed By <u>DMB</u> Barometric Pressure Gauge <u>1020</u> ID # <u>28427</u> <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2879</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1024</td> <td>16320</td> <td>10/21/2018</td> </tr> <tr> <td>0.100</td> <td>G2834</td> <td>17280</td> <td>09/11/2019</td> </tr> <tr> <td>0.200</td> <td>SD1011</td> <td>17090</td> <td>02/24/2019</td> </tr> <tr> <td>0.300</td> <td>DR1275</td> <td>17140</td> <td>05/15/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>22817080A5</td> <td>10/05/2019</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>G11739</td> <td>201707D</td> <td>07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3964</td> <td>201707E</td> <td>07/25/2019</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> <td>201707C</td> <td>07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG715202</td> <td>06/01/2019</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	G2879	N/A	N/A	0.040	SD1024	16320	10/21/2018	0.100	G2834	17280	09/11/2019	0.200	SD1011	17090	02/24/2019	0.300	DR1275	17140	05/15/2019	0.080 DGS	N/A	22817080A5	10/05/2019	Simulator	Serial Number	Lot Number	Expiration	0.050	G11739	201707D	07/25/2019	0.080	SD3964	201707E	07/25/2019	0.200	DR3856	201707C	07/24/2019	0.080 DGS	N/A	AG715202	06/01/2019	Department Inspection Performed By <u>DMB</u> Barometric Pressure ID# <u>28662</u> Gauge <u>1021</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.050</td> <td>G11739</td> </tr> <tr> <td>0.080</td> <td>SD3964</td> </tr> <tr> <td>0.200</td> <td>DR3856</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD1019	Interferent	SD1021	0.050	G11739	0.080	SD3964	0.200	DR3856
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	G2879	N/A	N/A																																																										
0.040	SD1024	16320	10/21/2018																																																										
0.100	G2834	17280	09/11/2019																																																										
0.200	SD1011	17090	02/24/2019																																																										
0.300	DR1275	17140	05/15/2019																																																										
0.080 DGS	N/A	22817080A5	10/05/2019																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	G11739	201707D	07/25/2019																																																										
0.080	SD3964	201707E	07/25/2019																																																										
0.200	DR3856	201707C	07/24/2019																																																										
0.080 DGS	N/A	AG715202	06/01/2019																																																										
Simulator	Serial Number																																																												
0.000	SD1019																																																												
Interferent	SD1021																																																												
0.050	G11739																																																												
0.080	SD3964																																																												
0.200	DR3856																																																												
Notes/Suggested Service: <u>Optical bench calibration adjustment performed to bring values closer to nominal. 03/14/18 DMB</u> _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="margin-top: 10px;"> <u>DMB 3/14/18</u> <u>JJ Drake 3/14/18</u> Tech Review / Date Admin Review / Date </div>																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 14:57

Date of Inspection: 03/14/2018

Serial Number: 80-003937
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG715202 Exp: 06/01/2019
0.000	0.049	0.081	0.198	0.078
0.000	0.049	0.081	0.199	0.078
0.000	0.049	0.081	0.200	0.079
0.000	0.049	0.081	0.200	0.078
0.000	0.049	0.082	0.200	0.078
0.000	0.050	0.082	0.199	0.078
0.000	0.049	0.082	0.199	0.078
0.000	0.050	0.082	0.199	0.078
0.000	0.049	0.082	0.200	0.078
0.000	0.049	0.082	0.200	0.077

Standard Deviations	0.0004	0.0005	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

PBM

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Daniele M Bell
 _____ DANIELLE M BELL
 Signature and Printed Name

03/14/2018
 Date

3/14/18
JO

Stability Checks #80-003937 Osceola County S.D. 3/13/18 ~~RMS~~

DGS

OSCEOLA COUNTY S.D.
Intoxilyzer - Alconol Analyzer
Model 8000
03/13/2018
SN 80-003937
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	16:35
Control Test	0.050	16:36
Air Blank	0.000	16:36
Control Test	0.049	16:37
Air Blank	0.000	16:38
Control Test	0.049	16:39
Air Blank	0.000	16:39
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel. Std Dev(%)	1.1703	

RMS
Operator's Signature

8/14/18
RMS

OSCEOLA COUNTY S.D.
Intoxilyzer - Alconol Analyzer
Model 8000
03/13/2018
SN 80-003937
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	16:40
Control Test	0.083	16:41
Air Blank	0.000	16:41
Control Test	0.083	16:42
Air Blank	0.000	16:42
Control Test	0.082	16:43
Air Blank	0.000	16:44
Control Test Stats		
Average	0.0827	
Std Dev	0.0006	
Rel. Std Dev(%)	0.6984	

RMS
Operator's Signature

OSCEOLA COUNTY S.D.
Intoxilyzer - Alconol Analyzer
Model 8000
03/13/2018
SN 80-003937
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	16:45
Control Test	0.202	16:46
Air Blank	0.000	16:46
Control Test	0.203	16:47
Air Blank	0.000	16:47
Control Test	0.203	16:48
Air Blank	0.000	16:49
Control Test Stats		
Average	0.2027	
Std Dev	0.0006	
Rel. Std Dev(%)	0.2849	

RMS
Operator's Signature

OSCEOLA COUNTY S.D.
Intoxilyzer - Alconol Analyzer
Model 8000
03/13/2018
SN 30-003937
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	16:50
Control Test	0.082	16:50
Air Blank	0.000	16:50
Control Test	0.082	16:51
Air Blank	0.000	16:51
Control Test	0.082	16:52
Air Blank	0.000	16:52
Control Test Stats		
Average	0.0820	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

RMS
Operator's Signature

RMS



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003937, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003937</u>	UNCERTAINTY* ±
Owning Agency:	<u>OSCEOLA COUNTY S.O.</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/14/2018</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>14:57</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/14/2018 _____
 Date
Danielle M Bell _____
 DANIELLE M BELL,
 Department Inspector

FDLE/ATP Form 69 March 2018
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Handwritten initials and date: JB 3/14/18

Handwritten initials: DGM

Optical bench calibration adjustment data # 80-003937 Osceola County S.O. 3/14/18 DWS

OSCEOLA COUNTY S.O.
Intoxilyzer - Alcotest 9510
Model 8000
03/14/2018

Auto Calibration

Max Power Res Value = 27
Auto Range Res Value = 24

SN 80-003937
09:54:40

Sample % Abs (% Abs Ref)
Sample #1 = 1.5240 (-0.0030)
Sample #2 = 1.5250 (0.0000)
Sample #3 = 1.5240 (-0.0110)
Sample #4 = 1.5250 (-0.0070)
Avg % Abs = 1.5247 (-0.0060)
STD DEV = 0.0006 (0.0056)
REL STD DEV = 0.038 (92.796)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12917, Sum Io = 13393

Sample % Abs (% Abs Ref)
Sample #1 = 1.8400 (-0.0130)
Sample #2 = 1.8200 (0.0200)
Sample #3 = 1.8120 (0.0130)
Sample #4 = 1.8030 (0.0290)
Avg % Abs = 1.8117 (0.0207)
STD DEV = 0.0085 (0.0080)
REL STD DEV = 0.469 (38.810)

Sample % Abs (% Abs Ref)
Sample #1 = 3.5220 (-0.0130)
Sample #2 = 3.5300 (-0.0030)
Sample #3 = 3.4990 (0.0190)
Sample #4 = 3.5140 (0.0100)
Avg % Abs = 3.5143 (0.0067)
STD DEV = 0.0155 (0.0143)
REL STD DEV = 0.441 (214.418)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12921, Sum Io = 13398

Sample % Abs (% Abs Ref)
Sample #1 = 3.4980 (0.0020)
Sample #2 = 3.5440 (-0.0030)
Sample #3 = 3.5260 (0.0180)
Sample #4 = 3.5360 (0.0160)
Avg % Abs = 3.5353 (0.0087)
STD DEV = 0.0090 (0.0145)
REL STD DEV = 0.255 (166.943)

Sample % Abs (% Abs Ref)
Sample #1 = 6.7020 (0.0050)
Sample #2 = 6.7020 (0.0200)
Sample #3 = 6.7220 (0.0250)
Sample #4 = 6.7320 (0.0150)
Avg % Abs = 6.7187 (0.0200)
STD DEV = 0.0153 (0.0050)
REL STD DEV = 0.227 (25.000)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12970, Sum Io = 13418

Sample % Abs (% Abs Ref)
Sample #1 = 5.1050 (-0.0030)
Sample #2 = 5.1150 (-0.0020)
Sample #3 = 5.1040 (0.0070)
Sample #4 = 5.1270 (0.0000)
Avg % Abs = 5.1153 (0.0017)
STD DEV = 0.0115 (0.0047)
REL STD DEV = 0.225 (282.549)

Sample % Abs (% Abs Ref)
Sample #1 = 9.6840 (-0.0030)
Sample #2 = 9.6830 (0.0190)
Sample #3 = 9.6660 (0.0220)
Sample #4 = 9.6950 (0.0150)
Avg % Abs = 9.6813 (0.0187)
STD DEV = 0.0146 (0.0035)
REL STD DEV = 0.151 (18.814)

Sol Value = 0.400 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12921, Sum Io = 13398

Sample % Abs (% Abs Ref)
Sample #1 = 3.4980 (0.0020)
Sample #2 = 3.5440 (-0.0030)
Sample #3 = 3.5260 (0.0180)
Sample #4 = 3.5360 (0.0160)
Avg % Abs = 3.5353 (0.0087)
STD DEV = 0.0090 (0.0145)
REL STD DEV = 0.255 (166.943)

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.100
Std Dev = 0.01 Rel Std Dev = 14.42
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.777
Std Dev = 0.03 Rel Std Dev = 3.70
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.812
Std Dev = 0.01 Rel Std Dev = 0.47
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.535
Std Dev = 0.01 Rel Std Dev = 0.26
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.115
Std Dev = 0.01 Rel Std Dev = 0.22
Zero Order Coef = -222.35
First Order Coef = 2661.83
Second Order Coef = 32.97
Standard Deviation = 56.93891

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.185
Std Dev = 0.02 Rel Std Dev = 5.69
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.525
Std Dev = 0.00 Rel Std Dev = 0.04
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.514
Std Dev = 0.02 Rel Std Dev = 0.44
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.719
Std Dev = 0.02 Rel Std Dev = 0.23
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.681
Std Dev = 0.01 Rel Std Dev = 0.15
Zero Order Coef = -230.40
First Order Coef = 1365.64
Second Order Coef = 13.66
Standard Deviation = 29.150187

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.185
Std Dev = 0.02 Rel Std Dev = 5.69
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.525
Std Dev = 0.00 Rel Std Dev = 0.04
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.514
Std Dev = 0.02 Rel Std Dev = 0.44
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.719
Std Dev = 0.02 Rel Std Dev = 0.23
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.681
Std Dev = 0.01 Rel Std Dev = 0.15
Zero Order Coef = -230.40
First Order Coef = 1365.64
Second Order Coef = 13.66
Standard Deviation = 29.150187

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12963, Sum Io = 13410

Sample % Abs (% Abs Ref)
Sample #1 = 0.8010 (-0.0110)
Sample #2 = 0.7560 (0.0250)
Sample #3 = 0.8100 (0.0120)
Sample #4 = 0.7660 (0.0400)
Avg % Abs = 0.7773 (0.0257)
STD DEV = 0.0287 (0.0140)
REL STD DEV = 3.696 (54.592)

Solution Stats Quadratic Fit Chan 1
Act g/210L Fit Residual g/210L
0.000 0.001 -0.0003
0.040 0.039 0.0008
0.100 0.099 0.0011
0.200 0.202 -0.0013
0.300 0.299 0.0016

Solution Stats Quadratic Fit Chan 2
Act g/210L Fit Residual g/210L
0.000 0.000 -0.0005
0.040 0.040 0.0004
0.100 0.099 0.0005
0.200 0.201 -0.0003
0.300 0.300 0.0003

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

Sample #1 = 3247.00
Sample #2 = 3272.00
Sample #3 = 3339.00
Sample #4 = 3227.00
Average Result = 3279.3333
STD DEV = 56.3590
REL STD DEV = 1.719

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

Sample #1 = 3259.00
Sample #2 = 3312.00
Sample #3 = 3310.00
Sample #4 = 3328.00
Average Result = 3316.6667
STD DEV = 9.6658
REL STD DEV = 0.297

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1020
3 um H2O Adjust (mg/l * 10,000) = 530
9 um H2O Adjust (mg/l * 10,000) = 493
**** AUTO CAL PASS

DWS

3/14/18

Post Calibration Adjustment

Stability Checks # 80-003937 Osceola County S.O. 3/14/18 ~~DBS~~

DBS

OSCEOLA COUNTY S.O.
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-003937
03/14/2018
Software: 8100.27

OSCEOLA COUNTY S.O.
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-003937
03/14/2018
Software: 8100.27

OSCEOLA COUNTY S.O.
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-003937
03/14/2018
Software: 8100.27

OSCEOLA COUNTY S.O.
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-003937
03/14/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:51
Control Test	0.079	11:51
Air Blank	0.000	11:52
Control Test	0.079	11:52
Air Blank	0.000	11:53
Control Test	0.079	11:53
Air Blank	0.000	11:55
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	11:45
Control Test	0.200	11:46
Air Blank	0.000	11:47
Control Test	0.200	11:47
Air Blank	0.000	11:48
Control Test	0.200	11:48
Air Blank	0.000	11:49
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	11:40
Control Test	0.081	11:41
Air Blank	0.000	11:41
Control Test	0.082	11:42
Air Blank	0.000	11:43
Control Test	0.081	11:43
Air Blank	0.000	11:44
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

Test	g/210L	Time
Air Blank	0.000	11:35
Control Test	0.050	11:35
Air Blank	0.000	11:36
Control Test	0.050	11:37
Air Blank	0.000	11:37
Control Test	0.049	11:38
Air Blank	0.000	11:38
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

DBS

Operator's Signature

3/14/18
DB

DBS

Operator's Signature

DBS

Operator's Signature

DBS

Operator's Signature

DBS