



# INSTRUMENT PROCESSING SHEET

Agency Hillsborough County Sheriff's Office s/N 80-005113

Florida Department of Law Enforcement Date In 04/11/2018 DI Completion Date 04/16/2018  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>TG</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>DMB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>252</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.156</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28662</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
--	---	--

**Final Release Date**

**FDLE**

APR 16 2018

Alcohol Testing Program

Simulator	Serial #	Lot #/Exp
0.050	G11739	201707D 07/25/2019
0.080	SD3964	201707E 07/25/2019
0.200	DR3856	201707C 07/24/2019
0.080 DGS	N/A	AG805701 02/26/2020

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By DMB

Lab Temp °C 21.8  
 External Digital Therm. ID#: 300502  
 34°C +/-2 Serial #: G2835  
 34°C +/-2 Serial #: G2840  
 34°C +/-2 Serial #: SD1025

**Calibration Adjustment** Performed By DMB

Barometric Pressure Gauge 1017 ID # 28427

Simulator	Serial Number	Lot Number	Expiration
0.000	G8144	N/A	N/A
0.040	G2403	16320	10/21/2018
0.100	G2879	17280	09/11/2019
0.200	G3709	17090	02/24/2019
0.300	G8149	17140	05/15/2019
0.080 DGS	N/A	22817080A5	10/05/2019

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050	G2835	201707D	07/25/2019
0.080	G2840	201707E	07/25/2019
0.200	SD1025	201707C	07/24/2019
0.080 DGS	N/A	AG805701	02/26/2020

**Department Inspection** Performed By DMB

Barometric Pressure ID# 28662  
 Gauge 1019 Instrument 1017  
 Mouth Alcohol Solution Lot # 2017-B  
 Acetone Stock Solution Lot # 2018-A

Simulator	Serial Number
0.000	G2880
Interferent	G2407
0.050	G2835
0.080	G2840
0.200	SD1025

**Attachments**

Form 41  Post-Stability Checks  
 Stability Checks  Flow Calibration  
 Calibration Certificate  Form 40  
 Calibration Adjustment  Other \_\_\_\_\_

Notes/Suggested Service: Performed optical bench calibration to bring barometric pressure value closer to nominal. DMB 04/13/18

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

DMB 4/16/18 J. Graham 4/16/18  
 Tech Review / Date Admin Review / Date

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO  
Time of Inspection: 11:37

Date of Inspection: 04/16/2018

Serial Number: 80-005113  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.049	0.080	0.200	0.080
0.000	0.050	0.082	0.201	0.080
0.000	0.050	0.082	0.202	0.079
0.000	0.050	0.082	0.201	0.080
0.000	0.050	0.082	0.202	0.080
0.000	0.050	0.082	0.202	0.079
0.000	0.050	0.082	0.202	0.079
0.000	0.050	0.082	0.201	0.079
0.000	0.050	0.082	0.202	0.079
0.000	0.050	0.082	0.202	0.079

Standard Deviations	0.0003	0.0006	0.0007	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

*DM*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Danielle M Bell*

DANIELLE M BELL

Signature and Printed Name

04/16/2018  
Date

*4/16/18  
DM*

# Stability Checks #80-005113 Hillsborough County S.O. 4/11/18 *RMS*

*DAS*

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005113  
04/11/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:37
Control Test	0.049	12:38
Air Blank	0.000	12:38
Control Test	0.049	12:39
Air Blank	0.000	12:39
Control Test	0.049	12:40
Air Blank	0.000	12:41
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005113  
04/11/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:41
Control Test	0.081	12:42
Air Blank	0.000	12:43
Control Test	0.080	12:43
Air Blank	0.000	12:44
Control Test	0.080	12:45
Air Blank	0.000	12:45
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005113  
04/11/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:46
Control Test	0.197	12:47
Air Blank	0.000	12:47
Control Test	0.197	12:48
Air Blank	0.000	12:48
Control Test	0.197	12:49
Air Blank	0.000	12:50
Control Test Stats		
Average	0.1970	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005113  
04/11/2018  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:56
Control Test	0.081	12:56
Air Blank	0.000	12:57
Control Test	0.081	12:57
Air Blank	0.000	12:57
Control Test	0.081	12:58
Air Blank	0.000	12:58
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*4/11/18*

*RMS*  
Operator's Signature

*RMS*  
Operator's Signature

*RMS*  
Operator's Signature

*RMS*  
Operator's Signature

*RMS*



Florida Department of Law Enforcement  
 Alcohol Testing Program  
 2729 Fort Knox Blvd.  
 Bldg. 2, Suite 1300  
 Tallahassee, FL 32308

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005113, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005113</u>	UNCERTAINTY* ±	
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/16/2018</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:37</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
 \*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/16/2018

Date

DANIELLE M BELL,  
 Department Inspector

FDLE/ATP Form 69 March 2018  
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*4/16/18*

*psm*

# Optical Bench Calibration Adjustment Data #80-005113 Hillsborough County S.O. 4/13/18 *CRS*

*CRS*

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
 <<<<< CHANNEL 1 >>>>>  
 Solution Stats Quadratic Fit Chan 1  
 Act Fit Residual  
 g/210L g/210L g/210L  
 0.000 0.000 -0.0002  
 0.040 0.040 0.0005  
 0.100 0.101 -0.0005  
 0.200 0.200 0.0003  
 0.300 0.300 -0.0001

Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.119  
 Std Dev = 0.03 Rel Std Dev = 21.56  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 0.871  
 Std Dev = 0.02 Rel Std Dev = 2.76  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 2.024  
 Std Dev = 0.03 Rel Std Dev = 1.24  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 3.873  
 Std Dev = 0.02 Rel Std Dev = 0.61  
 Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 5.708  
 Std Dev = 0.02 Rel Std Dev = 0.36  
 Zero Order Coef = -286.48  
 First Order Coef = 2479.35  
 Second Order Coef = 13.01  
 Standard Deviation = 18.162119

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 7.0750 (-0.00500)  
 Sample #2 = 7.0960 (0.0380)  
 Sample #3 = 7.0750 (0.0540)  
 Sample #4 = 7.0910 (0.0500)  
 Avg % Abs = 7.0873 (0.0473)  
 STD DEV = 0.0110 (0.0083)  
 REL STD DEV = 0.155 (17.592)

<<<<< CHANNEL 1 >>>>>  
 Sol Value = 0.300 g/210L \*\*\*  
 Fit value = 1.4286 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 11716, Sum Io = 14166  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 5.7350 (-0.0100)  
 Sample #2 = 5.6860 (0.0660)  
 Sample #3 = 5.7100 (0.0650)  
 Sample #4 = 5.7270 (0.0770)  
 Avg % Abs = 5.7077 (0.0693)  
 STD DEV = 0.0206 (0.0067)  
 REL STD DEV = 0.361 (9.603)

<<<<< CHANNEL 2 >>>>>  
 Sol Value = 0.100 g/210L \*\*\*  
 Fit value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 11742, Sum Io = 14178  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 2.0340 (-0.0080)  
 Sample #2 = 2.0180 (0.0420)  
 Sample #3 = 2.0030 (0.0830)  
 Sample #4 = 2.0520 (0.0850)  
 Avg % Abs = 2.0243 (0.0700)  
 STD DEV = 0.0251 (0.0243)  
 REL STD DEV = 1.240 (34.670)

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
 Sol Value = 0.080 g/210L \*\*\*  
 Fit value = 0.3810 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3230.00  
 Sample #2 = 3230.00  
 Sample #3 = 3215.00  
 Sample #4 = 3179.00  
 Average Result = 3208.0000  
 STD DEV = 26.2107  
 REL STD DEV = 0.817  
 \*\*\*\*\*  
 \*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sample #1 = 3475.00  
 Sample #2 = 3470.00  
 Sample #3 = 3477.00  
 Sample #4 = 3448.00  
 Average Result = 3465.0000  
 STD DEV = 15.1327  
 REL STD DEV = 0.437  
 \*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1017  
 3 um H2O Adjust (mg/l\*10,000) = 601  
 9 um H2O Adjust (mg/l\*10,000) = 344  
 \*\*\*\*\* AUTO CAL PASS \*\*\*\*\*

<<<<< CHANNEL 2 >>>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.190  
 Std Dev = 0.02 Rel Std Dev = 25.01  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 1.544  
 Std Dev = 0.01 Rel Std Dev = 0.36  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 3.705  
 Std Dev = 0.02 Rel Std Dev = 0.46  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 7.087  
 Std Dev = 0.01 Rel Std Dev = 0.15  
 Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 10.350  
 Std Dev = 0.00 Rel Std Dev = 0.02  
 Zero Order Coef = -114.30  
 First Order Coef = 1283.15  
 Second Order Coef = 10.50  
 Standard Deviation = 15.577538

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 10.3670 (0.0040)  
 Sample #2 = 10.3490 (0.0630)  
 Sample #3 = 10.3480 (0.0750)  
 Sample #4 = 10.3520 (0.0780)  
 Avg % Abs = 10.3497 (0.0720)  
 REL STD DEV = 0.0021 (0.0079)  
 REL STD DEV = 0.020 (11.024)

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.7160 (-0.0140)  
 Sample #2 = 3.7190 (0.0120)  
 Sample #3 = 3.6660 (0.0450)  
 Sample #4 = 3.7110 (0.0430)  
 Avg % Abs = 3.7053 (0.0333)  
 STD DEV = 0.0172 (0.0185)  
 REL STD DEV = 0.465 (55.507)

<<<<< CHANNEL 2 >>>>>  
 Sol Value = 0.200 g/210L \*\*\*  
 Fit value = 0.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 11728, Sum Io = 14171  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.8750 (-0.0040)  
 Sample #2 = 3.8970 (0.0420)  
 Sample #3 = 3.8500 (0.0890)  
 Sample #4 = 3.8720 (0.0840)  
 Avg % Abs = 3.8730 (0.0717)  
 STD DEV = 0.0235 (0.0258)  
 REL STD DEV = 0.607 (36.019)

<<<<< CHANNEL 1 >>>>>  
 Sol Value = 0.100 g/210L \*\*\*  
 Fit value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 11742, Sum Io = 14178  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 2.0340 (-0.0080)  
 Sample #2 = 2.0180 (0.0420)  
 Sample #3 = 2.0030 (0.0830)  
 Sample #4 = 2.0520 (0.0850)  
 Avg % Abs = 2.0243 (0.0700)  
 STD DEV = 0.0251 (0.0243)  
 REL STD DEV = 1.240 (34.670)

<<<<< CHANNEL 2 >>>>>  
 Sol Value = 0.100 g/210L \*\*\*  
 Fit value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 11742, Sum Io = 14178  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 2.0340 (-0.0080)  
 Sample #2 = 2.0180 (0.0420)  
 Sample #3 = 2.0030 (0.0830)  
 Sample #4 = 2.0520 (0.0850)  
 Avg % Abs = 2.0243 (0.0700)  
 STD DEV = 0.0251 (0.0243)  
 REL STD DEV = 1.240 (34.670)

<<<<< CHANNEL 2 >>>>>  
 Sol Value = 0.100 g/210L \*\*\*  
 Fit value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 11742, Sum Io = 14178  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 2.0340 (-0.0080)  
 Sample #2 = 2.0180 (0.0420)  
 Sample #3 = 2.0030 (0.0830)  
 Sample #4 = 2.0520 (0.0850)  
 Avg % Abs = 2.0243 (0.0700)  
 STD DEV = 0.0251 (0.0243)  
 REL STD DEV = 1.240 (34.670)

<<<<< CHANNEL 2 >>>>>  
 Sol Value = 0.100 g/210L \*\*\*  
 Fit value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 11742, Sum Io = 14178  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 2.0340 (-0.0080)  
 Sample #2 = 2.0180 (0.0420)  
 Sample #3 = 2.0030 (0.0830)  
 Sample #4 = 2.0520 (0.0850)  
 Avg % Abs = 2.0243 (0.0700)  
 STD DEV = 0.0251 (0.0243)  
 REL STD DEV = 1.240 (34.670)

<<<<< CHANNEL 2 >>>>>  
 Sol Value = 0.100 g/210L \*\*\*  
 Fit value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 11742, Sum Io = 14178  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 2.0340 (-0.0080)  
 Sample #2 = 2.0180 (0.0420)  
 Sample #3 = 2.0030 (0.0830)  
 Sample #4 = 2.0520 (0.0850)  
 Avg % Abs = 2.0243 (0.0700)  
 STD DEV = 0.0251 (0.0243)  
 REL STD DEV = 1.240 (34.670)

*4/16/18*  
*CRS*

Post Calibration Adjustment  
 Stability Checks #80-005113 Hillsborough County S.O. 4/13/18 ~~DOB~~

DOB

HILLSBOROUGH CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005113  
 04/13/2018  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:07
Control Test	0.050	11:08
Air Blank	0.000	11:08
Control Test	0.050	11:09
Air Blank	0.000	11:10
Control Test	0.050	11:10
Air Blank	0.000	11:11
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

*[Signature]*

Operator's Signature

4/16/18  
*[Signature]*

HILLSBOROUGH CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005113  
 04/13/2018  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:12
Control Test	0.082	11:13
Air Blank	0.000	11:13
Control Test	0.081	11:14
Air Blank	0.000	11:15
Control Test	0.082	11:15
Air Blank	0.000	11:16
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel. Std Dev(%)	0.7070	

*[Signature]*

Operator's Signature

HILLSBOROUGH CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005113  
 04/13/2018  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:17
Control Test	0.201	11:18
Air Blank	0.000	11:18
Control Test	0.201	11:19
Air Blank	0.000	11:19
Control Test	0.200	11:20
Air Blank	0.000	11:21
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel. Std Dev(%)	0.2877	

*[Signature]*

Operator's Signature

HILLSBOROUGH CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005113  
 04/13/2018  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:21
Control Test	0.079	11:22
Air Blank	0.000	11:22
Control Test	0.080	11:22
Air Blank	0.000	11:23
Control Test	0.080	11:23
Air Blank	0.000	11:24
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel. Std Dev(%)	0.7247	

*[Signature]*

Operator's Signature

*[Signature]*