



INSTRUMENT PROCESSING SHEET

Agency Martin County SO

S/N 80-006168

Florida Department of Law Enforcement

Date In 02/07/2018

DI Completion Date 2/13/18

Ship P/U H/D CMI EE

Intake Performed By <u>JE</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____	Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>196</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>1164</u> (.139 - .169) 36 mm <u>1179</u> (.156 - .190) 53 mm <u>1246</u> (.228 - .278) 103 mm <u>1515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28462</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)															
Final Release Date FDLE FEB 13 2018 Alcohol Testing Program	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>G2835</u></td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td><u>SD1013</u></td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG708807</u> 3-29-19</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>G2835</u>	201707D 07/25/2019	0.080	<u>SD1013</u>	201707E 07/25/2019	0.200	<u>SD1025</u>	201707C 07/24/2019	0.080 DGS	N/A	<u>AG708807</u> 3-29-19	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.0</u> External Digital Therm. ID#: <u>300500</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2835</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1013</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1025</u>
Simulator	Serial #	Lot #/Exp															
0.050	<u>G2835</u>	201707D 07/25/2019															
0.080	<u>SD1013</u>	201707E 07/25/2019															
0.200	<u>SD1025</u>	201707C 07/24/2019															
0.080 DGS	N/A	<u>AG708807</u> 3-29-19															

Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge <u>1029</u> ID # <u>28427</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>SD1016</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>SD1024</u></td> <td><u>16320</u></td> <td><u>10-21-18</u></td> </tr> <tr> <td>0.100</td> <td><u>DR3855</u></td> <td><u>17280</u></td> <td><u>9-11-19</u></td> </tr> <tr> <td>0.200</td> <td><u>G2407</u></td> <td><u>17090</u></td> <td><u>2-24-19</u></td> </tr> <tr> <td>0.300</td> <td><u>DR1275</u></td> <td><u>17140</u></td> <td><u>5-15-19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A <u>228</u></td> <td><u>17080A5</u></td> <td><u>10-5-19</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>G2835</u></td> <td><u>201707D</u></td> <td><u>7-25-19</u></td> </tr> <tr> <td>0.080</td> <td><u>SD1013</u></td> <td><u>201707E</u></td> <td><u>7-25-19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>201707C</u></td> <td><u>7-24-19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG715202</u></td> <td><u>6-1-19</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>SD1016</u>	N/A	N/A	0.040	<u>SD1024</u>	<u>16320</u>	<u>10-21-18</u>	0.100	<u>DR3855</u>	<u>17280</u>	<u>9-11-19</u>	0.200	<u>G2407</u>	<u>17090</u>	<u>2-24-19</u>	0.300	<u>DR1275</u>	<u>17140</u>	<u>5-15-19</u>	0.080 DGS	N/A <u>228</u>	<u>17080A5</u>	<u>10-5-19</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>G2835</u>	<u>201707D</u>	<u>7-25-19</u>	0.080	<u>SD1013</u>	<u>201707E</u>	<u>7-25-19</u>	0.200	<u>SD1025</u>	<u>201707C</u>	<u>7-24-19</u>	0.080 DGS	N/A	<u>AG715202</u>	<u>6-1-19</u>	Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>28462</u> Gauge <u>1030</u> Instrument <u>1028</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2018-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2880</u></td> </tr> <tr> <td>Interferent</td> <td><u>G8144</u></td> </tr> <tr> <td>0.050</td> <td><u>G2835</u></td> </tr> <tr> <td>0.080</td> <td><u>SD1013</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G2880</u>	Interferent	<u>G8144</u>	0.050	<u>G2835</u>	0.080	<u>SD1013</u>	0.200	<u>SD1025</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	<u>SD1016</u>	N/A	N/A																																																										
0.040	<u>SD1024</u>	<u>16320</u>	<u>10-21-18</u>																																																										
0.100	<u>DR3855</u>	<u>17280</u>	<u>9-11-19</u>																																																										
0.200	<u>G2407</u>	<u>17090</u>	<u>2-24-19</u>																																																										
0.300	<u>DR1275</u>	<u>17140</u>	<u>5-15-19</u>																																																										
0.080 DGS	N/A <u>228</u>	<u>17080A5</u>	<u>10-5-19</u>																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	<u>G2835</u>	<u>201707D</u>	<u>7-25-19</u>																																																										
0.080	<u>SD1013</u>	<u>201707E</u>	<u>7-25-19</u>																																																										
0.200	<u>SD1025</u>	<u>201707C</u>	<u>7-24-19</u>																																																										
0.080 DGS	N/A	<u>AG715202</u>	<u>6-1-19</u>																																																										
Simulator	Serial Number																																																												
0.000	<u>G2880</u>																																																												
Interferent	<u>G8144</u>																																																												
0.050	<u>G2835</u>																																																												
0.080	<u>SD1013</u>																																																												
0.200	<u>SD1025</u>																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>Pagan 2/13/18</u> <u>J. Baker 2/13/18</u> Tech Review / Date Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO
Time of Inspection: 13:47

Date of Inspection: 02/13/2018

Serial Number: 80-006168
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG715202 Exp: 06/01/2019
0.000	0.049	0.081	0.199	0.080
0.000	0.049	0.082	0.200	0.080
0.000	0.050	0.082	0.200	0.080
0.000	0.049	0.082	0.199	0.080
0.000	0.050	0.082	0.199	0.080
0.000	0.049	0.083	0.199	0.080
0.000	0.049	0.082	0.199	0.080
0.000	0.050	0.083	0.199	0.080
0.000	0.050	0.082	0.199	0.080
0.000	0.050	0.082	0.200	0.081

Standard Deviations	0.0005	0.0005	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

PSM

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

02/13/2018
Date

*2/13/18
JO*

STABILITY CHECKS #80-006168

MARTIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006168
02/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:07
Control Test	0.049	10:08
Air Blank	0.000	10:09
Control Test	0.050	10:09
Air Blank	0.000	10:10
Control Test	0.049	10:11
Air Blank	0.000	10:11
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

SP
Operator's Signature

2/13/18
AD

MARTIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006168
02/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:20
Control Test	0.082	10:21
Air Blank	0.000	10:22
Control Test	0.083	10:22
Air Blank	0.000	10:23
Control Test	0.082	10:23
Air Blank	0.000	10:24
Control Test Stats		
Average	0.0823	
Std Dev	0.0006	
Rel Std Dev(%)	0.7012	

SP
Operator's Signature

MARTIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006168
02/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:36
Control Test	0.201	10:36
Air Blank	0.000	10:37
Control Test	0.201	10:38
Air Blank	0.000	10:38
Control Test	0.201	10:39
Air Blank	0.000	10:39
Control Test Stats		
Average	0.2010	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
Operator's Signature

MARTIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006168
02/13/2018
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:00
Control Test	0.081	10:00
Air Blank	0.000	10:01
Control Test	0.080	10:01
Air Blank	0.000	10:02
Control Test	0.081	10:02
Air Blank	0.000	10:02
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

DGS

SP
Operator's Signature

Blair



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006168, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006168</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MARTIN COUNTY SO</u>	0.05 g/ 210 L	0.004
Calibration Date:	<u>02/13/2018</u>	0.08 g/ 210 L	0.005
Calibration Time:	<u>13:47</u>	0.20 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by Calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the Target Alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS) ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/13/2018

Date

Shayla Platt

SHAYLA D PLATT,
 Department Inspector

FDLE/ATP Form 69 January 2018
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*2/13/18
 SP*

SP

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5750 (-0.0010)
 Sample #2 = 1.6010 (0.0000)
 Sample #3 = 1.5740 (0.0220)
 Sample #4 = 1.5780 (0.0130)
 Avg % Abs = 1.5843 (0.0117)
 STD DEV = 0.0145 (0.0111)
 REL STD DEV = 0.920 (94.804)

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 6.9660 (0.0140)
 Sample #2 = 6.9960 (0.0140)
 Sample #3 = 6.9760 (0.0370)
 Sample #4 = 6.9660 (0.0400)
 Avg % Abs = 6.9793 (0.0303)
 STD DEV = 0.0153 (0.0142)
 REL STD DEV = 0.219 (46.894)

***** AUTO CAL DATA *****
 <<<< CHANNEL 1 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.042
 Std Dev = 0.03 Rel Std Dev = 74.27
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.766
 Std Dev = 0.03 Rel Std Dev = 4.05
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.851
 Std Dev = 0.03 Rel Std Dev = 1.69
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.564

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0000
0.040	0.040	0.0003
0.100	0.100	-0.0005
0.200	0.200	0.0003
0.300	0.300	-0.0001

MARTIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 02/13/2018
 SN 80-006168
 11:01:29
 Auto Calibration
 Max Power Res Value = 102
 Auto Range Res Value = 71

Sol Value = 0.000 g/210L ***
 Fit Value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12620, 9um Io = 13278
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.0550 (-0.0080)
 Sample #2 = 0.0780 (0.0170)
 Sample #3 = 0.0230 (0.0640)
 Sample #4 = 0.0250 (0.0780)
 Avg % Abs = 0.0420 (0.0530)
 STD DEV = 0.0312 (0.0320)
 REL STD DEV = 74.269 (60.289)

Sol Value = 0.100 g/210L ***
 Fit Value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12599, 9um Io = 13274
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.8090 (0.0000)
 Sample #2 = 1.8150 (0.0180)
 Sample #3 = 1.8700 (0.0000)
 Sample #4 = 1.8680 (0.0130)
 Avg % Abs = 1.8510 (0.0103)
 STD DEV = 0.0312 (0.0093)
 REL STD DEV = 1.685 (89.918)

Sol Value = 0.300 g/210L ***
 Fit Value = 1.4286 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12591, 9um Io = 13272
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 5.2470 (-0.0140)
 Sample #2 = 5.1970 (0.0360)
 Sample #3 = 5.2550 (0.0150)
 Sample #4 = 5.2470 (0.0420)
 Avg % Abs = 5.2330 (0.0310)
 STD DEV = 0.0314 (0.0142)
 REL STD DEV = 0.601 (45.734)

<<<< CHANNEL 2 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.165
 Std Dev = 0.02 Rel Std Dev = 11.45
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.584
 Std Dev = 0.01 Rel Std Dev = 0.92
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.681
 Std Dev = 0.01 Rel Std Dev = 0.21
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.979
 Std Dev = 0.02 Rel Std Dev = 0.22
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 10.078
 Std Dev = 0.03 Rel Std Dev = 0.30
 Zero Order Coef = -204.59
 First Order Coef = 1299.27
 Second Order Coef = 13.70
 Standard Deviation = 10.64924

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	0.0003
0.100	0.100	-0.0000
0.200	0.200	-0.0001
0.300	0.300	0.0001

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1710 (-0.0240)
 Sample #2 = 0.1870 (-0.0120)
 Sample #3 = 0.1570 (0.0080)
 Sample #4 = 0.1520 (0.0150)
 Avg % Abs = 0.1653 (0.0037)
 STD DEV = 0.0189 (0.0140)
 REL STD DEV = 11.449 (382.143)

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.6640 (-0.0060)
 Sample #2 = 3.6730 (0.0000)
 Sample #3 = 3.6820 (0.0010)
 Sample #4 = 3.6880 (0.0000)
 Avg % Abs = 3.6810 (0.0003)
 STD DEV = 0.0075 (0.0006)
 REL STD DEV = 0.205 (173.205)

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 10.0370 (-0.0010)
 Sample #2 = 10.0460 (0.0270)
 Sample #3 = 10.0830 (0.0200)
 Sample #4 = 10.1060 (0.0200)
 Avg % Abs = 10.0783 (0.0223)
 STD DEV = 0.0303 (0.0040)
 REL STD DEV = 0.300 (18.096)

<<<< CHANNEL 1 >>>>
 Sol Value = 0.080 g/210L ***
 Fit Value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1 *****
 Sample #1 = 3432.00
 Sample #2 = 3360.00
 Sample #3 = 3391.00
 Sample #4 = 3353.00
 Average Result = 3368.0000
 STD DEV = 20.2237
 REL STD DEV = 0.600
 ***** CHANNEL 2 *****
 Sample #1 = 3240.00
 Sample #2 = 3230.00
 Sample #3 = 3228.00
 Sample #4 = 3236.00
 Average Result = 3231.3333
 STD DEV = 4.1633
 REL STD DEV = 0.129
 ***** CHANNEL 1 *****
 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1028
 3 um H2O Adjust (mg/l*10,000) = 441
 9 um H2O Adjust (mg/l*10,000) = 578
 ***** AUTO CAL PASS *****

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	0.0003
0.100	0.100	-0.0000
0.200	0.200	-0.0001
0.300	0.300	0.0001

Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12616, 9um Io = 13277
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.7430 (-0.0090)
 Sample #2 = 0.7970 (-0.0120)
 Sample #3 = 0.7650 (0.0360)
 Sample #4 = 0.7350 (0.0420)
 Avg % Abs = 0.7657 (0.0220)
 STD DEV = 0.0310 (0.0296)
 REL STD DEV = 4.049 (134.533)

Sol Value = 0.200 g/210L ***
 Fit Value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12595, 9um Io = 13274
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.5720 (-0.0160)
 Sample #2 = 3.6040 (0.0080)
 Sample #3 = 3.5370 (0.0430)
 Sample #4 = 3.5500 (0.0460)
 Avg % Abs = 3.5637 (0.0323)
 STD DEV = 0.0355 (0.0211)
 REL STD DEV = 0.997 (65.340)

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 10.0370 (-0.0010)
 Sample #2 = 10.0460 (0.0270)
 Sample #3 = 10.0830 (0.0200)
 Sample #4 = 10.1060 (0.0200)
 Avg % Abs = 10.0783 (0.0223)
 STD DEV = 0.0303 (0.0040)
 REL STD DEV = 0.300 (18.096)

<<<< CHANNEL 1 >>>>
 Sol Value = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.165
 Std Dev = 0.02 Rel Std Dev = 11.45
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.584
 Std Dev = 0.01 Rel Std Dev = 0.92
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.681
 Std Dev = 0.01 Rel Std Dev = 0.21
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.979
 Std Dev = 0.02 Rel Std Dev = 0.22
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 10.078
 Std Dev = 0.03 Rel Std Dev = 0.30
 Zero Order Coef = -204.59
 First Order Coef = 1299.27
 Second Order Coef = 13.70
 Standard Deviation = 10.64924

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	0.0003
0.100	0.100	-0.0000
0.200	0.200	-0.0001
0.300	0.300	0.0001

CALC ADJUSTMENT
 #80-006168 SP
 2/13/18
 JB

PDM

#80-006168
 POST CAL ADJUST STABILITY CHECKS

MARTIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006168
 02/13/2018
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:51
Control Test	0.049	11:51
Air Blank	0.000	11:52
Control Test	0.049	11:53
Air Blank	0.000	11:53
Control Test	0.049	11:54
Air Blank	0.000	11:54
Control Test	0.049	11:54
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MARTIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006168
 02/13/2018
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:56
Control Test	0.081	11:56
Air Blank	0.000	11:57
Control Test	0.081	11:58
Air Blank	0.000	11:58
Control Test	0.082	11:59
Air Blank	0.000	11:59
Control Test	0.0813	11:59
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

MARTIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006168
 02/13/2018
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:00
Control Test	0.197	12:01
Air Blank	0.000	12:01
Control Test	0.199	12:02
Air Blank	0.000	12:03
Control Test	0.200	12:03
Air Blank	0.000	12:04
Control Test	0.1987	12:04
Average	0.1987	
Std Dev	0.0015	
Rel Std Dev(%)	0.7669	

MARTIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006168
 02/13/2018
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:06
Control Test	0.080	12:07
Air Blank	0.000	12:07
Control Test	0.081	12:07
Air Blank	0.000	12:08
Control Test	0.081	12:08
Air Blank	0.000	12:09
Control Test	0.080	12:09
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

DGS

SP

Operator's Signature

2/13/18
 DA

SP

Operator's Signature

SP

Operator's Signature

SP

Operator's Signature

SP